

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP100724\
 Data File : PP067549.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 07 Oct 2024 15:00
 Operator : YP\AJ
 Sample : P4324-01MS
 Misc :
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 LAW-24-00169-00170MS

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 08 02:08:36 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP092624.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Sep 27 06:59:05 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.756	4.054	13431541	14283984	9.993	9.560
2)	SA Decachlor...	10.671	9.222	17112897	17782059	11.788	12.904
Target Compounds							
3)	L1 AR-1016-1	5.920	5.163	12239216	12895691	256.696	250.019
4)	L1 AR-1016-2	5.943	5.183	16733952	17182167	241.674	241.648
5)	L1 AR-1016-3	6.006	5.363	11350644	9831487	254.099	247.036
6)	L1 AR-1016-4	6.104	5.402	9076756	8473000	252.586	245.387
7)	L1 AR-1016-5	6.399	5.622	9051507	10397788	245.633	242.316
8)	L2 AR-1221-1	4.955	4.266	1359664	1093217	81.219	57.183 #
9)	L2 AR-1221-2	5.045	4.357	1504769	1614396	118.775	110.038
10)	L2 AR-1221-3	5.122	4.436	6818678	7080205	176.187	163.275
11)	L3 AR-1232-1	5.122	4.436	6818678	7080205	212.665	196.830
12)	L3 AR-1232-2	5.655	5.183	9291200	17182167	521.994	518.667
13)	L3 AR-1232-3	5.943	5.363	16733952	9831487	529.345	547.234
14)	L3 AR-1232-4	6.104	5.446	9076756	10108879	562.486	578.022
15)	L3 AR-1232-5	6.194	5.622	8012451	10397788	592.257	553.776
16)	L4 AR-1242-1	5.920	5.163	12239216	12895691	344.160	336.298
17)	L4 AR-1242-2	5.943	5.183	16733952	17182167	319.842	328.102
18)	L4 AR-1242-3	6.006	5.363	11350644	9831487	333.047	330.796
19)	L4 AR-1242-4	6.104	5.446	9076756	10108879	330.955	314.148
20)	L4 AR-1242-5	6.840	5.979	1561173	8527342	52.456	223.356 #
21)	L5 AR-1248-1	5.920	5.163	12239216	12895691	433.476	421.072
22)	L5 AR-1248-2	6.194	5.402	8012451	8473000	180.481	186.763
23)	L5 AR-1248-3	6.399	5.446	9051507	10108879	186.536	213.557
24)	L5 AR-1248-4	6.775	5.622	8487340	10397788	156.421	191.460
25)	L5 AR-1248-5	6.840	6.019	1561173	1188486	28.311	24.385
26)	L6 AR-1254-1	6.775	5.979	8487340	8527342	136.226	105.196
27)	L6 AR-1254-2	6.992	6.125	10479307	8359895	115.494	113.984
28)	L6 AR-1254-3	7.358	6.550	4802866	18055484	51.204	158.355 #
29)	L6 AR-1254-4	7.642	6.763	2964091	1317375	44.477	21.107 #
30)	L6 AR-1254-5	8.058	7.184	26839439	27882332	326.530	271.574
31)	L7 AR-1260-1	7.523	6.667	19132909	20607272	242.857	251.912
32)	L7 AR-1260-2	7.776	6.853	22758984	24624998	249.699	262.123
33)	L7 AR-1260-3	8.138	7.011	14412985	23426785	228.161	258.587
34)	L7 AR-1260-4	8.376	7.485	22503862	16731614	298.043	250.922
35)	L7 AR-1260-5	8.706	7.724	37109164	37752650	286.612	259.316
36)	L8 AR-1262-1	8.376	7.222	22503862	17451136	236.017	165.919 #
37)	L8 AR-1262-2	8.706	7.485	37109164	16731614	228.718	176.725
38)	L8 AR-1262-3	9.061	8.011	24834674	7109134	211.395	96.919 #
39)	L8 AR-1262-4	9.127	8.074	13956761	28039319	150.906	212.744 #
40)	L8 AR-1262-5	9.847	8.602	9133395	9354698	149.183	153.963
41)	L9 AR-1268-1	9.061	8.011	24834674	7109134	128.991	35.297 #

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP100724\
Data File : PP067549.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 07 Oct 2024 15:00
Operator : YP\AJ
Sample : P4324-01MS
Misc :
ALS Vial : 14 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
LAW-24-00169-00170MS

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 08 02:08:36 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP092624.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Sep 27 06:59:05 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
42)	L9 AR-1268-2	9.127	8.074	13956761	28039319	81.814	156.761 #
43)	L9 AR-1268-3	9.397	8.296	1114952	601604	7.405	3.787 #
44)	L9 AR-1268-4	9.847	8.602	9133395	9354698	150.092	144.889
45)	L9 AR-1268-5	10.298	8.930	3129802	3889517	7.244	8.804



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP100724\
Data File : PP067549.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 07 Oct 2024 15:00
Operator : YP\AJ
Sample : P4324-01MS
Misc :
ALS Vial : 14 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
LAW-24-00169-00170MS

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 08 02:08:36 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP092624.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Sep 27 06:59:05 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.756 min
Delta R.T.: 0.003 min
Response: 13431541
Conc: 9.99 ng/ml

Instrument :
ECD_P
ClientSampleId :
LAW-24-00169-00170MS




#1 Tetrachloro-m-xylene

R.T.: 4.054 min
Delta R.T.: 0.001 min
Response: 14283984
Conc: 9.56 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.671 min
Delta R.T.: 0.010 min
Response: 17112897
Conc: 11.79 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.222 min
Delta R.T.: -0.004 min
Response: 17782059
Conc: 12.90 ng/ml



#3 AR-1016-1


R.T.: 5.920 min
Delta R.T.: 0.003 min
Response: 12239216
Conc: 256.70 ng/ml

Instrument :
ECD_P
ClientSampleId :
LAW-24-00169-00170MS




#3 AR-1016-1

R.T.: 5.163 min
Delta R.T.: -0.001 min
Response: 12895691
Conc: 250.02 ng/ml




#4 AR-1016-2

R.T.: 5.943 min
Delta R.T.: 0.003 min
Response: 16733952
Conc: 241.67 ng/ml



#4 AR-1016-2


R.T.: 5.183 min
Delta R.T.: -0.001 min
Response: 17182167
Conc: 241.65 ng/ml



#5 AR-1016-3


R.T.: 6.006 min
Delta R.T.: 0.003 min
Response: 11350644
Conc: 254.10 ng/ml

Instrument :
ECD_P
ClientSampleId :
LAW-24-00169-00170MS




#5 AR-1016-3

R.T.: 5.363 min
Delta R.T.: -0.001 min
Response: 9831487
Conc: 247.04 ng/ml




#6 AR-1016-4

R.T.: 6.104 min
Delta R.T.: 0.003 min
Response: 9076756
Conc: 252.59 ng/ml



#6 AR-1016-4


R.T.: 5.402 min
Delta R.T.: -0.002 min
Response: 8473000
Conc: 245.39 ng/ml



#7 AR-1016-5


R.T.: 6.399 min
Delta R.T.: 0.003 min
Response: 9051507
Conc: 245.63 ng/ml

Instrument :
ECD_P
ClientSampleId :
LAW-24-00169-00170MS




#7 AR-1016-5

R.T.: 5.622 min
Delta R.T.: -0.002 min
Response: 10397788
Conc: 242.32 ng/ml




#8 AR-1221-1

R.T.: 4.955 min
Delta R.T.: 0.000 min
Response: 1359664
Conc: 81.22 ng/ml



#8 AR-1221-1

R.T.: 4.266 min
Delta R.T.: -0.003 min
Response: 1093217
Conc: 57.18 ng/ml



#9 AR-1221-2

R.T.: 5.045 min
Delta R.T.: 0.004 min
Response: 1504769
Conc: 118.77 ng/ml

Instrument :
ECD_P
ClientSampleId :
LAW-24-00169-00170MS

#9 AR-1221-2


R.T.: 4.357 min
Delta R.T.: 0.000 min
Response: 1614396
Conc: 110.04 ng/ml

#10 AR-1221-3

R.T.: 5.122 min
Delta R.T.: 0.004 min
Response: 6818678
Conc: 176.19 ng/ml

#10 AR-1221-3


R.T.: 4.436 min
Delta R.T.: 0.000 min
Response: 7080205
Conc: 163.27 ng/ml



#11 AR-1232-1


R.T.: 5.122 min
Delta R.T.: 0.005 min
Response: 6818678
Conc: 212.67 ng/ml

Instrument :
ECD_P
ClientSampleId :
LAW-24-00169-00170MS




#11 AR-1232-1

R.T.: 4.436 min
Delta R.T.: 0.000 min
Response: 7080205
Conc: 196.83 ng/ml




#12 AR-1232-2

R.T.: 5.655 min
Delta R.T.: 0.004 min
Response: 9291200
Conc: 521.99 ng/ml



#12 AR-1232-2


R.T.: 5.183 min
Delta R.T.: 0.000 min
Response: 17182167
Conc: 518.67 ng/ml



#13 AR-1232-3


R.T.: 5.943 min
Delta R.T.: 0.004 min
Response: 16733952
Conc: 529.34 ng/ml

Instrument :
ECD_P
ClientSampleId :
LAW-24-00169-00170MS




#13 AR-1232-3

R.T.: 5.363 min
Delta R.T.: 0.000 min
Response: 9831487
Conc: 547.23 ng/ml




#14 AR-1232-4

R.T.: 6.104 min
Delta R.T.: 0.004 min
Response: 9076756
Conc: 562.49 ng/ml



#14 AR-1232-4


R.T.: 5.446 min
Delta R.T.: 0.000 min
Response: 10108879
Conc: 578.02 ng/ml



#15 AR-1232-5


R.T.: 6.194 min
Delta R.T.: 0.004 min
Response: 8012451
Conc: 592.26 ng/ml

Instrument :
ECD_P
ClientSampleId :
LAW-24-00169-00170MS




#15 AR-1232-5

R.T.: 5.622 min
Delta R.T.: 0.000 min
Response: 10397788
Conc: 553.78 ng/ml




#16 AR-1242-1

R.T.: 5.920 min
Delta R.T.: 0.002 min
Response: 12239216
Conc: 344.16 ng/ml



#16 AR-1242-1


R.T.: 5.163 min
Delta R.T.: -0.001 min
Response: 12895691
Conc: 336.30 ng/ml



#17 AR-1242-2


R.T.: 5.943 min
Delta R.T.: 0.002 min
Response: 16733952
Conc: 319.84 ng/ml

Instrument :
ECD_P
ClientSampleId :
LAW-24-00169-00170MS




#17 AR-1242-2

R.T.: 5.183 min
Delta R.T.: -0.001 min
Response: 17182167
Conc: 328.10 ng/ml




#18 AR-1242-3

R.T.: 6.006 min
Delta R.T.: 0.002 min
Response: 11350644
Conc: 333.05 ng/ml



#18 AR-1242-3


R.T.: 5.363 min
Delta R.T.: -0.002 min
Response: 9831487
Conc: 330.80 ng/ml



#19 AR-1242-4


R.T.: 6.104 min
Delta R.T.: 0.003 min
Response: 9076756
Conc: 330.96 ng/ml

Instrument :
ECD_P
ClientSampleId :
LAW-24-00169-00170MS




#19 AR-1242-4

R.T.: 5.446 min
Delta R.T.: -0.001 min
Response: 10108879
Conc: 314.15 ng/ml




#20 AR-1242-5

R.T.: 6.840 min
Delta R.T.: 0.004 min
Response: 1561173
Conc: 52.46 ng/ml



#20 AR-1242-5


R.T.: 5.979 min
Delta R.T.: 0.000 min
Response: 8527342
Conc: 223.36 ng/ml



#21 AR-1248-1


R.T.: 5.920 min
Delta R.T.: 0.004 min
Response: 12239216
Conc: 433.48 ng/ml

Instrument :
ECD_P
ClientSampleId :
LAW-24-00169-00170MS




#21 AR-1248-1

R.T.: 5.163 min
Delta R.T.: 0.000 min
Response: 12895691
Conc: 421.07 ng/ml




#22 AR-1248-2

R.T.: 6.194 min
Delta R.T.: 0.004 min
Response: 8012451
Conc: 180.48 ng/ml



#22 AR-1248-2


R.T.: 5.402 min
Delta R.T.: -0.001 min
Response: 8473000
Conc: 186.76 ng/ml



#23 AR-1248-3


R.T.: 6.399 min
Delta R.T.: 0.004 min
Response: 9051507
Conc: 186.54 ng/ml

Instrument :
ECD_P
ClientSampleId :
LAW-24-00169-00170MS




#23 AR-1248-3

R.T.: 5.446 min
Delta R.T.: 0.000 min
Response: 10108879
Conc: 213.56 ng/ml




#24 AR-1248-4

R.T.: 6.775 min
Delta R.T.: -0.019 min
Response: 8487340
Conc: 156.42 ng/ml



#24 AR-1248-4


R.T.: 5.622 min
Delta R.T.: 0.000 min
Response: 10397788
Conc: 191.46 ng/ml



#25 AR-1248-5


R.T.: 6.840 min
Delta R.T.: 0.005 min
Response: 1561173
Conc: 28.31 ng/ml

Instrument :
ECD_P
ClientSampleId :
LAW-24-00169-00170MS




#25 AR-1248-5

R.T.: 6.019 min
Delta R.T.: -0.001 min
Response: 1188486
Conc: 24.38 ng/ml




#26 AR-1254-1

R.T.: 6.775 min
Delta R.T.: 0.005 min
Response: 8487340
Conc: 136.23 ng/ml



#26 AR-1254-1


R.T.: 5.979 min
Delta R.T.: -0.002 min
Response: 8527342
Conc: 105.20 ng/ml



#27 AR-1254-2


R.T.: 6.992 min
Delta R.T.: 0.004 min
Response: 10479307
Conc: 115.49 ng/ml

Instrument :
ECD_P
ClientSampleId :
LAW-24-00169-00170MS




#27 AR-1254-2

R.T.: 6.125 min
Delta R.T.: -0.002 min
Response: 8359895
Conc: 113.98 ng/ml




#28 AR-1254-3

R.T.: 7.358 min
Delta R.T.: 0.006 min
Response: 4802866
Conc: 51.20 ng/ml



#28 AR-1254-3


R.T.: 6.550 min
Delta R.T.: 0.012 min
Response: 18055484
Conc: 158.36 ng/ml



#29 AR-1254-4


R.T.: 7.642 min
Delta R.T.: 0.007 min
Response: 2964091
Conc: 44.48 ng/ml

Instrument :
ECD_P
ClientSampleId :
LAW-24-00169-00170MS




#29 AR-1254-4

R.T.: 6.763 min
Delta R.T.: -0.002 min
Response: 1317375
Conc: 21.11 ng/ml




#30 AR-1254-5

R.T.: 8.058 min
Delta R.T.: 0.005 min
Response: 26839439
Conc: 326.53 ng/ml



#30 AR-1254-5


R.T.: 7.184 min
Delta R.T.: -0.003 min
Response: 27882332
Conc: 271.57 ng/ml



#31 AR-1260-1


R.T.: 7.523 min
Delta R.T.: 0.004 min
Response: 19132909
Conc: 242.86 ng/ml

Instrument :
ECD_P
ClientSampleId :
LAW-24-00169-00170MS




#31 AR-1260-1

R.T.: 6.667 min
Delta R.T.: -0.002 min
Response: 20607272
Conc: 251.91 ng/ml




#32 AR-1260-2

R.T.: 7.776 min
Delta R.T.: 0.004 min
Response: 22758984
Conc: 249.70 ng/ml



#32 AR-1260-2


R.T.: 6.853 min
Delta R.T.: -0.002 min
Response: 24624998
Conc: 262.12 ng/ml



#33 AR-1260-3


R.T.: 8.138 min
Delta R.T.: 0.005 min
Response: 14412985
Conc: 228.16 ng/ml

Instrument :
ECD_P
ClientSampleId :
LAW-24-00169-00170MS




#33 AR-1260-3

R.T.: 7.011 min
Delta R.T.: -0.002 min
Response: 23426785
Conc: 258.59 ng/ml




#34 AR-1260-4

R.T.: 8.376 min
Delta R.T.: 0.004 min
Response: 22503862
Conc: 298.04 ng/ml



#34 AR-1260-4


R.T.: 7.485 min
Delta R.T.: -0.003 min
Response: 16731614
Conc: 250.92 ng/ml



#35 AR-1260-5


R.T.: 8.706 min
Delta R.T.: -0.002 min
Response: 37109164
Conc: 286.61 ng/ml

Instrument :
ECD_P
ClientSampleId :
LAW-24-00169-00170MS




#35 AR-1260-5

R.T.: 7.724 min
Delta R.T.: -0.002 min
Response: 37752650
Conc: 259.32 ng/ml




#36 AR-1262-1

R.T.: 8.376 min
Delta R.T.: 0.004 min
Response: 22503862
Conc: 236.02 ng/ml



#36 AR-1262-1


R.T.: 7.222 min
Delta R.T.: -0.003 min
Response: 17451136
Conc: 165.92 ng/ml



#37 AR-1262-2


R.T.: 8.706 min
Delta R.T.: -0.002 min
Response: 37109164
Conc: 228.72 ng/ml

Instrument :
ECD_P
ClientSampleId :
LAW-24-00169-00170MS




#37 AR-1262-2

R.T.: 7.485 min
Delta R.T.: -0.003 min
Response: 16731614
Conc: 176.72 ng/ml




#38 AR-1262-3

R.T.: 9.061 min
Delta R.T.: 0.015 min
Response: 24834674
Conc: 211.40 ng/ml



#38 AR-1262-3


R.T.: 8.011 min
Delta R.T.: -0.003 min
Response: 7109134
Conc: 96.92 ng/ml



#39 AR-1262-4


R.T.: 9.127 min
Delta R.T.: -0.011 min
Response: 13956761
Conc: 150.91 ng/ml

Instrument :
ECD_P
ClientSampleId :
LAW-24-00169-00170MS




#39 AR-1262-4

R.T.: 8.074 min
Delta R.T.: -0.004 min
Response: 28039319
Conc: 212.74 ng/ml




#40 AR-1262-5

R.T.: 9.847 min
Delta R.T.: 0.008 min
Response: 9133395
Conc: 149.18 ng/ml



#40 AR-1262-5


R.T.: 8.602 min
Delta R.T.: -0.003 min
Response: 9354698
Conc: 153.96 ng/ml



#41 AR-1268-1


R.T.: 9.061 min
Delta R.T.: 0.019 min
Response: 24834674
Conc: 128.99 ng/ml

Instrument :
ECD_P
ClientSampleId :
LAW-24-00169-00170MS




#41 AR-1268-1

R.T.: 8.011 min
Delta R.T.: -0.003 min
Response: 7109134
Conc: 35.30 ng/ml




#42 AR-1268-2

R.T.: 9.127 min
Delta R.T.: -0.014 min
Response: 13956761
Conc: 81.81 ng/ml



#42 AR-1268-2


R.T.: 8.074 min
Delta R.T.: -0.004 min
Response: 28039319
Conc: 156.76 ng/ml



#43 AR-1268-3


R.T.: 9.397 min
Delta R.T.: 0.004 min
Response: 1114952
Conc: 7.41 ng/ml

Instrument :
ECD_P
ClientSampleId :
LAW-24-00169-00170MS




#43 AR-1268-3

R.T.: 8.296 min
Delta R.T.: -0.003 min
Response: 601604
Conc: 3.79 ng/ml




#44 AR-1268-4

R.T.: 9.847 min
Delta R.T.: 0.009 min
Response: 9133395
Conc: 150.09 ng/ml



#44 AR-1268-4


R.T.: 8.602 min
Delta R.T.: -0.004 min
Response: 9354698
Conc: 144.89 ng/ml



#45 AR-1268-5

R.T.: 10.298 min
Delta R.T.: 0.009 min
Response: 3129802
Conc: 7.24 ng/ml

Instrument :
ECD_P
ClientSampleId :
LAW-24-00169-00170MS



#45 AR-1268-5

R.T.: 8.930 min
Delta R.T.: -0.004 min
Response: 3889517
Conc: 8.80 ng/ml