

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP100724\
 Data File : PP067560.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 07 Oct 2024 18:00
 Operator : YP\AJ
 Sample : PB163944BL
 Misc :
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 PB163944BL

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 10/08/2024
 Supervised By :Ankita Jodhani 10/08/2024

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 08 02:26:20 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP092624.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Sep 27 06:59:05 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.753	4.052	19040339	20287376	14.166	13.578
2) SA Decachlor...	10.665	9.219	25168355	24152342	17.337	17.527m

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP100724\
Data File : PP067560.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 07 Oct 2024 18:00
Operator : YP\AJ
Sample : PB163944BL
Misc :
ALS Vial : 20 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
PB163944BL


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/08/2024
Supervised By :Ankita Jodhani 10/08/2024

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 08 02:26:20 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP092624.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Sep 27 06:59:05 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.753 min
Delta R.T.: 0.000 min
Response: 19040339
Conc: 14.17 ng/ml

Instrument :
ECD_P
ClientSampleId :
PB163944BL


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/08/2024
Supervised By :Ankita Jodhani 10/08/2024




#1 Tetrachloro-m-xylene

R.T.: 4.052 min
Delta R.T.: 0.000 min
Response: 20287376
Conc: 13.58 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.665 min
Delta R.T.: 0.004 min
Response: 25168355
Conc: 17.34 ng/ml



#2 Decachlorobiphenyl

R.T.: 9.219 min
Delta R.T.: -0.007 min
Response: 24152342
Conc: 17.53 ng/ml m