

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP100824\
 Data File : PP067597.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 08 Oct 2024 19:12
 Operator : YP\AJ
 Sample : AR1242ICC250
 Misc :
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1242ICC250

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 09 02:27:50 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP100824.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Oct 09 02:22:02 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.755	4.051	24714146	26992640	24.961	25.746
2) SA Decachlor...	10.666	9.218	31835407	30686591	26.897	26.819
Target Compounds						
16) L4 AR-1242-1	5.919	5.160	7532960	7966048	260.835	264.807
17) L4 AR-1242-2	5.942	5.181	10856012	10764640	259.569	263.482
18) L4 AR-1242-3	6.005	5.361	7003429	5902002	253.500	252.337
19) L4 AR-1242-4	6.103	5.444	5549729	6325659	252.030	256.788
20) L4 AR-1242-5	6.837	5.974	6362457	7698031	259.752	266.737



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP100824\
Data File : PP067597.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 08 Oct 2024 19:12
Operator : YP\AJ
Sample : AR1242ICC250
Misc :
ALS Vial : 13 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1242ICC250

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 09 02:27:50 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP100824.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Oct 09 02:22:02 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.755 min
Delta R.T.: 0.003 min
Response: 24714146
Conc: 24.96 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242ICC250




#1 Tetrachloro-m-xylene

R.T.: 4.051 min
Delta R.T.: 0.000 min
Response: 26992640
Conc: 25.75 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.666 min
Delta R.T.: 0.000 min
Response: 31835407
Conc: 26.90 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.218 min
Delta R.T.: 0.000 min
Response: 30686591
Conc: 26.82 ng/ml



#16 AR-1242-1


R.T.: 5.919 min
Delta R.T.: 0.003 min
Response: 7532960
Conc: 260.83 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242ICC250




#16 AR-1242-1

R.T.: 5.160 min
Delta R.T.: 0.000 min
Response: 7966048
Conc: 264.81 ng/ml




#17 AR-1242-2

R.T.: 5.942 min
Delta R.T.: 0.002 min
Response: 10856012
Conc: 259.57 ng/ml



#17 AR-1242-2


R.T.: 5.181 min
Delta R.T.: 0.000 min
Response: 10764640
Conc: 263.48 ng/ml



#18 AR-1242-3


R.T.: 6.005 min
Delta R.T.: 0.003 min
Response: 7003429
Conc: 253.50 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242ICC250




#18 AR-1242-3

R.T.: 5.361 min
Delta R.T.: 0.000 min
Response: 5902002
Conc: 252.34 ng/ml




#19 AR-1242-4

R.T.: 6.103 min
Delta R.T.: 0.002 min
Response: 5549729
Conc: 252.03 ng/ml



#19 AR-1242-4


R.T.: 5.444 min
Delta R.T.: 0.000 min
Response: 6325659
Conc: 256.79 ng/ml



#20 AR-1242-5

R.T.: 6.837 min
Delta R.T.: 0.002 min
Response: 6362457
Conc: 259.75 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242ICC250



#20 AR-1242-5

R.T.: 5.974 min
Delta R.T.: -0.001 min
Response: 7698031
Conc: 266.74 ng/ml