

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP100824\
 Data File : PP067610.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 08 Oct 2024 22:42
 Operator : YP\AJ
 Sample : AR1268ICC1000
 Misc :
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1268ICC1000

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 10/09/2024
 Supervised By :Ankita Jodhani 10/09/2024

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 09 04:57:52 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP100824.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Oct 09 04:46:09 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.755	4.051	90812575	97016985	95.024	96.337
2) SA Decachlor...	10.666	9.217	175.8E6	175.5E6	90.038	92.883
Target Compounds						
41) L9 AR-1268-1	9.045	8.006	139.5E6	147.0E6	925.475m	929.738m
42) L9 AR-1268-2	9.145	8.072	125.7E6	133.1E6	920.211	933.699
43) L9 AR-1268-3	9.396	8.291	110.7E6	117.4E6	925.055	927.652
44) L9 AR-1268-4	9.844	8.599	48063749	51956547	911.676	948.161
45) L9 AR-1268-5	10.295	8.926	348.4E6	359.3E6	936.060	959.229

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP100824\
Data File : PP067610.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 08 Oct 2024 22:42
Operator : YP\AJ
Sample : AR1268ICC1000
Misc :
ALS Vial : 26 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
AR1268ICC1000


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/09/2024
Supervised By :Ankita Jodhani 10/09/2024

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 09 04:57:52 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP100824.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Oct 09 04:46:09 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.755 min
Delta R.T.: 0.002 min
Response: 90812575
Conc: 95.02 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1268ICC1000


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/09/2024
Supervised By :Ankita Jodhani 10/09/2024




#1 Tetrachloro-m-xylene

R.T.: 4.051 min
Delta R.T.: 0.000 min
Response: 97016985
Conc: 96.34 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.666 min
Delta R.T.: 0.001 min
Response: 175837642
Conc: 90.04 ng/ml



#2 Decachlorobiphenyl

R.T.: 9.217 min
Delta R.T.: 0.000 min
Response: 175518672
Conc: 92.88 ng/ml




#41 AR-1268-1

R.T.: 9.045 min
Delta R.T.: 0.002 min
Response: 139515419
Conc: 925.48 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1268ICC1000


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/09/2024
Supervised By :Ankita Jodhani 10/09/2024




#41 AR-1268-1

R.T.: 8.006 min
Delta R.T.: -0.001 min
Response: 146986557
Conc: 929.74 ng/ml




#42 AR-1268-2

R.T.: 9.145 min
Delta R.T.: 0.000 min
Response: 125679796
Conc: 920.21 ng/ml



#42 AR-1268-2

R.T.: 8.072 min
Delta R.T.: -0.001 min
Response: 133073925
Conc: 933.70 ng/ml




#43 AR-1268-3

R.T.: 9.396 min
Delta R.T.: 0.002 min
Response: 110668977
Conc: 925.06 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1268ICC1000


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/09/2024
Supervised By :Ankita Jodhani 10/09/2024




#43 AR-1268-3

R.T.: 8.291 min
Delta R.T.: 0.000 min
Response: 117364027
Conc: 927.65 ng/ml




#44 AR-1268-4

R.T.: 9.844 min
Delta R.T.: 0.001 min
Response: 48063749
Conc: 911.68 ng/ml



#44 AR-1268-4

R.T.: 8.599 min
Delta R.T.: 0.000 min
Response: 51956547
Conc: 948.16 ng/ml




#45 AR-1268-5

R.T.: 10.295 min
Delta R.T.: 0.000 min
Response: 348373365
Conc: 936.06 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1268ICC1000

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/09/2024
Supervised By :Ankita Jodhani 10/09/2024



#45 AR-1268-5

R.T.: 8.926 min
Delta R.T.: 0.000 min
Response: 359344903
Conc: 959.23 ng/ml