

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP100824\
 Data File : PP067618.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 09 Oct 2024 00:51
 Operator : YP\AJ
 Sample : AR1254ICV500
 Misc :
 ALS Vial : 34 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 ICVPP100824AR1254

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 10/09/2024
 Supervised By :Ankita Jodhani 10/09/2024

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 09 03:43:22 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP100824.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Oct 09 03:24:30 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.756	4.051	48530224	52089649	51.723	51.478
2) SA Decachlor...	10.668	9.217	60024597	58780254	50.717	50.845
Target Compounds						
26) L6 AR-1254-1	6.775	5.975	22815828	29677404	509.481	502.293
27) L6 AR-1254-2	6.991	6.123	33399032	26325396	505.298	501.778
28) L6 AR-1254-3	7.357	6.532	34711516	42198184	499.786	507.141
29) L6 AR-1254-4	7.639	6.760	25405512	24704816	509.082	511.760m
30) L6 AR-1254-5	8.057	7.182	29963770	37426268	505.560m	508.846

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP100824\
Data File : PP067618.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 09 Oct 2024 00:51
Operator : YP\AJ
Sample : AR1254ICV500
Misc :
ALS Vial : 34 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
ICVPP100824AR1254


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/09/2024
Supervised By :Ankita Jodhani 10/09/2024

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 09 03:43:22 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP100824.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Oct 09 03:24:30 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm





#1 Tetrachloro-m-xylene

R.T.: 4.756 min
Delta R.T.: 0.002 min
Response: 48530224
Conc: 51.72 ng/ml

Instrument :


ECD_P

ClientSampleId :

ICVPP100824AR1254


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/09/2024
Supervised By :Ankita Jodhani 10/09/2024




#1 Tetrachloro-m-xylene

R.T.: 4.051 min
Delta R.T.: 0.000 min
Response: 52089649
Conc: 51.48 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.668 min
Delta R.T.: 0.004 min
Response: 60024597
Conc: 50.72 ng/ml



#2 Decachlorobiphenyl

R.T.: 9.217 min
Delta R.T.: 0.000 min
Response: 58780254
Conc: 50.85 ng/ml




#26 AR-1254-1

R.T.: 6.775 min
Delta R.T.: 0.002 min
Response: 22815828
Conc: 509.48 ng/ml

Instrument :
ECD_P
ClientSampleId :
ICVPP100824AR1254


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/09/2024
Supervised By :Ankita Jodhani 10/09/2024




#26 AR-1254-1

R.T.: 5.975 min
Delta R.T.: 0.000 min
Response: 29677404
Conc: 502.29 ng/ml




#27 AR-1254-2

R.T.: 6.991 min
Delta R.T.: 0.002 min
Response: 33399032
Conc: 505.30 ng/ml



#27 AR-1254-2

R.T.: 6.123 min
Delta R.T.: 0.000 min
Response: 26325396
Conc: 501.78 ng/ml




#28 AR-1254-3

R.T.: 7.357 min
Delta R.T.: 0.002 min
Response: 34711516
Conc: 499.79 ng/ml

Instrument :
ECD_P
ClientSampleId :
ICVPP100824AR1254


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/09/2024
Supervised By :Ankita Jodhani 10/09/2024




#28 AR-1254-3

R.T.: 6.532 min
Delta R.T.: 0.000 min
Response: 42198184
Conc: 507.14 ng/ml




#29 AR-1254-4

R.T.: 7.639 min
Delta R.T.: 0.002 min
Response: 25405512
Conc: 509.08 ng/ml



#29 AR-1254-4

R.T.: 6.760 min
Delta R.T.: 0.000 min
Response: 24704816
Conc: 511.76 ng/ml m




#30 AR-1254-5

R.T.: 8.057 min
Delta R.T.: 0.002 min
Response: 29963770
Conc: 505.56 ng/ml

Instrument :
ECD_P
ClientSampleId :
ICVPP100824AR1254

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/09/2024
Supervised By :Ankita Jodhani 10/09/2024



#30 AR-1254-5

R.T.: 7.182 min
Delta R.T.: 0.000 min
Response: 37426268
Conc: 508.85 ng/ml