

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP100918\
 Data File : PP027655.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 09 Oct 2018 19:15
 Operator : EI\MA
 Sample : HSTDICC750
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 10 09:13:29 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP100918.M
 Quant Title : 8080.M
 QLast Update : Wed Oct 10 09:10:34 2018
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-CLPesticides-1 Signal #2 Phase: ZB-CLPesticides-2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	12.046	12.159	191.2E6	322.8E6	750.000	750.000
Target Compounds							
1) T	Dalapon	3.312	2.984	420.8E6	371.8E6	682.500	682.500
2) T	3,5-DICHL...	10.460	10.252	339.8E6	366.4E6	697.500	697.500
3) T	4-Nitroph...	11.622	11.304	276.2E6	130.6E6	682.500	682.500
5) T	DICAMBA	12.367	12.469	1104.2E6	1174.7E6	705.000	705.000
6) T	MCPD	12.709	12.662	32949245	62342527	70.500	70.500
7) T	MCPA	12.953	13.051	48005982	97519879	69.750	69.750
8) T	DICHLORPROP	13.564	13.585	259.8E6	291.9E6	705.000	705.000
9) T	2,4-D	13.914	14.071	256.6E6	354.6E6	705.000	705.000
10) T	Pentachlo...	14.331	14.681	5804.8E6	4723.9E6	712.500	712.500
11) T	2,4,5-TP ...	15.193	15.252	2077.4E6	1800.7E6	712.500	712.500
12) T	2,4,5-T	15.583	15.892	1662.8E6	1250.0E6	712.500	712.500
13) T	2,4-DB	16.346	16.497	109.5E6	143.2E6	712.500	712.500
14) T	DINOSEB	17.834	16.886	1427.1E6	1181.1E6	705.000	705.000
15) T	Picloram	17.596	18.267	1581.5E6	1276.1E6	712.500	712.500
16) T	DCPA	18.197	18.146	3072.6E6	2283.5E6	720.000	720.000



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP100918\
Data File : PP027655.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 09 Oct 2018 19:15
Operator : EI\MA
Sample : HSTDICC750
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampled :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 10 09:13:29 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP100918.M
Quant Title : 8080.M
QLast Update : Wed Oct 10 09:10:34 2018
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-CLPesticides-1 Signal #2 Phase: ZB-CLPesticides-2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Dalapon


R.T.: 3.312 min
Delta R.T.: 0.000 min
Response: 420808526
Conc: 682.50 ng/ml

Instrument :
ECD_P
ClientSampleId :




#1 Dalapon

R.T.: 2.984 min
Delta R.T.: 0.000 min
Response: 371794177
Conc: 682.50 ng/ml




#2 3,5-DICHLOROBENZOIC ACID

R.T.: 10.460 min
Delta R.T.: 0.000 min
Response: 339767295
Conc: 697.50 ng/ml



#2 3,5-DICHLOROBENZOIC ACID


R.T.: 10.252 min
Delta R.T.: 0.000 min
Response: 366353669
Conc: 697.50 ng/ml



#3 4-Nitrophenol


R.T.: 11.622 min
Delta R.T.: 0.000 min
Response: 276188300
Conc: 682.50 ng/ml

Instrument :
ECD_P
ClientSampleId :




#3 4-Nitrophenol

R.T.: 11.304 min
Delta R.T.: 0.000 min
Response: 130578450
Conc: 682.50 ng/ml




#4 2,4-DCAA

R.T.: 12.046 min
Delta R.T.: 0.000 min
Response: 191190343
Conc: 750.00 ng/ml



#4 2,4-DCAA


R.T.: 12.159 min
Delta R.T.: 0.000 min
Response: 322767933
Conc: 750.00 ng/ml



#5 DICAMBA


R.T.: 12.367 min
Delta R.T.: 0.000 min
Response: 1104212102
Conc: 705.00 ng/ml

Instrument :
ECD_P
ClientSampleId :




#5 DICAMBA

R.T.: 12.469 min
Delta R.T.: 0.000 min
Response: 1174729155
Conc: 705.00 ng/ml




#6 MCP

R.T.: 12.709 min
Delta R.T.: 0.000 min
Response: 32949245
Conc: 70.50 ug/ml



#6 MCP


R.T.: 12.662 min
Delta R.T.: 0.000 min
Response: 62342527
Conc: 70.50 ug/ml



#7 MCPA


R.T.: 12.953 min
Delta R.T.: 0.000 min
Response: 48005982
Conc: 69.75 ug/ml

Instrument :
ECD_P
ClientSampleId :




#7 MCPA

R.T.: 13.051 min
Delta R.T.: 0.000 min
Response: 97519879
Conc: 69.75 ug/ml




#8 DICHLORPROP

R.T.: 13.564 min
Delta R.T.: 0.000 min
Response: 259825752
Conc: 705.00 ng/ml



#8 DICHLORPROP


R.T.: 13.585 min
Delta R.T.: 0.000 min
Response: 291935707
Conc: 705.00 ng/ml



#9 2,4-D


R.T.: 13.914 min
Delta R.T.: 0.000 min
Response: 256587766
Conc: 705.00 ng/ml

Instrument :
ECD_P
ClientSampleId :




#9 2,4-D

R.T.: 14.071 min
Delta R.T.: 0.000 min
Response: 354597229
Conc: 705.00 ng/ml




#10 Pentachlorophenol

R.T.: 14.331 min
Delta R.T.: 0.000 min
Response: 5804786919
Conc: 712.50 ng/ml



#10 Pentachlorophenol


R.T.: 14.681 min
Delta R.T.: 0.000 min
Response: 4723898876
Conc: 712.50 ng/ml



#11 2,4,5-TP (SILVEX)


R.T.: 15.193 min
Delta R.T.: 0.000 min
Response: 2077350656
Conc: 712.50 ng/ml

Instrument :
ECD_P
ClientSampleId :




#11 2,4,5-TP (SILVEX)

R.T.: 15.252 min
Delta R.T.: 0.000 min
Response: 1800737781
Conc: 712.50 ng/ml




#12 2,4,5-T

R.T.: 15.583 min
Delta R.T.: 0.000 min
Response: 1662809559
Conc: 712.50 ng/ml



#12 2,4,5-T


R.T.: 15.892 min
Delta R.T.: 0.000 min
Response: 1249994807
Conc: 712.50 ng/ml



#13 2,4-DB


R.T.: 16.346 min
Delta R.T.: 0.000 min
Response: 109474690
Conc: 712.50 ng/ml

Instrument :
ECD_P
ClientSampleId :




#13 2,4-DB

R.T.: 16.497 min
Delta R.T.: 0.000 min
Response: 143236487
Conc: 712.50 ng/ml




#14 DINOSEB

R.T.: 17.834 min
Delta R.T.: 0.000 min
Response: 1427110713
Conc: 705.00 ng/ml



#14 DINOSEB


R.T.: 16.886 min
Delta R.T.: 0.000 min
Response: 1181084275
Conc: 705.00 ng/ml



#15 Picloram


R.T.: 17.596 min
Delta R.T.: 0.000 min
Response: 1581488759
Conc: 712.50 ng/ml

Instrument :
ECD_P
ClientSampleId :




#15 Picloram

R.T.: 18.267 min
Delta R.T.: 0.000 min
Response: 1276129006
Conc: 712.50 ng/ml



#16 DCPA

R.T.: 18.197 min
Delta R.T.: 0.000 min
Response: 3072574173
Conc: 720.00 ng/ml



#16 DCPA

R.T.: 18.146 min
Delta R.T.: 0.000 min
Response: 2283515544
Conc: 720.00 ng/ml