

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP100918\
 Data File : PP027656.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 09 Oct 2018 19:40
 Operator : EI\MA
 Sample : HSTDICC1000
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 HSTDICC1000

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 10 09:13:50 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP100918.M
 Quant Title : 8080.M
 QLast Update : Wed Oct 10 09:10:34 2018
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-CLPesticides-1 Signal #2 Phase: ZB-CLPesticides-2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	12.045	12.159	246.3E6	417.5E6	966.242	970.071
Target Compounds							
1) T	Dalapon	3.312	2.984	543.3E6	483.1E6	881.110	886.898
2) T	3,5-DICHL...	10.459	10.251	435.8E6	472.0E6	894.589	898.632
3) T	4-Nitroph...	11.620	11.303	367.2E6	176.6E6	907.310	922.893
5) T	DICAMBA	12.366	12.469	1417.8E6	1501.7E6	905.208	901.203
6) T	MCPD	12.710	12.663	44035285	81416542	94.220	92.070
7) T	MCPA	12.955	13.052	68009107	130.0E6	98.813	92.951
8) T	DICHLORPROP	13.563	13.583	333.5E6	375.3E6	904.937	906.315
9) T	2,4-D	13.912	14.070	337.7E6	465.8E6	927.890	926.067
10) T	Pentachlo...	14.330	14.680	7273.3E6	5864.2E6	892.756	884.484
11) T	2,4,5-TP ...	15.192	15.252	2665.0E6	2299.2E6	914.071	909.746
12) T	2,4,5-T	15.582	15.891	2193.3E6	1682.9E6	939.791	959.230
13) T	2,4-DB	16.345	16.495	157.9E6	188.9E6	1027.701	939.677
14) T	DINOSEB	17.833	16.886	1845.1E6	1531.2E6	911.513	913.997
15) T	Picloram	17.593	18.264	2283.8E6	1865.5E6	1028.897	1041.583
16) T	DCPA	18.196	18.146	3919.2E6	2887.7E6	918.394	910.516

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP100918\
Data File : PP027656.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 09 Oct 2018 19:40
Operator : EI\MA
Sample : HSTDICC1000
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
HSTDICC1000

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 10 09:13:50 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP100918.M
Quant Title : 8080.M
QLast Update : Wed Oct 10 09:10:34 2018
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-CLPesticides-1 Signal #2 Phase: ZB-CLPesticides-2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

