

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\
 Method File : PP100919.M
 Title : 8080.M
 Last Update : Thu Oct 10 04:06:58 2019
 Response Via : Initial Calibration

Calibration Files

750 =PP030424.D 1000 =PP030425.D 1500 =PP030426.D
 500 =PP030423.D 200 =PP030422.D

	Compound	750	1000	1500	500	200	Avg	%RSD
1) T	Dalapon	5.466	5.467	5.310	5.583	5.704	5.506 E5	2.68
2) T	3,5-DICHLOROBENZOI...	5.998	5.805	5.476	6.264	6.572	6.023 E5	6.98
3) T	4-Nitrophenol	4.188	4.104	3.910	4.276	4.168	4.129 E5	3.32
4) S	2,4-DCAA	3.589	3.458	3.208	3.782	4.042	3.616 E5	8.77
5) T	DICAMBA	1.478	1.443	1.361	1.529	1.584	1.479 E6	5.72
6) T	MCPP	9.173	9.274	8.691	9.302	9.123	9.112 E5	2.71
7) T	MCPA	1.251	1.215	1.146	1.304	1.392	1.262 E6	7.36
8) T	DICHLORPROP	3.543	3.404	3.135	3.756	4.051	3.578 E5	9.71
9) T	2,4-D	3.967	3.832	3.563	4.163	4.366	3.978 E5	7.73
10) T	Pentachlorophenol	5.765	5.589	5.167	5.979	5.987	5.697 E6	5.96
11) T	2,4,5-TP (SILVEX)	2.014	1.972	1.844	2.076	2.084	1.998 E6	4.88
12) T	2,4,5-T	1.956	1.941	1.819	2.012	1.977	1.941 E6	3.77
13) T	2,4-DB	2.582	2.513	2.352	2.598	2.682	2.546 E5	4.85
14) T	DINOSEB	1.138	1.175	1.135	1.110	0.965	1.104 E6	7.36
15) T	Picloram	2.356	2.351	2.361	2.285	2.091	2.288 E6	5.02
16) T	DCPA	2.674	2.626	2.471	2.721	2.657	2.630 E6	3.62

Signal #2 Calibration Files

750 =PP030424.D 1000 =PP030425.D 1500 =PP030426.D
 500 =PP030423.D 200 =PP030422.D

	Compound	750	1000	1500	500	200	Avg	%RSD
1) T	Dalapon	2.907	2.904	2.851	2.924	2.928	2.903 E5	1.07
2) T	3,5-DICHLOROBENZOI...	2.953	2.893	2.759	3.033	2.910	2.910 E5	3.45
3) T	4-Nitrophenol	1.177	1.193	1.175	1.124	0.783	1.090 E5	15.93
4) S	2,4-DCAA	2.207	2.175	2.064	2.273	2.269	2.198 E5	3.88
5) T	DICAMBA	8.625	8.517	8.147	8.778	8.733	8.560 E5	2.94
6) T	MCPP	4.497	4.571	4.493	4.373	3.352	4.257 E5	12.01
7) T	MCPA	7.525	7.391	7.227	7.539	6.984	7.333 E5	3.17
8) T	DICHLORPROP	2.084	2.041	1.925	2.154	2.188	2.079 E5	4.97
9) T	2,4-D	2.361	2.324	2.213	2.424	2.290	2.322 E5	3.38
10) T	Pentachlorophenol	3.386	3.279	3.025	3.524	3.496	3.342 E6	6.05
11) T	2,4,5-TP (SILVEX)	1.245	1.221	1.153	1.275	1.253	1.229 E6	3.80
12) T	2,4,5-T	1.076	1.074	1.034	1.079	0.988	1.050 E6	3.77
13) T	2,4-DB	1.445	1.420	1.383	1.469	1.392	1.422 E5	2.54
14) T	DINOSEB	6.574	6.817	6.784	6.380	5.328	6.376 E5	9.60
15) T	Picloram	1.465	1.475	1.496	1.394	1.182	1.402 E6	9.20
16) T	DCPA	1.619	1.579	1.486	1.665	1.651	1.600 E6	4.48

(#) = Out of Range