

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP100919\
 Data File : PP030427.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 09 Oct 2019 20:49
 Operator : DDAJ
 Sample : HSTDICV750
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 10 04:14:24 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP100919.M
 Quant Title : 8080.M
 QLast Update : Thu Oct 10 04:06:58 2019
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-CLPesticides-1 Signal #2 Phase: ZB-CLPesticides-2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 15% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(Min)
1 T	Dalapon	682.500	679.852	0.4	100	0.00
2 T	3,5-DICHLORO BENZOIC ACID	697.500	698.043	-0.1	100	0.00
3 T	4-Nitrophenol	682.500	697.642	-2.2	101	0.00
4 S	2,4-DCAA	750.000	746.863	0.4	100	0.00
5 T	DICAMBA	705.000	706.557	-0.2	100	0.00
6 T	MCP P	70.500	70.477	0.0	99	0.00
7 T	MCPA	69.750	68.876	1.3	100	0.00
8 T	DICHLORPROP	705.000	700.618	0.6	100	0.00
9 T	2,4-D	705.000	708.216	-0.5	101	0.00
10 T	Pentachlorophenol	712.500	722.935	-1.5	100	0.00
11 T	2,4,5-TP (SILVEX)	712.500	722.346	-1.4	101	0.00
12 T	2,4,5-T	712.500	723.060	-1.5	101	0.00
13 T	2,4-DB	712.500	724.270	-1.7	100	0.00
14 T	DINOSEB	705.000	756.150	-7.3	104	0.00
15 T	Picloram	712.500	740.010	-3.9	101	0.00
16 T	DCPA	720.000	736.426	-2.3	101	0.00

Signal #2

1 T	Dalapon	682.500	683.862	-0.2	100	0.00
2 T	3,5-DICHLORO BENZOIC ACID	697.500	711.095	-1.9	100	0.00
3 T	4-Nitrophenol	682.500	746.797	-9.4	101	0.00
4 S	2,4-DCAA	750.000	759.060	-1.2	101	0.00
5 T	DICAMBA	705.000	712.621	-1.1	100	0.00
6 T	MCP P	70.500	75.204	-6.7	101	0.00
7 T	MCPA	69.750	72.036	-3.3	101	0.00
8 T	DICHLORPROP	705.000	710.873	-0.8	101	0.00
9 T	2,4-D	705.000	726.003	-3.0	101	0.00
10 T	Pentachlorophenol	712.500	726.394	-2.0	101	0.00
11 T	2,4,5-TP (SILVEX)	712.500	726.599	-2.0	101	0.00
12 T	2,4,5-T	712.500	742.506	-4.2	102	0.00
13 T	2,4-DB	712.500	729.585	-2.4	101	0.00
14 T	DINOSEB	705.000	761.122	-8.0	105	0.00
15 T	Picloram	712.500	750.805	-5.4	101	0.00
16 T	DCPA	720.000	732.195	-1.7	101	0.00

Evaluate Continuing Calibration Report - Not Found

Signal #2

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP100919\
Data File : PP030427.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 09 Oct 2019 20:49
Operator : DDAJ
Sample : HSTDICV750
Misc :
ALS Vial : 8 Sample Multiplier: 1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 10 04:14:24 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP100919.M
Quant Title : 8080.M
QLast Update : Thu Oct 10 04:06:58 2019
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-CLPesticides-1 Signal #2 Phase: ZB-CLPesticides-2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 15% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(Min)

(#) = Out of Range

SPCC's out = 0 CCC's out = 0