

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP100923\
 Data File : PP060747.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 09 Oct 2023 11:44
 Operator : YP\AJ
 Sample : AR1254CCC500
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1254CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 09 12:33:02 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP092623.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Sep 27 11:30:16 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.531	3.682	109.5E6	78784961	48.890	52.929
2) SA Decachlor...	10.461	8.747	86463602	76811721	48.973	47.089
Target Compounds						
26) L6 AR-1254-1	6.586	5.589	40112321	44669693	470.028	451.579
27) L6 AR-1254-2	6.807	5.738	59697746	38910943	467.754	452.753
28) L6 AR-1254-3	7.178	6.145	62754796	64953786	472.138	453.187
29) L6 AR-1254-4	7.465	6.375	46057783	39109048	468.357	448.688
30) L6 AR-1254-5	7.889	6.796	51192642	58474537	472.504	461.446



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP100923\
Data File : PP060747.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 09 Oct 2023 11:44
Operator : YP\AJ
Sample : AR1254CCC500
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 09 12:33:02 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP092623.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Sep 27 11:30:16 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.531 min
Delta R.T.: -0.001 min
Response: 109489101
Conc: 48.89 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.682 min
Delta R.T.: -0.002 min
Response: 78784961
Conc: 52.93 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.461 min
Delta R.T.: 0.006 min
Response: 86463602
Conc: 48.97 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.747 min
Delta R.T.: -0.003 min
Response: 76811721
Conc: 47.09 ng/ml



#26 AR-1254-1


R.T.: 6.586 min
Delta R.T.: 0.000 min
Response: 40112321
Conc: 470.03 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500




#26 AR-1254-1

R.T.: 5.589 min
Delta R.T.: -0.004 min
Response: 44669693
Conc: 451.58 ng/ml




#27 AR-1254-2

R.T.: 6.807 min
Delta R.T.: 0.000 min
Response: 59697746
Conc: 467.75 ng/ml



#27 AR-1254-2


R.T.: 5.738 min
Delta R.T.: -0.004 min
Response: 38910943
Conc: 452.75 ng/ml



#28 AR-1254-3


R.T.: 7.178 min
Delta R.T.: 0.001 min
Response: 62754796
Conc: 472.14 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500




#28 AR-1254-3

R.T.: 6.145 min
Delta R.T.: -0.004 min
Response: 64953786
Conc: 453.19 ng/ml




#29 AR-1254-4

R.T.: 7.465 min
Delta R.T.: 0.001 min
Response: 46057783
Conc: 468.36 ng/ml



#29 AR-1254-4


R.T.: 6.375 min
Delta R.T.: -0.003 min
Response: 39109048
Conc: 448.69 ng/ml



#30 AR-1254-5

R.T.: 7.889 min
Delta R.T.: 0.002 min
Response: 51192642
Conc: 472.50 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500



#30 AR-1254-5

R.T.: 6.796 min
Delta R.T.: -0.003 min
Response: 58474537
Conc: 461.45 ng/ml