

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP100924\
 Data File : PP067664.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 09 Oct 2024 21:15
 Operator : YP\AJ
 Sample : P4364-08
 Misc :
 ALS Vial : 35 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 SB-03-2.0-2.5

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 10/10/2024
 Supervised By :Ankita Jodhani 10/10/2024

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 10 00:05:07 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP100824.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Oct 09 05:48:32 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.753	4.051	20076230	21625804	21.689m	21.403
2) SA Decachlor...	10.665	9.216	24225198	21050692	21.042	18.771m
Target Compounds						
31) L7 AR-1260-1	7.514	6.665	8461529	9623999	156.626m	168.955
32) L7 AR-1260-2	7.756	6.851	29347981	8575154	463.847	128.555m#
33) L7 AR-1260-3	8.134	7.008	5414586	6894668	104.430	108.235m
34) L7 AR-1260-4	8.370	7.481	6439993	4572782	106.987m	82.701
35) L7 AR-1260-5	8.708	7.721	9000482	14138513	83.561	115.607 #

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP100924\
Data File : PP067664.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 09 Oct 2024 21:15
Operator : YP\AJ
Sample : P4364-08
Misc :
ALS Vial : 35 Sample Multiplier: 1

Instrument :

ECD_P

ClientSampleId :

SB-03-2.0-2.5

Manual Integrations



APPROVED


Reviewed By :Yogesh Patel 10/10/2024

Supervised By :Ankita Jodhani 10/10/2024

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 10 00:05:07 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP100824.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Oct 09 05:48:32 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.753 min
Delta R.T.: -0.002 min
Response: 20076230
Conc: 21.69 ng/ml

Instrument :
ECD_P
ClientSampleId :
SB-03-2.0-2.5


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/10/2024
Supervised By :Ankita Jodhani 10/10/2024




#1 Tetrachloro-m-xylene

R.T.: 4.051 min
Delta R.T.: 0.000 min
Response: 21625804
Conc: 21.40 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.665 min
Delta R.T.: -0.004 min
Response: 24225198
Conc: 21.04 ng/ml



#2 Decachlorobiphenyl

R.T.: 9.216 min
Delta R.T.: -0.003 min
Response: 21050692
Conc: 18.77 ng/ml m




#31 AR-1260-1

R.T.: 7.514 min
Delta R.T.: -0.007 min
Response: 8461529
Conc: 156.63 ng/ml

Instrument :
ECD_P
ClientSampleId :
SB-03-2.0-2.5


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/10/2024
Supervised By :Ankita Jodhani 10/10/2024




#31 AR-1260-1

R.T.: 6.665 min
Delta R.T.: 0.000 min
Response: 9623999
Conc: 168.95 ng/ml




#32 AR-1260-2

R.T.: 7.756 min
Delta R.T.: -0.019 min
Response: 29347981
Conc: 463.85 ng/ml



#32 AR-1260-2

R.T.: 6.851 min
Delta R.T.: 0.000 min
Response: 8575154
Conc: 128.55 ng/ml m




#33 AR-1260-3

R.T.: 8.134 min
Delta R.T.: -0.003 min
Response: 5414586
Conc: 104.43 ng/ml

Instrument :
ECD_P
ClientSampleId :
SB-03-2.0-2.5


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/10/2024
Supervised By :Ankita Jodhani 10/10/2024




#33 AR-1260-3

R.T.: 7.008 min
Delta R.T.: 0.000 min
Response: 6894668
Conc: 108.23 ng/ml m




#34 AR-1260-4

R.T.: 8.370 min
Delta R.T.: -0.005 min
Response: 6439993
Conc: 106.99 ng/ml m



#34 AR-1260-4

R.T.: 7.481 min
Delta R.T.: -0.002 min
Response: 4572782
Conc: 82.70 ng/ml




#35 AR-1260-5

R.T.: 8.708 min
Delta R.T.: -0.003 min
Response: 9000482
Conc: 83.56 ng/ml

Instrument :
ECD_P
ClientSampleId :
SB-03-2.0-2.5

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/10/2024
Supervised By :Ankita Jodhani 10/10/2024



#35 AR-1260-5

R.T.: 7.721 min
Delta R.T.: 0.000 min
Response: 14138513
Conc: 115.61 ng/ml