

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP100924\
 Data File : PP067678.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 10 Oct 2024 01:44
 Operator : YP\AJ
 Sample : AR1254CCC500
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1254CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 10 02:19:50 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP100824.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Oct 09 05:48:32 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.754	4.051	46968441	48869857	50.741	48.367
2) SA Decachlor...	10.665	9.215	56747145	54546919	49.289	48.639
Target Compounds						
26) L6 AR-1254-1	6.773	5.975	22958778	28878460	512.673	488.770
27) L6 AR-1254-2	6.990	6.122	33106022	25625743	500.865	488.443
28) L6 AR-1254-3	7.355	6.532	34257938	40469441	493.256	486.364
29) L6 AR-1254-4	7.637	6.760	25062137	23896638	502.201	495.018
30) L6 AR-1254-5	8.056	7.182	28763291	35523484	485.107	482.976



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP100924\
Data File : PP067678.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Oct 2024 01:44
Operator : YP\AJ
Sample : AR1254CCC500
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 10 02:19:50 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP100824.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Oct 09 05:48:32 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.754 min
Delta R.T.: -0.001 min
Response: 46968441
Conc: 50.74 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500




#1 Tetrachloro-m-xylene

R.T.: 4.051 min
Delta R.T.: 0.000 min
Response: 48869857
Conc: 48.37 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.665 min
Delta R.T.: -0.004 min
Response: 56747145
Conc: 49.29 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.215 min
Delta R.T.: -0.004 min
Response: 54546919
Conc: 48.64 ng/ml



#26 AR-1254-1


R.T.: 6.773 min
Delta R.T.: 0.000 min
Response: 22958778
Conc: 512.67 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500




#26 AR-1254-1

R.T.: 5.975 min
Delta R.T.: 0.000 min
Response: 28878460
Conc: 488.77 ng/ml




#27 AR-1254-2

R.T.: 6.990 min
Delta R.T.: 0.000 min
Response: 33106022
Conc: 500.86 ng/ml



#27 AR-1254-2


R.T.: 6.122 min
Delta R.T.: 0.000 min
Response: 25625743
Conc: 488.44 ng/ml



#28 AR-1254-3


R.T.: 7.355 min
Delta R.T.: 0.000 min
Response: 34257938
Conc: 493.26 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500




#28 AR-1254-3

R.T.: 6.532 min
Delta R.T.: 0.000 min
Response: 40469441
Conc: 486.36 ng/ml




#29 AR-1254-4

R.T.: 7.637 min
Delta R.T.: 0.000 min
Response: 25062137
Conc: 502.20 ng/ml



#29 AR-1254-4


R.T.: 6.760 min
Delta R.T.: 0.000 min
Response: 23896638
Conc: 495.02 ng/ml



#30 AR-1254-5

R.T.: 8.056 min
Delta R.T.: 0.001 min
Response: 28763291
Conc: 485.11 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500



#30 AR-1254-5

R.T.: 7.182 min
Delta R.T.: 0.000 min
Response: 35523484
Conc: 482.98 ng/ml