

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP101023\
 Data File : PP060850.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 10 Oct 2023 16:46
 Operator : YP\AJ
 Sample : AR1248CCC500
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1248CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 11 02:09:52 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP092623.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Sep 27 11:30:16 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.533	3.684	112.5E6	78417437	50.243	52.682
2) SA Decachlor...	10.466	8.747	88386707	79557536	50.062	48.773
Target Compounds						
21) L5 AR-1248-1	5.720	4.781	24111742	17547356	518.260	480.833
22) L5 AR-1248-2	5.998	5.020	34763084	24182149	501.377	462.622
23) L5 AR-1248-3	6.206	5.063	38275593	25671335	503.961	462.344
24) L5 AR-1248-4	6.614	5.236	41130519	29705960	497.311	451.272
25) L5 AR-1248-5	6.653	5.632	40404234	29137329	502.022	452.990



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP101023\
Data File : PP060850.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Oct 2023 16:46
Operator : YP\AJ
Sample : AR1248CCC500
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 11 02:09:52 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP092623.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Sep 27 11:30:16 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.533 min
Delta R.T.: 0.000 min
Response: 112519798
Conc: 50.24 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.684 min
Delta R.T.: -0.001 min
Response: 78417437
Conc: 52.68 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.466 min
Delta R.T.: 0.010 min
Response: 88386707
Conc: 50.06 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.747 min
Delta R.T.: -0.003 min
Response: 79557536
Conc: 48.77 ng/ml



#21 AR-1248-1


R.T.: 5.720 min
Delta R.T.: 0.001 min
Response: 24111742
Conc: 518.26 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500




#21 AR-1248-1

R.T.: 4.781 min
Delta R.T.: -0.002 min
Response: 17547356
Conc: 480.83 ng/ml




#22 AR-1248-2

R.T.: 5.998 min
Delta R.T.: 0.001 min
Response: 34763084
Conc: 501.38 ng/ml



#22 AR-1248-2


R.T.: 5.020 min
Delta R.T.: -0.003 min
Response: 24182149
Conc: 462.62 ng/ml



#23 AR-1248-3


R.T.: 6.206 min
Delta R.T.: 0.002 min
Response: 38275593
Conc: 503.96 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500




#23 AR-1248-3

R.T.: 5.063 min
Delta R.T.: -0.002 min
Response: 25671335
Conc: 462.34 ng/ml




#24 AR-1248-4

R.T.: 6.614 min
Delta R.T.: 0.002 min
Response: 41130519
Conc: 497.31 ng/ml



#24 AR-1248-4


R.T.: 5.236 min
Delta R.T.: -0.003 min
Response: 29705960
Conc: 451.27 ng/ml



#25 AR-1248-5

R.T.: 6.653 min
Delta R.T.: 0.002 min
Response: 40404234
Conc: 502.02 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500



#25 AR-1248-5

R.T.: 5.632 min
Delta R.T.: -0.003 min
Response: 29137329
Conc: 452.99 ng/ml