

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP101024\
 Data File : PP067682.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 10 Oct 2024 09:49
 Operator : YP\AJ
 Sample : AR1242CCC500
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1242CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 11 01:38:45 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP100824.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Oct 09 05:48:32 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.756	4.051	48011345	51278804	51.868	50.751
2) SA Decachlor...	10.669	9.216	58510714	55767976	50.821	49.728
Target Compounds						
16) L4 AR-1242-1	5.919	5.159	14383605	14538770	514.856	500.019
17) L4 AR-1242-2	5.942	5.180	20895232	19876329	514.221	495.080
18) L4 AR-1242-3	6.006	5.360	13696653	11253908	507.947	500.284
19) L4 AR-1242-4	6.104	5.444	10934566	12012406	509.759	506.048
20) L4 AR-1242-5	6.838	5.973	12867977	14740526	533.368	514.511



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP101024\
Data File : PP067682.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Oct 2024 09:49
Operator : YP\AJ
Sample : AR1242CCC500
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 11 01:38:45 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP100824.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Oct 09 05:48:32 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.756 min
Delta R.T.: 0.000 min
Response: 48011345
Conc: 51.87 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500




#1 Tetrachloro-m-xylene

R.T.: 4.051 min
Delta R.T.: 0.000 min
Response: 51278804
Conc: 50.75 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.669 min
Delta R.T.: 0.000 min
Response: 58510714
Conc: 50.82 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.216 min
Delta R.T.: -0.003 min
Response: 55767976
Conc: 49.73 ng/ml



#16 AR-1242-1


R.T.: 5.919 min
Delta R.T.: 0.003 min
Response: 14383605
Conc: 514.86 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500




#16 AR-1242-1

R.T.: 5.159 min
Delta R.T.: -0.001 min
Response: 14538770
Conc: 500.02 ng/ml




#17 AR-1242-2

R.T.: 5.942 min
Delta R.T.: 0.003 min
Response: 20895232
Conc: 514.22 ng/ml



#17 AR-1242-2


R.T.: 5.180 min
Delta R.T.: -0.001 min
Response: 19876329
Conc: 495.08 ng/ml



#18 AR-1242-3


R.T.: 6.006 min
Delta R.T.: 0.003 min
Response: 13696653
Conc: 507.95 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500




#18 AR-1242-3

R.T.: 5.360 min
Delta R.T.: -0.001 min
Response: 11253908
Conc: 500.28 ng/ml




#19 AR-1242-4

R.T.: 6.104 min
Delta R.T.: 0.003 min
Response: 10934566
Conc: 509.76 ng/ml



#19 AR-1242-4


R.T.: 5.444 min
Delta R.T.: -0.001 min
Response: 12012406
Conc: 506.05 ng/ml



#20 AR-1242-5

R.T.: 6.838 min
Delta R.T.: 0.002 min
Response: 12867977
Conc: 533.37 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500



#20 AR-1242-5

R.T.: 5.973 min
Delta R.T.: -0.002 min
Response: 14740526
Conc: 514.51 ng/ml