

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP101024\
 Data File : PP067724.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 10 Oct 2024 22:11
 Operator : YP\AJ
 Sample : AR1254CCC500
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1254CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 11 01:51:32 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP100824.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Oct 09 05:48:32 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.754	4.051	48911437	51617945	52.841	51.087
2) SA Decachlor...	10.664	9.214	61930708	59145049	53.792	52.739
Target Compounds						
26) L6 AR-1254-1	6.773	5.974	25966139	31503137	579.828	533.193
27) L6 AR-1254-2	6.990	6.121	36359168	27654244	550.082	527.107
28) L6 AR-1254-3	7.354	6.531	37491627	43033924	539.815	517.185
29) L6 AR-1254-4	7.637	6.758	27234916	25121835	545.740	520.398
30) L6 AR-1254-5	8.056	7.179	31118876	37706759	524.835	512.660



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP101024\
Data File : PP067724.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Oct 2024 22:11
Operator : YP\AJ
Sample : AR1254CCC500
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 11 01:51:32 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP100824.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Oct 09 05:48:32 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.754 min
Delta R.T.: 0.000 min
Response: 48911437
Conc: 52.84 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500




#1 Tetrachloro-m-xylene

R.T.: 4.051 min
Delta R.T.: -0.001 min
Response: 51617945
Conc: 51.09 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.664 min
Delta R.T.: -0.005 min
Response: 61930708
Conc: 53.79 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.214 min
Delta R.T.: -0.005 min
Response: 59145049
Conc: 52.74 ng/ml



#26 AR-1254-1


R.T.: 6.773 min
Delta R.T.: 0.000 min
Response: 25966139
Conc: 579.83 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500




#26 AR-1254-1

R.T.: 5.974 min
Delta R.T.: -0.002 min
Response: 31503137
Conc: 533.19 ng/ml




#27 AR-1254-2

R.T.: 6.990 min
Delta R.T.: 0.000 min
Response: 36359168
Conc: 550.08 ng/ml



#27 AR-1254-2


R.T.: 6.121 min
Delta R.T.: -0.001 min
Response: 27654244
Conc: 527.11 ng/ml



#28 AR-1254-3


R.T.: 7.354 min
Delta R.T.: 0.000 min
Response: 37491627
Conc: 539.82 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500




#28 AR-1254-3

R.T.: 6.531 min
Delta R.T.: -0.002 min
Response: 43033924
Conc: 517.18 ng/ml




#29 AR-1254-4

R.T.: 7.637 min
Delta R.T.: 0.000 min
Response: 27234916
Conc: 545.74 ng/ml



#29 AR-1254-4


R.T.: 6.758 min
Delta R.T.: -0.002 min
Response: 25121835
Conc: 520.40 ng/ml



#30 AR-1254-5

R.T.: 8.056 min
Delta R.T.: 0.001 min
Response: 31118876
Conc: 524.83 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500



#30 AR-1254-5

R.T.: 7.179 min
Delta R.T.: -0.002 min
Response: 37706759
Conc: 512.66 ng/ml