

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP101024\
 Data File : PP067736.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 11 Oct 2024 01:47
 Operator : YP\AJ
 Sample : AR1248CCC500
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1248CCC500

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 10/11/2024
 Supervised By :Ankita Jodhani 10/11/2024

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 11 03:50:32 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP100824.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Oct 09 05:48:32 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.754	4.053	45466476	49756294	49.119	49.244
2) SA Decachlor...	10.665	9.215	57344481	52929345	49.808	47.197
Target Compounds						
21) L5 AR-1248-1	5.917	5.161	10623989	11357962	495.918	501.150
22) L5 AR-1248-2	6.191	5.401	16117085	14928395	479.437	451.066
23) L5 AR-1248-3	6.395	5.444	17912324	17036117	489.068m	494.164
24) L5 AR-1248-4	6.795	5.619	19097289	20264049	472.112m	494.777m
25) L5 AR-1248-5	6.835	6.017	18788703	19314779	473.733m	515.843m

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP101024\
Data File : PP067736.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 11 Oct 2024 01:47
Operator : YP\AJ
Sample : AR1248CCC500
Misc :
ALS Vial : 5 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
AR1248CCC500


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/11/2024
Supervised By :Ankita Jodhani 10/11/2024

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 11 03:50:32 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP100824.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Oct 09 05:48:32 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.754 min
Delta R.T.: -0.001 min
Response: 45466476
Conc: 49.12 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/11/2024
Supervised By :Ankita Jodhani 10/11/2024




#1 Tetrachloro-m-xylene

R.T.: 4.053 min
Delta R.T.: 0.000 min
Response: 49756294
Conc: 49.24 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.665 min
Delta R.T.: -0.004 min
Response: 57344481
Conc: 49.81 ng/ml



#2 Decachlorobiphenyl

R.T.: 9.215 min
Delta R.T.: -0.004 min
Response: 52929345
Conc: 47.20 ng/ml




#21 AR-1248-1

R.T.: 5.917 min
Delta R.T.: 0.000 min
Response: 10623989
Conc: 495.92 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/11/2024
Supervised By :Ankita Jodhani 10/11/2024




#21 AR-1248-1

R.T.: 5.161 min
Delta R.T.: 0.000 min
Response: 11357962
Conc: 501.15 ng/ml




#22 AR-1248-2

R.T.: 6.191 min
Delta R.T.: 0.000 min
Response: 16117085
Conc: 479.44 ng/ml



#22 AR-1248-2

R.T.: 5.401 min
Delta R.T.: 0.001 min
Response: 14928395
Conc: 451.07 ng/ml




#23 AR-1248-3

R.T.: 6.395 min
Delta R.T.: -0.001 min
Response: 17912324
Conc: 489.07 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/11/2024
Supervised By :Ankita Jodhani 10/11/2024




#23 AR-1248-3

R.T.: 5.444 min
Delta R.T.: 0.000 min
Response: 17036117
Conc: 494.16 ng/ml




#24 AR-1248-4

R.T.: 6.795 min
Delta R.T.: 0.000 min
Response: 19097289
Conc: 472.11 ng/ml m



#24 AR-1248-4

R.T.: 5.619 min
Delta R.T.: 0.000 min
Response: 20264049
Conc: 494.78 ng/ml m




#25 AR-1248-5

R.T.: 6.835 min
Delta R.T.: -0.001 min
Response: 18788703
Conc: 473.73 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/11/2024
Supervised By :Ankita Jodhani 10/11/2024



#25 AR-1248-5

R.T.: 6.017 min
Delta R.T.: 0.000 min
Response: 19314779
Conc: 515.84 ng/ml m