

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP101122\
 Data File : PP052191.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 11 Oct 2022 22:54
 Operator : YP\AJ
 Sample : AR1248ICC1000
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1248ICC1000

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 12 03:35:14 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP101122.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Oct 12 03:33:29 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.424	3.675	164.6E6	126.1E6	97.314	97.914
2) SA Decachlor...	10.235	8.764	135.5E6	113.2E6	92.358	93.419
Target Compounds						
21) L5 AR-1248-1	5.596	4.776	26059496	26327572	922.251	935.323
22) L5 AR-1248-2	5.872	5.016	44178862	35294347	916.102	918.754
23) L5 AR-1248-3	6.076	5.059	53381300	36202733	916.973	919.472
24) L5 AR-1248-4	6.481	5.234	59825137	44029493	908.868	937.049
25) L5 AR-1248-5	6.520	5.632	58455532	41195845	913.608	948.529



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP101122\
Data File : PP052191.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 11 Oct 2022 22:54
Operator : YP\AJ
Sample : AR1248ICC1000
Misc :
ALS Vial : 15 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1248ICC1000

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 12 03:35:14 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP101122.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Oct 12 03:33:29 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.424 min
Delta R.T.: 0.000 min
Response: 164601509
Conc: 97.31 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248ICC1000




#1 Tetrachloro-m-xylene

R.T.: 3.675 min
Delta R.T.: 0.000 min
Response: 126140685
Conc: 97.91 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.235 min
Delta R.T.: -0.002 min
Response: 135518055
Conc: 92.36 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.764 min
Delta R.T.: -0.001 min
Response: 113186074
Conc: 93.42 ng/ml



#21 AR-1248-1


R.T.: 5.596 min
Delta R.T.: 0.000 min
Response: 26059496
Conc: 922.25 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248ICC1000




#21 AR-1248-1

R.T.: 4.776 min
Delta R.T.: 0.000 min
Response: 26327572
Conc: 935.32 ng/ml




#22 AR-1248-2

R.T.: 5.872 min
Delta R.T.: 0.000 min
Response: 44178862
Conc: 916.10 ng/ml



#22 AR-1248-2


R.T.: 5.016 min
Delta R.T.: 0.000 min
Response: 35294347
Conc: 918.75 ng/ml



#23 AR-1248-3


R.T.: 6.076 min
Delta R.T.: 0.000 min
Response: 53381300
Conc: 916.97 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248ICC1000




#23 AR-1248-3

R.T.: 5.059 min
Delta R.T.: 0.000 min
Response: 36202733
Conc: 919.47 ng/ml




#24 AR-1248-4

R.T.: 6.481 min
Delta R.T.: 0.000 min
Response: 59825137
Conc: 908.87 ng/ml



#24 AR-1248-4


R.T.: 5.234 min
Delta R.T.: 0.000 min
Response: 44029493
Conc: 937.05 ng/ml



#25 AR-1248-5

R.T.: 6.520 min
Delta R.T.: 0.000 min
Response: 58455532
Conc: 913.61 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248ICC1000



#25 AR-1248-5

R.T.: 5.632 min
Delta R.T.: 0.000 min
Response: 41195845
Conc: 948.53 ng/ml