

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP101122\
 Data File : PP052205.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 12 Oct 2022 02:44
 Operator : YP\AJ
 Sample : AR1268ICC250
 Misc :
 ALS Vial : 29 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1268ICC250

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 29 05:35:52 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP101122.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Oct 12 05:20:17 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.423	3.675	43968961	33230798	27.973	29.006
2)	SA Decachlor...	10.235	8.763	66275296	54178648	47.754	48.488
Target Compounds							
3)	L1 AR-1016-1	5.595	4.776	100077	50429	2.201	1.134 #
4)	L1 AR-1016-2	5.615	4.776f	138165	50429	2.099	0.793 #
5)	L1 AR-1016-3	5.682	0.000	142055	0	3.414	N.D. #
6)	L1 AR-1016-4	5.782	0.000	62879	0	1.731	N.D. #
7)	L1 AR-1016-5	6.078	5.236	194868	259937	4.566	7.347 #
8)	L2 AR-1221-1	4.661f	0.000	52595	0	2.545	N.D. #
9)	L2 AR-1221-2	4.730	3.980	563976	431652	36.853	35.993
10)	L2 AR-1221-3	4.790	0.000	54670	0	1.168	N.D. #
11)	L3 AR-1232-1	4.790	0.000	54670	0	1.517	N.D. #
12)	L3 AR-1232-2	5.327	4.776f	129254	50429	6.370	1.683 #
13)	L3 AR-1232-3	5.615	0.000	138165	0	4.588	N.D. #
14)	L3 AR-1232-4	5.782	0.000	62879	0	3.770	N.D. #
15)	L3 AR-1232-5	5.872	5.236	190747	259937	12.718	16.420 #
16)	L4 AR-1242-1	5.595	4.776	100077	50429	2.877	1.448 #
17)	L4 AR-1242-2	5.615	4.776f	138165	50429	2.791	1.022 #
18)	L4 AR-1242-3	5.682	0.000	142055	0	4.480	N.D. #
19)	L4 AR-1242-4	5.782	0.000	62879	0	2.302	N.D. #
20)	L4 AR-1242-5	6.519	5.590	404258	273816	10.874	8.125 #
21)	L5 AR-1248-1	5.595	4.776	100077	50429	3.723	1.860 #
22)	L5 AR-1248-2	5.872	0.000	190747	0	4.104	N.D. #
23)	L5 AR-1248-3	6.078	0.000	194868	0	3.519	N.D. #
24)	L5 AR-1248-4	6.482	5.236	317649	259937	5.003	5.610
25)	L5 AR-1248-5	6.519	5.632	404258	241908	6.527	5.752
26)	L6 AR-1254-1	6.454	5.590	262343	273816	3.685	3.914
27)	L6 AR-1254-2	6.678	0.000	390159	0	3.658	N.D. #
28)	L6 AR-1254-3	7.043	0.000	356555	0	3.393	N.D. #
29)	L6 AR-1254-4	7.329	6.380	268588	168807	3.676	2.824
30)	L6 AR-1254-5	7.737	0.000	175308	0	2.113	N.D. #
31)	L7 AR-1260-1	7.205	0.000	91374	0	1.102	N.D. #
32)	L7 AR-1260-2	7.468	6.469	384788	359973	4.077	4.349
33)	L7 AR-1260-3	7.825	6.632	10608567	628203	176.636	8.137 #
34)	L7 AR-1260-4	8.072f	7.100	14514361	12025211	189.251	213.916
35)	L7 AR-1260-5	8.374	7.345	5541594	4481527	40.427	33.983
36)	L8 AR-1262-1	7.825	6.894	10608567	534874	109.407	14.869 #
37)	L8 AR-1262-2	8.374	7.100	5541594	12025211	34.418	151.864 #
38)	L8 AR-1262-3	8.691	7.631	52066484	47207402	473.763	764.467 #
39)	L8 AR-1262-4	8.786	7.696	48606428	42785384	950.855	374.773 #
40)	L8 AR-1262-5	9.455	8.197	20137095	14961735	319.895	294.357
41)	L9 AR-1268-1	8.691	7.631	52066484	47207402	262.990	260.225

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP101122\
Data File : PP052205.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 12 Oct 2022 02:44
Operator : YP\AJ
Sample : AR1268ICC250
Misc :
ALS Vial : 29 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1268ICC250

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 29 05:35:52 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP101122.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Oct 12 05:20:17 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
42)	L9 AR-1268-2	8.786	7.696	48606428	42785384	264.709	261.865
43)	L9 AR-1268-3	9.021	7.906	42945316	36605954	263.446	260.314
44)	L9 AR-1268-4	9.455	8.197	20137095	14961735	275.995	264.756
45)	L9 AR-1268-5	9.884	8.498	135.5E6	108.4E6	259.904	257.767



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP101122\
Data File : PP052205.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 12 Oct 2022 02:44
Operator : YP\AJ
Sample : AR1268ICC250
Misc :
ALS Vial : 29 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1268ICC250

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 29 05:35:52 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP101122.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Oct 12 05:20:17 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.423 min
Delta R.T.: -0.003 min
Response: 43968961
Conc: 27.97 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1268ICC250




#1 Tetrachloro-m-xylene

R.T.: 3.675 min
Delta R.T.: -0.002 min
Response: 33230798
Conc: 29.01 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.235 min
Delta R.T.: -0.002 min
Response: 66275296
Conc: 47.75 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.763 min
Delta R.T.: -0.003 min
Response: 54178648
Conc: 48.49 ng/ml



#3 AR-1016-1


R.T.: 5.595 min
Delta R.T.: -0.003 min
Response: 100077
Conc: 2.20 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1268ICC250




#3 AR-1016-1

R.T.: 4.776 min
Delta R.T.: -0.003 min
Response: 50429
Conc: 1.13 ng/ml




#4 AR-1016-2

R.T.: 5.615 min
Delta R.T.: -0.005 min
Response: 138165
Conc: 2.10 ng/ml



#4 AR-1016-2


R.T.: 4.776 min
Delta R.T.: -0.021 min
Response: 50429
Conc: 0.79 ng/ml



#5 AR-1016-3


R.T.: 5.682 min
Delta R.T.: -0.001 min
Response: 142055
Conc: 3.41 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1268ICC250




#5 AR-1016-3

R.T.: 0.000 min
Exp R.T. : 4.976 min
Response: 0
Conc: N.D.




#6 AR-1016-4

R.T.: 5.782 min
Delta R.T.: 0.000 min
Response: 62879
Conc: 1.73 ng/ml



#6 AR-1016-4


R.T.: 0.000 min
Exp R.T. : 5.018 min
Response: 0
Conc: N.D.



#7 AR-1016-5


R.T.: 6.078 min
Delta R.T.: 0.000 min
Response: 194868
Conc: 4.57 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1268ICC250




#7 AR-1016-5

R.T.: 5.236 min
Delta R.T.: 0.000 min
Response: 259937
Conc: 7.35 ng/ml




#8 AR-1221-1

R.T.: 4.661 min
Delta R.T.: 0.029 min
Response: 52595
Conc: 2.55 ng/ml



#8 AR-1221-1


R.T.: 0.000 min
Exp R.T. : 3.892 min
Response: 0
Conc: N.D.



#9 AR-1221-2


R.T.: 4.730 min
Delta R.T.: 0.011 min
Response: 563976
Conc: 36.85 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1268ICC250




#9 AR-1221-2

R.T.: 3.980 min
Delta R.T.: 0.000 min
Response: 431652
Conc: 35.99 ng/ml




#10 AR-1221-3

R.T.: 4.790 min
Delta R.T.: -0.006 min
Response: 54670
Conc: 1.17 ng/ml



#10 AR-1221-3


R.T.: 0.000 min
Exp R.T. : 4.056 min
Response: 0
Conc: N.D.



#11 AR-1232-1


R.T.: 4.790 min
Delta R.T.: -0.006 min
Response: 54670
Conc: 1.52 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1268ICC250




#11 AR-1232-1

R.T.: 0.000 min
Exp R.T. : 4.056 min
Response: 0
Conc: N.D.




#12 AR-1232-2

R.T.: 5.327 min
Delta R.T.: 0.000 min
Response: 129254
Conc: 6.37 ng/ml



#12 AR-1232-2


R.T.: 4.776 min
Delta R.T.: -0.021 min
Response: 50429
Conc: 1.68 ng/ml



#13 AR-1232-3


R.T.: 5.615 min
Delta R.T.: -0.003 min
Response: 138165
Conc: 4.59 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1268ICC250




#13 AR-1232-3

R.T.: 0.000 min
Exp R.T. : 4.975 min
Response: 0
Conc: N.D.




#14 AR-1232-4

R.T.: 5.782 min
Delta R.T.: 0.001 min
Response: 62879
Conc: 3.77 ng/ml



#14 AR-1232-4


R.T.: 0.000 min
Exp R.T. : 5.060 min
Response: 0
Conc: N.D.



#15 AR-1232-5


R.T.: 5.872 min
Delta R.T.: 0.000 min
Response: 190747
Conc: 12.72 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1268ICC250




#15 AR-1232-5

R.T.: 5.236 min
Delta R.T.: 0.000 min
Response: 259937
Conc: 16.42 ng/ml




#16 AR-1242-1

R.T.: 5.595 min
Delta R.T.: -0.002 min
Response: 100077
Conc: 2.88 ng/ml



#16 AR-1242-1


R.T.: 4.776 min
Delta R.T.: -0.001 min
Response: 50429
Conc: 1.45 ng/ml



#17 AR-1242-2


R.T.: 5.615 min
Delta R.T.: -0.004 min
Response: 138165
Conc: 2.79 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1268ICC250




#17 AR-1242-2

R.T.: 4.776 min
Delta R.T.: -0.020 min
Response: 50429
Conc: 1.02 ng/ml




#18 AR-1242-3

R.T.: 5.682 min
Delta R.T.: 0.000 min
Response: 142055
Conc: 4.48 ng/ml



#18 AR-1242-3


R.T.: 0.000 min
Exp R.T. : 4.975 min
Response: 0
Conc: N.D.



#19 AR-1242-4


R.T.: 5.782 min
Delta R.T.: 0.002 min
Response: 62879
Conc: 2.30 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1268ICC250




#19 AR-1242-4

R.T.: 0.000 min
Exp R.T. : 5.060 min
Response: 0
Conc: N.D.




#20 AR-1242-5

R.T.: 6.519 min
Delta R.T.: -0.003 min
Response: 404258
Conc: 10.87 ng/ml



#20 AR-1242-5


R.T.: 5.590 min
Delta R.T.: -0.001 min
Response: 273816
Conc: 8.13 ng/ml



#21 AR-1248-1


R.T.: 5.595 min
Delta R.T.: 0.000 min
Response: 100077
Conc: 3.72 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1268ICC250




#21 AR-1248-1

R.T.: 4.776 min
Delta R.T.: 0.000 min
Response: 50429
Conc: 1.86 ng/ml




#22 AR-1248-2

R.T.: 5.872 min
Delta R.T.: 0.000 min
Response: 190747
Conc: 4.10 ng/ml



#22 AR-1248-2


R.T.: 0.000 min
Exp R.T. : 5.016 min
Response: 0
Conc: N.D.



#23 AR-1248-3


R.T.: 6.078 min
Delta R.T.: 0.002 min
Response: 194868
Conc: 3.52 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1268ICC250




#23 AR-1248-3

R.T.: 0.000 min
Exp R.T. : 5.059 min
Response: 0
Conc: N.D.




#24 AR-1248-4

R.T.: 6.482 min
Delta R.T.: 0.000 min
Response: 317649
Conc: 5.00 ng/ml



#24 AR-1248-4


R.T.: 5.236 min
Delta R.T.: 0.002 min
Response: 259937
Conc: 5.61 ng/ml



#25 AR-1248-5


R.T.: 6.519 min
Delta R.T.: -0.002 min
Response: 404258
Conc: 6.53 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1268ICC250




#25 AR-1248-5

R.T.: 5.632 min
Delta R.T.: 0.000 min
Response: 241908
Conc: 5.75 ng/ml




#26 AR-1254-1

R.T.: 6.454 min
Delta R.T.: -0.001 min
Response: 262343
Conc: 3.69 ng/ml



#26 AR-1254-1


R.T.: 5.590 min
Delta R.T.: 0.000 min
Response: 273816
Conc: 3.91 ng/ml



#27 AR-1254-2


R.T.: 6.678 min
Delta R.T.: 0.003 min
Response: 390159
Conc: 3.66 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1268ICC250




#27 AR-1254-2

R.T.: 0.000 min
Exp R.T. : 5.740 min
Response: 0
Conc: N.D.




#28 AR-1254-3

R.T.: 7.043 min
Delta R.T.: 0.000 min
Response: 356555
Conc: 3.39 ng/ml



#28 AR-1254-3


R.T.: 0.000 min
Exp R.T. : 6.150 min
Response: 0
Conc: N.D.



#29 AR-1254-4


R.T.: 7.329 min
Delta R.T.: 0.000 min
Response: 268588
Conc: 3.68 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1268ICC250




#29 AR-1254-4

R.T.: 6.380 min
Delta R.T.: 0.000 min
Response: 168807
Conc: 2.82 ng/ml




#30 AR-1254-5

R.T.: 7.737 min
Delta R.T.: -0.012 min
Response: 175308
Conc: 2.11 ng/ml



#30 AR-1254-5


R.T.: 0.000 min
Exp R.T.: 6.803 min
Response: 0
Conc: N.D.



#31 AR-1260-1


R.T.: 7.205 min
Delta R.T.: -0.004 min
Response: 91374
Conc: 1.10 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1268ICC250




#31 AR-1260-1

R.T.: 0.000 min
Exp R.T. : 6.283 min
Response: 0
Conc: N.D.




#32 AR-1260-2

R.T.: 7.468 min
Delta R.T.: 0.002 min
Response: 384788
Conc: 4.08 ng/ml



#32 AR-1260-2


R.T.: 6.469 min
Delta R.T.: -0.003 min
Response: 359973
Conc: 4.35 ng/ml



#33 AR-1260-3


R.T.: 7.825 min
Delta R.T.: -0.002 min
Response: 10608567
Conc: 176.64 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1268ICC250




#33 AR-1260-3

R.T.: 6.632 min
Delta R.T.: 0.003 min
Response: 628203
Conc: 8.14 ng/ml




#34 AR-1260-4

R.T.: 8.072 min
Delta R.T.: 0.019 min
Response: 14514361
Conc: 189.25 ng/ml



#34 AR-1260-4


R.T.: 7.100 min
Delta R.T.: -0.005 min
Response: 12025211
Conc: 213.92 ng/ml



#35 AR-1260-5


R.T.: 8.374 min
Delta R.T.: -0.003 min
Response: 5541594
Conc: 40.43 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1268ICC250




#35 AR-1260-5

R.T.: 7.345 min
Delta R.T.: -0.003 min
Response: 4481527
Conc: 33.98 ng/ml




#36 AR-1262-1

R.T.: 7.825 min
Delta R.T.: 0.000 min
Response: 10608567
Conc: 109.41 ng/ml



#36 AR-1262-1


R.T.: 6.894 min
Delta R.T.: 0.000 min
Response: 534874
Conc: 14.87 ng/ml



#37 AR-1262-2


R.T.: 8.374 min
Delta R.T.: -0.001 min
Response: 5541594
Conc: 34.42 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1268ICC250




#37 AR-1262-2

R.T.: 7.100 min
Delta R.T.: -0.003 min
Response: 12025211
Conc: 151.86 ng/ml




#38 AR-1262-3

R.T.: 8.691 min
Delta R.T.: -0.004 min
Response: 52066484
Conc: 473.76 ng/ml



#38 AR-1262-3


R.T.: 7.631 min
Delta R.T.: 0.000 min
Response: 47207402
Conc: 764.47 ng/ml



#39 AR-1262-4


R.T.: 8.786 min
Delta R.T.: 0.000 min
Response: 48606428
Conc: 950.85 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1268ICC250




#39 AR-1262-4

R.T.: 7.696 min
Delta R.T.: 0.000 min
Response: 42785384
Conc: 374.77 ng/ml




#40 AR-1262-5

R.T.: 9.455 min
Delta R.T.: -0.002 min
Response: 20137095
Conc: 319.90 ng/ml



#40 AR-1262-5


R.T.: 8.197 min
Delta R.T.: 0.000 min
Response: 14961735
Conc: 294.36 ng/ml



#41 AR-1268-1


R.T.: 8.691 min
Delta R.T.: 0.000 min
Response: 52066484
Conc: 262.99 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1268ICC250




#41 AR-1268-1

R.T.: 7.631 min
Delta R.T.: 0.000 min
Response: 47207402
Conc: 260.22 ng/ml




#42 AR-1268-2

R.T.: 8.786 min
Delta R.T.: -0.001 min
Response: 48606428
Conc: 264.71 ng/ml



#42 AR-1268-2


R.T.: 7.696 min
Delta R.T.: 0.000 min
Response: 42785384
Conc: 261.87 ng/ml



#43 AR-1268-3


R.T.: 9.021 min
Delta R.T.: 0.000 min
Response: 42945316
Conc: 263.45 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1268ICC250




#43 AR-1268-3

R.T.: 7.906 min
Delta R.T.: 0.000 min
Response: 36605954
Conc: 260.31 ng/ml




#44 AR-1268-4

R.T.: 9.455 min
Delta R.T.: -0.002 min
Response: 20137095
Conc: 275.99 ng/ml



#44 AR-1268-4


R.T.: 8.197 min
Delta R.T.: 0.000 min
Response: 14961735
Conc: 264.76 ng/ml



#45 AR-1268-5

R.T.: 9.884 min
Delta R.T.: 0.000 min
Response: 135525285
Conc: 259.90 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1268ICC250



#45 AR-1268-5

R.T.: 8.498 min
Delta R.T.: 0.000 min
Response: 108432384
Conc: 257.77 ng/ml