

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP101122\
 Data File : PP052203.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 12 Oct 2022 02:11
 Operator : YP\AJ
 Sample : AR1268ICC750
 Misc :
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1268ICC750

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 10/12/2022
 Supervised By :Ankita Jodhani 10/13/2022

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 12 04:39:08 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP101122.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Oct 12 04:39:00 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.423	3.674	130.8E6	100.3E6	75.275	75.236
2) SA Decachlor...	10.234	8.763	181.8E6	156.4E6	74.841	75.119
Target Compounds						
41) L9 AR-1268-1	8.691	7.631	145.2E6	138.4E6	747.574	753.865
42) L9 AR-1268-2	8.786	7.696	135.2E6	126.2E6	748.547	754.707
43) L9 AR-1268-3	9.021	7.905	119.4E6	107.0E6	750.170	750.874
44) L9 AR-1268-4	9.456	8.196	54838810	43528455	753.274	751.728
45) L9 AR-1268-5	9.885	8.497	389.8E6	329.5E6	750.925	762.637m

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP101122\
Data File : PP052203.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 12 Oct 2022 02:11
Operator : YP\AJ
Sample : AR1268ICC750
Misc :
ALS Vial : 27 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
AR1268ICC750


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/12/2022
Supervised By :Ankita Jodhani 10/13/2022

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 12 04:39:08 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP101122.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Oct 12 04:39:00 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.423 min
Delta R.T.: 0.000 min
Response: 130789151
Conc: 75.28 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1268ICC750


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/12/2022
Supervised By :Ankita Jodhani 10/13/2022




#1 Tetrachloro-m-xylene

R.T.: 3.674 min
Delta R.T.: 0.000 min
Response: 100333008
Conc: 75.24 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.234 min
Delta R.T.: 0.000 min
Response: 181840509
Conc: 74.84 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.763 min
Delta R.T.: 0.000 min
Response: 156434575
Conc: 75.12 ng/ml




#41 AR-1268-1

R.T.: 8.691 min
Delta R.T.: 0.000 min
Response: 145248637
Conc: 747.57 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1268ICC750


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/12/2022
Supervised By :Ankita Jodhani 10/13/2022




#41 AR-1268-1

R.T.: 7.631 min
Delta R.T.: 0.000 min
Response: 138409646
Conc: 753.86 ng/ml




#42 AR-1268-2

R.T.: 8.786 min
Delta R.T.: 0.000 min
Response: 135221155
Conc: 748.55 ng/ml



#42 AR-1268-2

R.T.: 7.696 min
Delta R.T.: 0.000 min
Response: 126185354
Conc: 754.71 ng/ml




#43 AR-1268-3

R.T.: 9.021 min
Delta R.T.: 0.000 min
Response: 119354485
Conc: 750.17 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1268ICC750


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/12/2022
Supervised By :Ankita Jodhani 10/13/2022




#43 AR-1268-3

R.T.: 7.905 min
Delta R.T.: 0.000 min
Response: 107033368
Conc: 750.87 ng/ml




#44 AR-1268-4

R.T.: 9.456 min
Delta R.T.: 0.000 min
Response: 54838810
Conc: 753.27 ng/ml



#44 AR-1268-4

R.T.: 8.196 min
Delta R.T.: 0.000 min
Response: 43528455
Conc: 751.73 ng/ml




#45 AR-1268-5

R.T.: 9.885 min
Delta R.T.: 0.000 min
Response: 389788577
Conc: 750.92 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1268ICC750

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/12/2022
Supervised By :Ankita Jodhani 10/13/2022



#45 AR-1268-5

R.T.: 8.497 min
Delta R.T.: 0.000 min
Response: 329518769
Conc: 762.64 ng/ml m