

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP101123\
 Data File : PP060942.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 12 Oct 2023 02:11
 Operator : YP\AJ
 Sample : AR1660CCC500
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1660CCC500

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 10/12/2023
 Supervised By :Ankita Jodhani 10/12/2023

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 12 03:18:15 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP092623.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Sep 27 11:30:16 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.538	3.687	122.0E6	85932678	54.464	57.731m
2) SA Decachlor...	10.481	8.756	91366015	75605264	51.750m	46.350
Target Compounds						
3) L1 AR-1016-1	5.727	4.787	38738723	29310864	526.308	504.406
4) L1 AR-1016-2	5.751	4.806	54615601	40293905	521.474	500.511
5) L1 AR-1016-3	5.814	4.983	33841163	22543352	517.795	504.409
6) L1 AR-1016-4	5.914	5.026	28552164	17622417	525.054	489.733
7) L1 AR-1016-5	6.213	5.241	28738708	23712730	516.672	488.017
31) L7 AR-1260-1	7.354	6.283	52703897	44880578	502.781	447.128
32) L7 AR-1260-2	7.612	6.473	59778890	52749670	494.424	442.143
33) L7 AR-1260-3	7.976	6.628	47118747	51142780	508.346	446.683
34) L7 AR-1260-4	8.208	7.102	53946664	42134467	508.687	442.674
35) L7 AR-1260-5	8.544	7.345	98427345	94789784	508.613	437.396

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP101123\
Data File : PP060942.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 12 Oct 2023 02:11
Operator : YP\AJ
Sample : AR1660CCC500
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/12/2023
Supervised By :Ankita Jodhani 10/12/2023

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 12 03:18:15 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP092623.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Sep 27 11:30:16 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

