

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP101321\
 Data File : PP040038.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 13 Oct 2021 18:32
 Operator : AJ\MA
 Sample : I.BLK
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 14 02:31:28 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP101221.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Oct 13 08:03:59 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.896	3.984	713461	391330	23.494	23.147
2)	SA Decachlor...	10.916	9.314	589466	525820	23.849	24.776
Target Compounds							
10)	L2 AR-1221-3	0.000	4.423	0	8851	N.D.	16.344 #
11)	L3 AR-1232-1	0.000	4.423	0	8851	N.D.	21.792 #
28)	L6 AR-1254-3	7.760f	0.000	27489	0	19.295	N.D. #
32)	L7 AR-1260-2	0.000	7.020f	0	23294	N.D.	19.673 #
33)	L7 AR-1260-3	8.557f	7.196	6960	25422	7.638	21.516 #
36)	L8 AR-1262-1	8.557f	0.000	6960	0	5.104	N.D. #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP101321\
Data File : PP040038.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 13 Oct 2021 18:32
Operator : AJ\MA
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 14 02:31:28 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP101221.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Oct 13 08:03:59 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.896 min
Delta R.T.: 0.003 min
Response: 713461
Conc: 23.49 ng/ml

Instrument :
ECD_P
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 3.984 min
Delta R.T.: 0.002 min
Response: 391330
Conc: 23.15 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.916 min
Delta R.T.: 0.005 min
Response: 589466
Conc: 23.85 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.314 min
Delta R.T.: 0.000 min
Response: 525820
Conc: 24.78 ng/ml



#10 AR-1221-3


R.T.: 0.000 min
Exp R.T.: 5.331 min
Response: 0
Conc: N.D.

Instrument :
ECD_P
ClientSampleId :




#10 AR-1221-3

R.T.: 4.423 min
Delta R.T.: -0.009 min
Response: 8851
Conc: 16.34 ng/ml




#11 AR-1232-1

R.T.: 0.000 min
Exp R.T.: 5.330 min
Response: 0
Conc: N.D.



#11 AR-1232-1


R.T.: 4.423 min
Delta R.T.: -0.008 min
Response: 8851
Conc: 21.79 ng/ml



#28 AR-1254-3


R.T.: 7.760 min
Delta R.T.: 0.019 min
Response: 27489
Conc: 19.30 ng/ml

Instrument :
ECD_P
ClientSampleId :




#28 AR-1254-3

R.T.: 0.000 min
Exp R.T. : 6.706 min
Response: 0
Conc: N.D.




#32 AR-1260-2

R.T.: 0.000 min
Exp R.T. : 8.175 min
Response: 0
Conc: N.D.



#32 AR-1260-2


R.T.: 7.020 min
Delta R.T.: -0.022 min
Response: 23294
Conc: 19.67 ng/ml



#33 AR-1260-3


R.T.: 8.557 min
Delta R.T.: 0.017 min
Response: 6960
Conc: 7.64 ng/ml

Instrument :
ECD_P
ClientSampleId :




#33 AR-1260-3

R.T.: 7.196 min
Delta R.T.: -0.003 min
Response: 25422
Conc: 21.52 ng/ml



#36 AR-1262-1

R.T.: 8.557 min
Delta R.T.: 0.017 min
Response: 6960
Conc: 5.10 ng/ml



#36 AR-1262-1

R.T.: 0.000 min
Exp R.T.: 7.376 min
Response: 0
Conc: N.D.