

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP101321\
 Data File : PP040015.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 13 Oct 2021 11:27
 Operator : AJ\MA
 Sample : PB139837BS
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 PB139837BS

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 14 02:25:14 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP101221.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Oct 13 08:03:59 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.895	3.983	667825	375846	21.991	22.232
2) SA Decachlor...	10.914	9.312	574498	546249	23.243	25.739
Target Compounds						
3) L1 AR-1016-1	6.218	5.247	473776	310597	479.816	475.938
4) L1 AR-1016-2	6.242	5.267	689709	432642	482.654	473.230
5) L1 AR-1016-3	6.308	5.461	428268	238170	491.432	470.766
6) L1 AR-1016-4	6.416	5.512	351843	192366	494.480	469.813
7) L1 AR-1016-5	6.731	5.743	328200	244098	513.135	487.549
31) L7 AR-1260-1	7.909	6.844	551392	493153	504.757	492.346
32) L7 AR-1260-2	8.176	7.041	649235	587369	475.015	496.044
33) L7 AR-1260-3	8.541	7.198	423479	556449	464.723	470.966
34) L7 AR-1260-4	8.772	7.683	529563	422532	501.715	480.740
35) L7 AR-1260-5	9.097	7.931	1008402	979522	486.949	479.288



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP101321\
Data File : PP040015.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 13 Oct 2021 11:27
Operator : AJ\MA
Sample : PB139837BS
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
PB139837BS

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 14 02:25:14 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP101221.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Oct 13 08:03:59 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.895 min
Delta R.T.: 0.002 min
Response: 667825
Conc: 21.99 ng/ml

Instrument :
ECD_P
ClientSampleId :
PB139837BS




#1 Tetrachloro-m-xylene

R.T.: 3.983 min
Delta R.T.: 0.000 min
Response: 375846
Conc: 22.23 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.914 min
Delta R.T.: 0.003 min
Response: 574498
Conc: 23.24 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.312 min
Delta R.T.: 0.000 min
Response: 546249
Conc: 25.74 ng/ml



#3 AR-1016-1


R.T.: 6.218 min
Delta R.T.: 0.000 min
Response: 473776
Conc: 479.82 ng/ml

Instrument :
ECD_P
ClientSampleId :
PB139837BS




#3 AR-1016-1

R.T.: 5.247 min
Delta R.T.: 0.000 min
Response: 310597
Conc: 475.94 ng/ml




#4 AR-1016-2

R.T.: 6.242 min
Delta R.T.: 0.001 min
Response: 689709
Conc: 482.65 ng/ml



#4 AR-1016-2


R.T.: 5.267 min
Delta R.T.: 0.000 min
Response: 432642
Conc: 473.23 ng/ml



#5 AR-1016-3


R.T.: 6.308 min
Delta R.T.: 0.001 min
Response: 428268
Conc: 491.43 ng/ml

Instrument :
ECD_P
ClientSampleId :
PB139837BS




#5 AR-1016-3

R.T.: 5.461 min
Delta R.T.: 0.000 min
Response: 238170
Conc: 470.77 ng/ml




#6 AR-1016-4

R.T.: 6.416 min
Delta R.T.: 0.002 min
Response: 351843
Conc: 494.48 ng/ml



#6 AR-1016-4


R.T.: 5.512 min
Delta R.T.: 0.000 min
Response: 192366
Conc: 469.81 ng/ml



#7 AR-1016-5


R.T.: 6.731 min
Delta R.T.: 0.001 min
Response: 328200
Conc: 513.14 ng/ml

Instrument :
ECD_P
ClientSampleId :
PB139837BS




#7 AR-1016-5

R.T.: 5.743 min
Delta R.T.: 0.000 min
Response: 244098
Conc: 487.55 ng/ml




#31 AR-1260-1

R.T.: 7.909 min
Delta R.T.: 0.000 min
Response: 551392
Conc: 504.76 ng/ml



#31 AR-1260-1


R.T.: 6.844 min
Delta R.T.: 0.000 min
Response: 493153
Conc: 492.35 ng/ml



#32 AR-1260-2


R.T.: 8.176 min
Delta R.T.: 0.000 min
Response: 649235
Conc: 475.02 ng/ml

Instrument :
ECD_P
ClientSampleId :
PB139837BS




#32 AR-1260-2

R.T.: 7.041 min
Delta R.T.: 0.000 min
Response: 587369
Conc: 496.04 ng/ml




#33 AR-1260-3

R.T.: 8.541 min
Delta R.T.: 0.002 min
Response: 423479
Conc: 464.72 ng/ml



#33 AR-1260-3


R.T.: 7.198 min
Delta R.T.: 0.000 min
Response: 556449
Conc: 470.97 ng/ml



#34 AR-1260-4


R.T.: 8.772 min
Delta R.T.: 0.000 min
Response: 529563
Conc: 501.71 ng/ml

Instrument :
ECD_P
ClientSampleId :
PB139837BS




#34 AR-1260-4

R.T.: 7.683 min
Delta R.T.: 0.000 min
Response: 422532
Conc: 480.74 ng/ml



#35 AR-1260-5

R.T.: 9.097 min
Delta R.T.: 0.000 min
Response: 1008402
Conc: 486.95 ng/ml



#35 AR-1260-5

R.T.: 7.931 min
Delta R.T.: 0.000 min
Response: 979522
Conc: 479.29 ng/ml