

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP101323\
 Data File : PP060973.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 13 Oct 2023 22:06
 Operator : YP\AJ
 Sample : AR1268ICC750
 Misc :
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 14 13:08:26 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP101323.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sat Oct 14 13:06:31 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.551	3.688	181.4E6	105.5E6	74.685	72.693
2)	SA Decachlor...	10.510	8.756	233.6E6	197.3E6	73.889	73.194
Target Compounds							
41)	L9 AR-1268-1	8.892	7.631	199.2E6	178.0E6	738.615	735.996
42)	L9 AR-1268-2	8.994	7.696	180.3E6	162.3E6	739.225	743.651
43)	L9 AR-1268-3	9.237	7.905	157.2E6	140.6E6	739.812	734.696
44)	L9 AR-1268-4	9.692	8.196	62896568	57631155	740.957	732.909
45)	L9 AR-1268-5	10.141	8.494	500.7E6	422.3E6	749.967	742.962



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP101323\
Data File : PP060973.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 13 Oct 2023 22:06
Operator : YP\AJ
Sample : AR1268ICC750
Misc :
ALS Vial : 27 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 14 13:08:26 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP101323.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat Oct 14 13:06:31 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.551 min
Delta R.T.: 0.000 min
Response: 181399385
Conc: 74.69 ng/ml

Instrument :
ECD_P
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 3.688 min
Delta R.T.: 0.000 min
Response: 105525709
Conc: 72.69 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.510 min
Delta R.T.: 0.000 min
Response: 233564794
Conc: 73.89 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.756 min
Delta R.T.: 0.000 min
Response: 197336673
Conc: 73.19 ng/ml



#41 AR-1268-1


R.T.: 8.892 min
Delta R.T.: 0.000 min
Response: 199152106
Conc: 738.62 ng/ml

Instrument :
ECD_P
ClientSampleId :




#41 AR-1268-1

R.T.: 7.631 min
Delta R.T.: 0.000 min
Response: 178012172
Conc: 736.00 ng/ml




#42 AR-1268-2

R.T.: 8.994 min
Delta R.T.: 0.000 min
Response: 180328238
Conc: 739.23 ng/ml



#42 AR-1268-2


R.T.: 7.696 min
Delta R.T.: 0.000 min
Response: 162316039
Conc: 743.65 ng/ml



#43 AR-1268-3


R.T.: 9.237 min
Delta R.T.: 0.000 min
Response: 157237421
Conc: 739.81 ng/ml

Instrument :
ECD_P
ClientSampleId :




#43 AR-1268-3

R.T.: 7.905 min
Delta R.T.: 0.000 min
Response: 140564792
Conc: 734.70 ng/ml




#44 AR-1268-4

R.T.: 9.692 min
Delta R.T.: 0.000 min
Response: 62896568
Conc: 740.96 ng/ml



#44 AR-1268-4


R.T.: 8.196 min
Delta R.T.: 0.000 min
Response: 57631155
Conc: 732.91 ng/ml



#45 AR-1268-5

R.T.: 10.141 min
Delta R.T.: -0.001 min
Response: 500680966
Conc: 749.97 ng/ml

Instrument :
ECD_P
ClientSampleId :



#45 AR-1268-5

R.T.: 8.494 min
Delta R.T.: 0.000 min
Response: 422250691
Conc: 742.96 ng/ml