

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP101421\
 Data File : PP040069.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 14 Oct 2021 17:12
 Operator : AJ\MA
 Sample : M4191-01MSD
 Misc :
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 15 06:37:50 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP101221.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Oct 13 08:03:59 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.895	3.983	681398	368833	22.438	21.817
2)	SA Decachlor...	10.915	9.311	558959	511021	22.615	24.079
Target Compounds							
3)	L1 AR-1016-1	6.218	5.247	551581	339443	558.613	520.140
4)	L1 AR-1016-2	6.242	5.266	811368	473226	567.789	517.621
5)	L1 AR-1016-3	6.308	5.460	508533	263211	583.535	520.264
6)	L1 AR-1016-4	6.415	5.512	432810	210116	608.272	513.164
7)	L1 AR-1016-5	6.730	5.743	381698	263482	596.779	526.266
8)	L2 AR-1221-1	5.144	4.242	78409	36769	225.146	179.670
9)	L2 AR-1221-2	5.244	4.343	66905	45270	300.300	269.132
10)	L2 AR-1221-3	5.332	4.432	311704	185938	386.642	343.351
11)	L3 AR-1232-1	5.332	4.432	311704	185938	500.933	457.796
12)	L3 AR-1232-2	5.922	5.266	425006	473226	1238.294	1210.499
13)	L3 AR-1232-3	6.242	5.460	811368	263211	1283.627	1270.719
14)	L3 AR-1232-4	6.415	5.557	432810	262484	1379.993	1387.259
15)	L3 AR-1232-5	6.513	5.743	317718	263482	1354.925	1276.182
16)	L4 AR-1242-1	6.218	5.247	551581	339443	702.488	682.189
17)	L4 AR-1242-2	6.242	5.266	811368	473226	715.307	692.866
18)	L4 AR-1242-3	6.308	5.460	508533	263211	731.691	697.629
19)	L4 AR-1242-4	6.415	5.557	432810	262484	751.156	708.599
20)	L4 AR-1242-5	7.200	6.124	43994	222074	84.118	465.599 #
21)	L5 AR-1248-1	6.218	5.247	551581	339443	894.119	870.643
22)	L5 AR-1248-2	6.513	5.512	317718	210116	396.891	390.915
23)	L5 AR-1248-3	6.730	5.557	381698	262484	425.693	464.336
24)	L5 AR-1248-4	7.160	5.743	52079	263482	56.593	412.294 #
25)	L5 AR-1248-5	7.200	6.166	43994	31742	49.733	50.882
26)	L6 AR-1254-1	7.129	6.124	248184	222074	267.498	231.804
27)	L6 AR-1254-2	7.358	6.283	277709	215422	200.603	244.077
28)	L6 AR-1254-3	7.742	6.724f	165114	393404	115.897	288.553 #
29)	L6 AR-1254-4	8.038	6.946	96179	52256	90.890	60.924 #
30)	L6 AR-1254-5	8.468	7.377	874669	758224	748.322	587.402
31)	L7 AR-1260-1	7.909	6.843	596703	547030	546.236	546.134
32)	L7 AR-1260-2	8.175	7.040	713861	644572	522.299	544.353
33)	L7 AR-1260-3	8.541	7.197	471267	602181	517.166	509.673
34)	L7 AR-1260-4	8.772	7.682	567111	458611	537.289	521.789
35)	L7 AR-1260-5	9.097	7.929	1072524	1041533	517.913	509.630
36)	L8 AR-1262-1	8.541	7.377	471267	758224	345.589	1079.365 #
37)	L8 AR-1262-2	9.097	7.929	1072524	1041533	432.491	462.817
38)	L8 AR-1262-3	9.429f	8.218	736180	224171	640.103	230.182 #
39)	L8 AR-1262-4	9.481f	8.281	422208	789400	516.724	445.281
40)	L8 AR-1262-5	10.167	8.781	285174	272222	300.149	300.609
41)	L9 AR-1268-1	9.429f	8.218	736180	224171	249.294	83.079 #

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP101421\
 Data File : PP040069.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 14 Oct 2021 17:12
 Operator : AJ\MA
 Sample : M4191-01MSD
 Misc :
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 15 06:37:50 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP101221.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Oct 13 08:03:59 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
42) L9	AR-1268-2	9.481f	8.281	422208	789400	149.541	315.290 #
43) L9	AR-1268-3	0.000	8.492	0	15653	N.D.	7.148 #
44) L9	AR-1268-4	10.167	8.781	285174	272222	276.702	274.519
45) L9	AR-1268-5	10.580	9.064	137699	97767	16.009	13.714



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP101421\
Data File : PP040069.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 14 Oct 2021 17:12
Operator : AJ\MA
Sample : M4191-01MSD
Misc :
ALS Vial : 18 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampled :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 15 06:37:50 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP101221.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Oct 13 08:03:59 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.895 min
Delta R.T.: 0.002 min
Response: 681398
Conc: 22.44 ng/ml

Instrument :
ECD_P
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 3.983 min
Delta R.T.: 0.000 min
Response: 368833
Conc: 21.82 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.915 min
Delta R.T.: 0.004 min
Response: 558959
Conc: 22.61 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.311 min
Delta R.T.: -0.001 min
Response: 511021
Conc: 24.08 ng/ml



#3 AR-1016-1


R.T.: 6.218 min
Delta R.T.: 0.000 min
Response: 551581
Conc: 558.61 ng/ml

Instrument :
ECD_P
ClientSampleId :




#3 AR-1016-1

R.T.: 5.247 min
Delta R.T.: 0.000 min
Response: 339443
Conc: 520.14 ng/ml




#4 AR-1016-2

R.T.: 6.242 min
Delta R.T.: 0.001 min
Response: 811368
Conc: 567.79 ng/ml



#4 AR-1016-2


R.T.: 5.266 min
Delta R.T.: 0.000 min
Response: 473226
Conc: 517.62 ng/ml



#5 AR-1016-3


R.T.: 6.308 min
Delta R.T.: 0.000 min
Response: 508533
Conc: 583.54 ng/ml

Instrument :
ECD_P
ClientSampleId :




#5 AR-1016-3

R.T.: 5.460 min
Delta R.T.: 0.000 min
Response: 263211
Conc: 520.26 ng/ml




#6 AR-1016-4

R.T.: 6.415 min
Delta R.T.: 0.000 min
Response: 432810
Conc: 608.27 ng/ml



#6 AR-1016-4


R.T.: 5.512 min
Delta R.T.: 0.000 min
Response: 210116
Conc: 513.16 ng/ml



#7 AR-1016-5


R.T.: 6.730 min
Delta R.T.: 0.000 min
Response: 381698
Conc: 596.78 ng/ml

Instrument :
ECD_P
ClientSampleId :




#7 AR-1016-5

R.T.: 5.743 min
Delta R.T.: 0.000 min
Response: 263482
Conc: 526.27 ng/ml




#8 AR-1221-1

R.T.: 5.144 min
Delta R.T.: -0.001 min
Response: 78409
Conc: 225.15 ng/ml



#8 AR-1221-1


R.T.: 4.242 min
Delta R.T.: -0.001 min
Response: 36769
Conc: 179.67 ng/ml



#9 AR-1221-2


R.T.: 5.244 min
Delta R.T.: 0.000 min
Response: 66905
Conc: 300.30 ng/ml

Instrument :
ECD_P
ClientSampleId :




#9 AR-1221-2

R.T.: 4.343 min
Delta R.T.: 0.000 min
Response: 45270
Conc: 269.13 ng/ml




#10 AR-1221-3

R.T.: 5.332 min
Delta R.T.: 0.001 min
Response: 311704
Conc: 386.64 ng/ml



#10 AR-1221-3


R.T.: 4.432 min
Delta R.T.: 0.000 min
Response: 185938
Conc: 343.35 ng/ml



#11 AR-1232-1


R.T.: 5.332 min
Delta R.T.: 0.002 min
Response: 311704
Conc: 500.93 ng/ml

Instrument :
ECD_P
ClientSampleId :




#11 AR-1232-1

R.T.: 4.432 min
Delta R.T.: 0.000 min
Response: 185938
Conc: 457.80 ng/ml




#12 AR-1232-2

R.T.: 5.922 min
Delta R.T.: 0.001 min
Response: 425006
Conc: 1238.29 ng/ml



#12 AR-1232-2


R.T.: 5.266 min
Delta R.T.: 0.000 min
Response: 473226
Conc: 1210.50 ng/ml



#13 AR-1232-3


R.T.: 6.242 min
Delta R.T.: 0.001 min
Response: 811368
Conc: 1283.63 ng/ml

Instrument :
ECD_P
ClientSampleId :




#13 AR-1232-3

R.T.: 5.460 min
Delta R.T.: 0.000 min
Response: 263211
Conc: 1270.72 ng/ml




#14 AR-1232-4

R.T.: 6.415 min
Delta R.T.: 0.000 min
Response: 432810
Conc: 1379.99 ng/ml



#14 AR-1232-4


R.T.: 5.557 min
Delta R.T.: 0.000 min
Response: 262484
Conc: 1387.26 ng/ml



#15 AR-1232-5


R.T.: 6.513 min
Delta R.T.: 0.001 min
Response: 317718
Conc: 1354.93 ng/ml

Instrument :
ECD_P
ClientSampleId :




#15 AR-1232-5

R.T.: 5.743 min
Delta R.T.: 0.000 min
Response: 263482
Conc: 1276.18 ng/ml




#16 AR-1242-1

R.T.: 6.218 min
Delta R.T.: 0.001 min
Response: 551581
Conc: 702.49 ng/ml



#16 AR-1242-1


R.T.: 5.247 min
Delta R.T.: 0.000 min
Response: 339443
Conc: 682.19 ng/ml



#17 AR-1242-2


R.T.: 6.242 min
Delta R.T.: 0.000 min
Response: 811368
Conc: 715.31 ng/ml

Instrument :
ECD_P
ClientSampleId :




#17 AR-1242-2

R.T.: 5.266 min
Delta R.T.: 0.000 min
Response: 473226
Conc: 692.87 ng/ml




#18 AR-1242-3

R.T.: 6.308 min
Delta R.T.: 0.002 min
Response: 508533
Conc: 731.69 ng/ml



#18 AR-1242-3


R.T.: 5.460 min
Delta R.T.: 0.000 min
Response: 263211
Conc: 697.63 ng/ml



#19 AR-1242-4


R.T.: 6.415 min
Delta R.T.: 0.001 min
Response: 432810
Conc: 751.16 ng/ml

Instrument :
ECD_P
ClientSampleId :




#19 AR-1242-4

R.T.: 5.557 min
Delta R.T.: 0.000 min
Response: 262484
Conc: 708.60 ng/ml




#20 AR-1242-5

R.T.: 7.200 min
Delta R.T.: 0.002 min
Response: 43994
Conc: 84.12 ng/ml



#20 AR-1242-5


R.T.: 6.124 min
Delta R.T.: 0.001 min
Response: 222074
Conc: 465.60 ng/ml



#21 AR-1248-1


R.T.: 6.218 min
Delta R.T.: 0.000 min
Response: 551581
Conc: 894.12 ng/ml

Instrument :
ECD_P
ClientSampleId :




#21 AR-1248-1

R.T.: 5.247 min
Delta R.T.: 0.000 min
Response: 339443
Conc: 870.64 ng/ml




#22 AR-1248-2

R.T.: 6.513 min
Delta R.T.: 0.002 min
Response: 317718
Conc: 396.89 ng/ml



#22 AR-1248-2


R.T.: 5.512 min
Delta R.T.: 0.000 min
Response: 210116
Conc: 390.92 ng/ml



#23 AR-1248-3


R.T.: 6.730 min
Delta R.T.: 0.001 min
Response: 381698
Conc: 425.69 ng/ml

Instrument :
ECD_P
ClientSampleId :




#23 AR-1248-3

R.T.: 5.557 min
Delta R.T.: 0.000 min
Response: 262484
Conc: 464.34 ng/ml




#24 AR-1248-4

R.T.: 7.160 min
Delta R.T.: 0.001 min
Response: 52079
Conc: 56.59 ng/ml



#24 AR-1248-4


R.T.: 5.743 min
Delta R.T.: 0.000 min
Response: 263482
Conc: 412.29 ng/ml



#25 AR-1248-5


R.T.: 7.200 min
Delta R.T.: 0.002 min
Response: 43994
Conc: 49.73 ng/ml

Instrument :
ECD_P
ClientSampleId :




#25 AR-1248-5

R.T.: 6.166 min
Delta R.T.: 0.000 min
Response: 31742
Conc: 50.88 ng/ml




#26 AR-1254-1

R.T.: 7.129 min
Delta R.T.: 0.001 min
Response: 248184
Conc: 267.50 ng/ml



#26 AR-1254-1


R.T.: 6.124 min
Delta R.T.: 0.000 min
Response: 222074
Conc: 231.80 ng/ml



#27 AR-1254-2


R.T.: 7.358 min
Delta R.T.: 0.000 min
Response: 277709
Conc: 200.60 ng/ml

Instrument :
ECD_P
ClientSampleId :




#27 AR-1254-2

R.T.: 6.283 min
Delta R.T.: 0.000 min
Response: 215422
Conc: 244.08 ng/ml




#28 AR-1254-3

R.T.: 7.742 min
Delta R.T.: 0.002 min
Response: 165114
Conc: 115.90 ng/ml



#28 AR-1254-3


R.T.: 6.724 min
Delta R.T.: 0.017 min
Response: 393404
Conc: 288.55 ng/ml



#29 AR-1254-4


R.T.: 8.038 min
Delta R.T.: 0.000 min
Response: 96179
Conc: 90.89 ng/ml

Instrument :
ECD_P
ClientSampleId :




#29 AR-1254-4

R.T.: 6.946 min
Delta R.T.: 0.000 min
Response: 52256
Conc: 60.92 ng/ml




#30 AR-1254-5

R.T.: 8.468 min
Delta R.T.: 0.000 min
Response: 874669
Conc: 748.32 ng/ml



#30 AR-1254-5


R.T.: 7.377 min
Delta R.T.: 0.000 min
Response: 758224
Conc: 587.40 ng/ml



#31 AR-1260-1


R.T.: 7.909 min
Delta R.T.: 0.000 min
Response: 596703
Conc: 546.24 ng/ml

Instrument :
ECD_P
ClientSampleId :




#31 AR-1260-1

R.T.: 6.843 min
Delta R.T.: -0.001 min
Response: 547030
Conc: 546.13 ng/ml




#32 AR-1260-2

R.T.: 8.175 min
Delta R.T.: 0.000 min
Response: 713861
Conc: 522.30 ng/ml



#32 AR-1260-2


R.T.: 7.040 min
Delta R.T.: -0.001 min
Response: 644572
Conc: 544.35 ng/ml



#33 AR-1260-3


R.T.: 8.541 min
Delta R.T.: 0.002 min
Response: 471267
Conc: 517.17 ng/ml

Instrument :
ECD_P
ClientSampleId :




#33 AR-1260-3

R.T.: 7.197 min
Delta R.T.: -0.001 min
Response: 602181
Conc: 509.67 ng/ml




#34 AR-1260-4

R.T.: 8.772 min
Delta R.T.: 0.000 min
Response: 567111
Conc: 537.29 ng/ml



#34 AR-1260-4


R.T.: 7.682 min
Delta R.T.: -0.001 min
Response: 458611
Conc: 521.79 ng/ml



#35 AR-1260-5


R.T.: 9.097 min
Delta R.T.: 0.000 min
Response: 1072524
Conc: 517.91 ng/ml

Instrument :
ECD_P
ClientSampleId :




#35 AR-1260-5

R.T.: 7.929 min
Delta R.T.: -0.001 min
Response: 1041533
Conc: 509.63 ng/ml




#36 AR-1262-1

R.T.: 8.541 min
Delta R.T.: 0.002 min
Response: 471267
Conc: 345.59 ng/ml



#36 AR-1262-1


R.T.: 7.377 min
Delta R.T.: 0.000 min
Response: 758224
Conc: 1079.36 ng/ml



#37 AR-1262-2


R.T.: 9.097 min
Delta R.T.: 0.001 min
Response: 1072524
Conc: 432.49 ng/ml

Instrument :
ECD_P
ClientSampleId :




#37 AR-1262-2

R.T.: 7.929 min
Delta R.T.: 0.000 min
Response: 1041533
Conc: 462.82 ng/ml




#38 AR-1262-3

R.T.: 9.429 min
Delta R.T.: 0.024 min
Response: 736180
Conc: 640.10 ng/ml



#38 AR-1262-3


R.T.: 8.218 min
Delta R.T.: 0.000 min
Response: 224171
Conc: 230.18 ng/ml



#39 AR-1262-4


R.T.: 9.481 min
Delta R.T.: -0.017 min
Response: 422208
Conc: 516.72 ng/ml

Instrument :
ECD_P
ClientSampleId :




#39 AR-1262-4

R.T.: 8.281 min
Delta R.T.: 0.000 min
Response: 789400
Conc: 445.28 ng/ml




#40 AR-1262-5

R.T.: 10.167 min
Delta R.T.: 0.003 min
Response: 285174
Conc: 300.15 ng/ml



#40 AR-1262-5


R.T.: 8.781 min
Delta R.T.: 0.001 min
Response: 272222
Conc: 300.61 ng/ml



#41 AR-1268-1


R.T.: 9.429 min
Delta R.T.: 0.026 min
Response: 736180
Conc: 249.29 ng/ml

Instrument :
ECD_P
ClientSampleId :




#41 AR-1268-1

R.T.: 8.218 min
Delta R.T.: 0.000 min
Response: 224171
Conc: 83.08 ng/ml




#42 AR-1268-2

R.T.: 9.481 min
Delta R.T.: -0.019 min
Response: 422208
Conc: 149.54 ng/ml



#42 AR-1268-2


R.T.: 8.281 min
Delta R.T.: -0.002 min
Response: 789400
Conc: 315.29 ng/ml



#43 AR-1268-3


R.T.: 0.000 min
Exp R.T.: 9.725 min
Response: 0
Conc: N.D.

Instrument :
ECD_P
ClientSampleId :




#43 AR-1268-3

R.T.: 8.492 min
Delta R.T.: 0.000 min
Response: 15653
Conc: 7.15 ng/ml




#44 AR-1268-4

R.T.: 10.167 min
Delta R.T.: 0.003 min
Response: 285174
Conc: 276.70 ng/ml



#44 AR-1268-4


R.T.: 8.781 min
Delta R.T.: 0.000 min
Response: 272222
Conc: 274.52 ng/ml



#45 AR-1268-5

R.T.: 10.580 min
Delta R.T.: 0.004 min
Response: 137699
Conc: 16.01 ng/ml

Instrument :
ECD_P
ClientSampleId :



#45 AR-1268-5

R.T.: 9.064 min
Delta R.T.: 0.000 min
Response: 97767
Conc: 13.71 ng/ml