

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP101421\
 Data File : PP040070.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 14 Oct 2021 17:28
 Operator : AJ\MA
 Sample : M4191-02
 Misc :
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 COMP-5

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 15 06:38:19 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP101221.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Oct 13 08:03:59 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.896	3.984	735864	412620	24.231	24.407
2) SA Decachlor...	10.918	9.313	528299	483572	21.374	22.785

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP101421\
Data File : PP040070.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 14 Oct 2021 17:28
Operator : AJ\MA
Sample : M4191-02
Misc :
ALS Vial : 19 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
COMP-5

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 15 06:38:19 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP101221.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Oct 13 08:03:59 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.896 min
Delta R.T.: 0.003 min
Response: 735864
Conc: 24.23 ng/ml

Instrument :
ECD_P
ClientSampleId :
COMP-5




#1 Tetrachloro-m-xylene

R.T.: 3.984 min
Delta R.T.: 0.001 min
Response: 412620
Conc: 24.41 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.918 min
Delta R.T.: 0.006 min
Response: 528299
Conc: 21.37 ng/ml



#2 Decachlorobiphenyl

R.T.: 9.313 min
Delta R.T.: 0.000 min
Response: 483572
Conc: 22.79 ng/ml