

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP101424\
 Data File : PP067815.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 14 Oct 2024 22:36
 Operator : YP\AJ
 Sample : AR1248CCC500
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1248CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 15 02:04:51 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP100824.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Oct 09 05:48:32 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.755	4.052	46535845	48440167	50.274	47.941
2) SA Decachlor...	10.668	9.217	55820137	52494618	48.484	46.809
Target Compounds						
21) L5 AR-1248-1	5.918	5.159	10845813	10957228	506.272	483.469
22) L5 AR-1248-2	6.193	5.400	16358525	15817077	486.619	477.918
23) L5 AR-1248-3	6.397	5.444	17681127	16526516	482.755	479.382
24) L5 AR-1248-4	6.798	5.618	19663014	19468459	486.098	475.352
25) L5 AR-1248-5	6.837	6.016	19166638	18131560	483.262	484.242



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP101424\
Data File : PP067815.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 14 Oct 2024 22:36
Operator : YP\AJ
Sample : AR1248CCC500
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 15 02:04:51 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP100824.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Oct 09 05:48:32 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.755 min
Delta R.T.: 0.000 min
Response: 46535845
Conc: 50.27 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500




#1 Tetrachloro-m-xylene

R.T.: 4.052 min
Delta R.T.: 0.000 min
Response: 48440167
Conc: 47.94 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.668 min
Delta R.T.: -0.001 min
Response: 55820137
Conc: 48.48 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.217 min
Delta R.T.: -0.003 min
Response: 52494618
Conc: 46.81 ng/ml



#21 AR-1248-1


R.T.: 5.918 min
Delta R.T.: 0.001 min
Response: 10845813
Conc: 506.27 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500




#21 AR-1248-1

R.T.: 5.159 min
Delta R.T.: 0.000 min
Response: 10957228
Conc: 483.47 ng/ml




#22 AR-1248-2

R.T.: 6.193 min
Delta R.T.: 0.002 min
Response: 16358525
Conc: 486.62 ng/ml



#22 AR-1248-2


R.T.: 5.400 min
Delta R.T.: 0.000 min
Response: 15817077
Conc: 477.92 ng/ml



#23 AR-1248-3


R.T.: 6.397 min
Delta R.T.: 0.001 min
Response: 17681127
Conc: 482.76 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500




#23 AR-1248-3

R.T.: 5.444 min
Delta R.T.: 0.000 min
Response: 16526516
Conc: 479.38 ng/ml




#24 AR-1248-4

R.T.: 6.798 min
Delta R.T.: 0.002 min
Response: 19663014
Conc: 486.10 ng/ml



#24 AR-1248-4


R.T.: 5.618 min
Delta R.T.: 0.000 min
Response: 19468459
Conc: 475.35 ng/ml



#25 AR-1248-5

R.T.: 6.837 min
Delta R.T.: 0.000 min
Response: 19166638
Conc: 483.26 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500



#25 AR-1248-5

R.T.: 6.016 min
Delta R.T.: 0.000 min
Response: 18131560
Conc: 484.24 ng/ml