

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP101722\
 Data File : PP052378.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 17 Oct 2022 16:38
 Operator : YP\AJ
 Sample : AR1248CCC500
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1248CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 18 00:02:58 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP101122.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Oct 12 05:20:17 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.423	3.674	79407350	62219964	50.518	54.310
2) SA Decachlor...	10.247	8.770	65529786	53502742	47.217	47.883
Target Compounds						
21) L5 AR-1248-1	5.598	4.777	13254324	13448567	493.044	496.048
22) L5 AR-1248-2	5.874	5.017	22429612	18133133	482.540	487.391
23) L5 AR-1248-3	6.079	5.060	26953162	18531511	486.765	486.438
24) L5 AR-1248-4	6.485	5.235	29383177	21988835	462.779	474.561
25) L5 AR-1248-5	6.524	5.635	28921949	20015839	466.973	475.963



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP101722\
Data File : PP052378.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 17 Oct 2022 16:38
Operator : YP\AJ
Sample : AR1248CCC500
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 18 00:02:58 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP101122.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Oct 12 05:20:17 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.423 min
Delta R.T.: -0.003 min
Response: 79407350
Conc: 50.52 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.674 min
Delta R.T.: -0.002 min
Response: 62219964
Conc: 54.31 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.247 min
Delta R.T.: 0.011 min
Response: 65529786
Conc: 47.22 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.770 min
Delta R.T.: 0.004 min
Response: 53502742
Conc: 47.88 ng/ml



#21 AR-1248-1


R.T.: 5.598 min
Delta R.T.: 0.002 min
Response: 13254324
Conc: 493.04 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500




#21 AR-1248-1

R.T.: 4.777 min
Delta R.T.: 0.000 min
Response: 13448567
Conc: 496.05 ng/ml




#22 AR-1248-2

R.T.: 5.874 min
Delta R.T.: 0.002 min
Response: 22429612
Conc: 482.54 ng/ml



#22 AR-1248-2


R.T.: 5.017 min
Delta R.T.: 0.001 min
Response: 18133133
Conc: 487.39 ng/ml



#23 AR-1248-3


R.T.: 6.079 min
Delta R.T.: 0.003 min
Response: 26953162
Conc: 486.77 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500




#23 AR-1248-3

R.T.: 5.060 min
Delta R.T.: 0.001 min
Response: 18531511
Conc: 486.44 ng/ml




#24 AR-1248-4

R.T.: 6.485 min
Delta R.T.: 0.004 min
Response: 29383177
Conc: 462.78 ng/ml



#24 AR-1248-4


R.T.: 5.235 min
Delta R.T.: 0.000 min
Response: 21988835
Conc: 474.56 ng/ml



#25 AR-1248-5

R.T.: 6.524 min
Delta R.T.: 0.004 min
Response: 28921949
Conc: 466.97 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500



#25 AR-1248-5

R.T.: 5.635 min
Delta R.T.: 0.002 min
Response: 20015839
Conc: 475.96 ng/ml