

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP101723\
 Data File : PP061001.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 17 Oct 2023 09:27
 Operator : YP\AJ
 Sample : HEXANE
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 17 11:09:57 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP101323.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sat Oct 14 13:32:01 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
----------	------	------	--------	--------	-------	-------

System Monitoring Compounds

Target Compounds							
24)	L5	AR-1248-4	6.615f	0.000	386185	0	5.054 N.D. #
26)	L6	AR-1254-1	6.615	0.000	386185	0	4.883 N.D. #
28)	L6	AR-1254-3	7.211	0.000	900983	0	7.365 N.D. #
29)	L6	AR-1254-4	7.485	0.000	1127275	0	12.807 N.D. #
30)	L6	AR-1254-5	7.934f	0.000	370714	0	3.733 N.D. #
31)	L7	AR-1260-1	7.373	0.000	72927	0	0.779 N.D. #
33)	L7	AR-1260-3	8.007	0.000	1173923	0	16.891 N.D. #
36)	L8	AR-1262-1	8.007f	0.000	1173923	0	9.346 N.D. #
38)	L8	AR-1262-3	8.884	0.000	277080	0	1.845 N.D. #
39)	L8	AR-1262-4	9.012f	0.000	305512	0	2.701 N.D. #
41)	L9	AR-1268-1	8.884	0.000	277080	0	1.063 N.D. #
42)	L9	AR-1268-2	9.012f	0.000	305512	0	1.303 N.D. #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP101723\
Data File : PP061001.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 17 Oct 2023 09:27
Operator : YP\AJ
Sample : HEXANE
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 17 11:09:57 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP101323.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat Oct 14 13:32:01 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 0.000 min
Exp R.T. : 4.551 min
Response: 0
Conc: N.D.

Instrument :
ECD_P
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 0.000 min
Exp R.T. : 3.687 min
Response: 0
Conc: N.D.




#2 Decachlorobiphenyl

R.T.: 0.000 min
Exp R.T. : 10.507 min
Response: 0
Conc: N.D.



#2 Decachlorobiphenyl


R.T.: 0.000 min
Exp R.T. : 8.754 min
Response: 0
Conc: N.D.



#24 AR-1248-4


R.T.: 6.615 min
Delta R.T.: -0.021 min
Response: 386185
Conc: 5.05 ng/ml

Instrument :
ECD_P
ClientSampleId :




#24 AR-1248-4

R.T.: 0.000 min
Exp R.T. : 5.242 min
Response: 0
Conc: N.D.




#26 AR-1254-1

R.T.: 6.615 min
Delta R.T.: 0.004 min
Response: 386185
Conc: 4.88 ng/ml



#26 AR-1254-1


R.T.: 0.000 min
Exp R.T. : 5.597 min
Response: 0
Conc: N.D.



#28 AR-1254-3


R.T.: 7.211 min
Delta R.T.: 0.008 min
Response: 900983
Conc: 7.36 ng/ml

Instrument :
ECD_P
ClientSampleId :




#28 AR-1254-3

R.T.: 0.000 min
Exp R.T. : 6.153 min
Response: 0
Conc: N.D.




#29 AR-1254-4

R.T.: 7.485 min
Delta R.T.: -0.006 min
Response: 1127275
Conc: 12.81 ng/ml



#29 AR-1254-4


R.T.: 0.000 min
Exp R.T. : 6.382 min
Response: 0
Conc: N.D.



#30 AR-1254-5


R.T.: 7.934 min
Delta R.T.: 0.019 min
Response: 370714
Conc: 3.73 ng/ml

Instrument :
ECD_P
ClientSampleId :




#30 AR-1254-5

R.T.: 0.000 min
Exp R.T. : 6.803 min
Response: 0
Conc: N.D.




#31 AR-1260-1

R.T.: 7.373 min
Delta R.T.: 0.005 min
Response: 72927
Conc: 0.78 ng/ml



#31 AR-1260-1


R.T.: 0.000 min
Exp R.T. : 6.282 min
Response: 0
Conc: N.D.



#33 AR-1260-3


R.T.: 8.007 min
Delta R.T.: 0.017 min
Response: 1173923
Conc: 16.89 ng/ml

Instrument :
ECD_P
ClientSampleId :




#33 AR-1260-3

R.T.: 0.000 min
Exp R.T. : 6.626 min
Response: 0
Conc: N.D.




#36 AR-1262-1

R.T.: 8.007 min
Delta R.T.: 0.015 min
Response: 1173923
Conc: 9.35 ng/ml



#36 AR-1262-1


R.T.: 0.000 min
Exp R.T. : 6.896 min
Response: 0
Conc: N.D.



#38 AR-1262-3


R.T.: 8.884 min
Delta R.T.: -0.010 min
Response: 277080
Conc: 1.85 ng/ml

Instrument :
ECD_P
ClientSampleId :




#38 AR-1262-3

R.T.: 0.000 min
Exp R.T. : 7.632 min
Response: 0
Conc: N.D.




#39 AR-1262-4

R.T.: 9.012 min
Delta R.T.: 0.024 min
Response: 305512
Conc: 2.70 ng/ml



#39 AR-1262-4


R.T.: 0.000 min
Exp R.T. : 7.696 min
Response: 0
Conc: N.D.



#41 AR-1268-1


R.T.: 8.884 min
Delta R.T.: -0.010 min
Response: 277080
Conc: 1.06 ng/ml

Instrument :
ECD_P
ClientSampleId :




#41 AR-1268-1

R.T.: 0.000 min
Exp R.T. : 7.633 min
Response: 0
Conc: N.D.



#42 AR-1268-2

R.T.: 9.012 min
Delta R.T.: 0.018 min
Response: 305512
Conc: 1.30 ng/ml



#42 AR-1268-2

R.T.: 0.000 min
Exp R.T. : 7.697 min
Response: 0
Conc: N.D.