

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP101723\
 Data File : PP061015.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 17 Oct 2023 15:10
 Operator : YP\AJ
 Sample : 04908-07MS
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 TP-3MS

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 10/18/2023
 Supervised By :Ankita Jodhani 10/18/2023

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 17 15:50:06 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP101323.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sat Oct 14 13:32:01 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.553	3.688	49014625	28731599	22.296	21.251
2) SA Decachlor...	10.510	8.755	34987190	28715988	22.342	19.099
Target Compounds						
3) L1 AR-1016-1	5.744	4.787	38500720	25500840	536.298	489.984
4) L1 AR-1016-2	5.767	4.807	54815126	34869531	534.417	486.875
5) L1 AR-1016-3	5.830	4.984	33117712	19660409	529.967	486.557
6) L1 AR-1016-4	5.930	5.026	27494137	15526310	538.220m	480.005
7) L1 AR-1016-5	6.230	5.242	27361452	20043100	541.180	480.036
31) L7 AR-1260-1	7.370	6.284	47070590	37676447	502.861	455.300
32) L7 AR-1260-2	7.629	6.473	53009143	44417440	488.259	451.064
33) L7 AR-1260-3	7.993	6.627	34078625	42703584	490.327	451.567
34) L7 AR-1260-4	8.226	7.103	43611578	32388916	511.230	451.901
35) L7 AR-1260-5	8.563	7.345	81623343	76864677	505.898	458.292

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP101723\
Data File : PP061015.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 17 Oct 2023 15:10
Operator : YP\AJ
Sample : 04908-07MS
Misc :
ALS Vial : 15 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
TP-3MS

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/18/2023
Supervised By :Ankita Jodhani 10/18/2023

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 17 15:50:06 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP101323.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat Oct 14 13:32:01 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

