

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP101821\
 Data File : PP040161.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 18 Oct 2021 19:13
 Operator : AJ\MA
 Sample : M4249-04MSD
 Misc :
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 TP1-(D)MSD

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 19 02:18:43 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP101221.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Oct 13 08:03:59 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.891	3.979	682555	365758	22.476	21.635
2)	SA Decachlor...	10.913	9.308	548390	491774	22.187	23.172
Target Compounds							
3)	L1 AR-1016-1	6.216	5.243	538022	333234	544.881	510.625
4)	L1 AR-1016-2	6.239	5.263	812249	466394	568.406	510.148
5)	L1 AR-1016-3	6.305	5.457	506573	260999	581.286	515.890
6)	L1 AR-1016-4	6.413	5.508	411356	207739	578.120	507.358
7)	L1 AR-1016-5	6.728	5.739	383217	265173	599.153	529.645
31)	L7 AR-1260-1	7.907	6.840	597969	530390	547.395	529.521
32)	L7 AR-1260-2	8.174	7.038	691172	626721	505.699	529.277
33)	L7 AR-1260-3	8.540	7.194	455928	590700	500.333	499.956
34)	L7 AR-1260-4	8.771	7.679	571376	454090	541.329	516.645
35)	L7 AR-1260-5	9.095	7.927	1073390	1033850	518.331	505.871



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP101821\
Data File : PP040161.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 18 Oct 2021 19:13
Operator : AJ\MA
Sample : M4249-04MSD
Misc :
ALS Vial : 24 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
TP1-(D)MSD

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 19 02:18:43 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP101221.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Oct 13 08:03:59 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.891 min
Delta R.T.: -0.002 min
Response: 682555
Conc: 22.48 ng/ml

Instrument :
ECD_P
ClientSampleId :
TP1-(D)MSD




#1 Tetrachloro-m-xylene

R.T.: 3.979 min
Delta R.T.: -0.003 min
Response: 365758
Conc: 21.63 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.913 min
Delta R.T.: 0.002 min
Response: 548390
Conc: 22.19 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.308 min
Delta R.T.: -0.005 min
Response: 491774
Conc: 23.17 ng/ml



#3 AR-1016-1


R.T.: 6.216 min
Delta R.T.: -0.001 min
Response: 538022
Conc: 544.88 ng/ml

Instrument :
ECD_P
ClientSampleId :
TP1-(D)MSD




#3 AR-1016-1

R.T.: 5.243 min
Delta R.T.: -0.004 min
Response: 333234
Conc: 510.62 ng/ml




#4 AR-1016-2

R.T.: 6.239 min
Delta R.T.: -0.001 min
Response: 812249
Conc: 568.41 ng/ml



#4 AR-1016-2


R.T.: 5.263 min
Delta R.T.: -0.004 min
Response: 466394
Conc: 510.15 ng/ml



#5 AR-1016-3


R.T.: 6.305 min
Delta R.T.: -0.002 min
Response: 506573
Conc: 581.29 ng/ml

Instrument :
ECD_P
ClientSampleId :
TP1-(D)MSD




#5 AR-1016-3

R.T.: 5.457 min
Delta R.T.: -0.004 min
Response: 260999
Conc: 515.89 ng/ml




#6 AR-1016-4

R.T.: 6.413 min
Delta R.T.: -0.001 min
Response: 411356
Conc: 578.12 ng/ml



#6 AR-1016-4


R.T.: 5.508 min
Delta R.T.: -0.003 min
Response: 207739
Conc: 507.36 ng/ml



#7 AR-1016-5


R.T.: 6.728 min
Delta R.T.: -0.002 min
Response: 383217
Conc: 599.15 ng/ml

Instrument :
ECD_P
ClientSampleId :
TP1-(D)MSD




#7 AR-1016-5

R.T.: 5.739 min
Delta R.T.: -0.004 min
Response: 265173
Conc: 529.64 ng/ml




#31 AR-1260-1

R.T.: 7.907 min
Delta R.T.: -0.003 min
Response: 597969
Conc: 547.40 ng/ml



#31 AR-1260-1


R.T.: 6.840 min
Delta R.T.: -0.004 min
Response: 530390
Conc: 529.52 ng/ml



#32 AR-1260-2


R.T.: 8.174 min
Delta R.T.: -0.001 min
Response: 691172
Conc: 505.70 ng/ml

Instrument :
ECD_P
ClientSampleId :
TP1-(D)MSD




#32 AR-1260-2

R.T.: 7.038 min
Delta R.T.: -0.004 min
Response: 626721
Conc: 529.28 ng/ml




#33 AR-1260-3

R.T.: 8.540 min
Delta R.T.: 0.000 min
Response: 455928
Conc: 500.33 ng/ml



#33 AR-1260-3


R.T.: 7.194 min
Delta R.T.: -0.004 min
Response: 590700
Conc: 499.96 ng/ml



#34 AR-1260-4


R.T.: 8.771 min
Delta R.T.: -0.001 min
Response: 571376
Conc: 541.33 ng/ml

Instrument :
ECD_P
ClientSampleId :
TP1-(D)MSD




#34 AR-1260-4

R.T.: 7.679 min
Delta R.T.: -0.004 min
Response: 454090
Conc: 516.65 ng/ml



#35 AR-1260-5

R.T.: 9.095 min
Delta R.T.: -0.001 min
Response: 1073390
Conc: 518.33 ng/ml



#35 AR-1260-5

R.T.: 7.927 min
Delta R.T.: -0.004 min
Response: 1033850
Conc: 505.87 ng/ml