

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP101822\
 Data File : PP052409.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 18 Oct 2022 09:44
 Operator : YP\AJ
 Sample : AR1660CCC500
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1660CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 19 04:34:25 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP101122.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Oct 12 05:20:17 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.422	3.671	77623611	58859811	49.383	51.377
2) SA Decachlor...	10.248	8.768	66640201	53343113	48.017	47.740
Target Compounds						
3) L1 AR-1016-1	5.597	4.774	21897972	22126682	481.546	497.497
4) L1 AR-1016-2	5.619	4.794	31240896	31672346	474.555	497.741
5) L1 AR-1016-3	5.682	4.973	20028437	16984967	481.277	510.975
6) L1 AR-1016-4	5.781	5.015	16000363	13885316	440.361	509.455
7) L1 AR-1016-5	6.078	5.233	20617607	17557056	483.053	496.254
31) L7 AR-1260-1	7.210	6.281	39643028	32416642	478.285	470.607
32) L7 AR-1260-2	7.469	6.471	44004825	38024995	466.227	459.427
33) L7 AR-1260-3	7.830	6.627	29274869	35607365	487.436	461.189
34) L7 AR-1260-4	8.058	7.103	34635879	25667389	451.612	456.597
35) L7 AR-1260-5	8.382	7.346	59319772	58510061	432.744	443.671



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP101822\
Data File : PP052409.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 18 Oct 2022 09:44
Operator : YP\AJ
Sample : AR1660CCC500
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 19 04:34:25 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP101122.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Oct 12 05:20:17 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.422 min
Delta R.T.: -0.004 min
Response: 77623611
Conc: 49.38 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.671 min
Delta R.T.: -0.005 min
Response: 58859811
Conc: 51.38 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.248 min
Delta R.T.: 0.012 min
Response: 66640201
Conc: 48.02 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.768 min
Delta R.T.: 0.002 min
Response: 53343113
Conc: 47.74 ng/ml



#3 AR-1016-1


R.T.: 5.597 min
Delta R.T.: 0.000 min
Response: 21897972
Conc: 481.55 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#3 AR-1016-1

R.T.: 4.774 min
Delta R.T.: -0.004 min
Response: 22126682
Conc: 497.50 ng/ml




#4 AR-1016-2

R.T.: 5.619 min
Delta R.T.: 0.000 min
Response: 31240896
Conc: 474.55 ng/ml



#4 AR-1016-2


R.T.: 4.794 min
Delta R.T.: -0.004 min
Response: 31672346
Conc: 497.74 ng/ml



#5 AR-1016-3


R.T.: 5.682 min
Delta R.T.: 0.000 min
Response: 20028437
Conc: 481.28 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#5 AR-1016-3

R.T.: 4.973 min
Delta R.T.: -0.003 min
Response: 16984967
Conc: 510.98 ng/ml




#6 AR-1016-4

R.T.: 5.781 min
Delta R.T.: 0.000 min
Response: 16000363
Conc: 440.36 ng/ml



#6 AR-1016-4


R.T.: 5.015 min
Delta R.T.: -0.003 min
Response: 13885316
Conc: 509.46 ng/ml



#7 AR-1016-5


R.T.: 6.078 min
Delta R.T.: 0.000 min
Response: 20617607
Conc: 483.05 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#7 AR-1016-5

R.T.: 5.233 min
Delta R.T.: -0.003 min
Response: 17557056
Conc: 496.25 ng/ml




#31 AR-1260-1

R.T.: 7.210 min
Delta R.T.: 0.000 min
Response: 39643028
Conc: 478.29 ng/ml



#31 AR-1260-1


R.T.: 6.281 min
Delta R.T.: -0.002 min
Response: 32416642
Conc: 470.61 ng/ml



#32 AR-1260-2


R.T.: 7.469 min
Delta R.T.: 0.003 min
Response: 44004825
Conc: 466.23 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#32 AR-1260-2

R.T.: 6.471 min
Delta R.T.: -0.002 min
Response: 38024995
Conc: 459.43 ng/ml




#33 AR-1260-3

R.T.: 7.830 min
Delta R.T.: 0.003 min
Response: 29274869
Conc: 487.44 ng/ml



#33 AR-1260-3


R.T.: 6.627 min
Delta R.T.: -0.002 min
Response: 35607365
Conc: 461.19 ng/ml



#34 AR-1260-4


R.T.: 8.058 min
Delta R.T.: 0.004 min
Response: 34635879
Conc: 451.61 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#34 AR-1260-4

R.T.: 7.103 min
Delta R.T.: -0.001 min
Response: 25667389
Conc: 456.60 ng/ml



#35 AR-1260-5

R.T.: 8.382 min
Delta R.T.: 0.005 min
Response: 59319772
Conc: 432.74 ng/ml



#35 AR-1260-5

R.T.: 7.346 min
Delta R.T.: -0.001 min
Response: 58510061
Conc: 443.67 ng/ml