

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP101823\
 Data File : PP061036.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 18 Oct 2023 16:52
 Operator : YP\AJ
 Sample : AR1660CCC500
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1660CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 18 17:05:11 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP101323.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sat Oct 14 13:32:01 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.552	3.688	111.0E6	65173937	50.487	48.206
2) SA Decachlor...	10.508	8.753	84023170	64586398	53.655	42.956
Target Compounds						
3) L1 AR-1016-1	5.743	4.786	38178659	24194472	531.812	464.882
4) L1 AR-1016-2	5.765	4.805	54059148	33584707	527.047	468.935
5) L1 AR-1016-3	5.829	4.982	33330927	18659139	533.379	461.778
6) L1 AR-1016-4	5.929	5.025	27353843	14820873	535.473	458.196
7) L1 AR-1016-5	6.228	5.240	26211884	19645555	518.443	470.515
31) L7 AR-1260-1	7.369	6.282	44342487	36517008	473.716	441.289
32) L7 AR-1260-2	7.628	6.471	51519975	43225026	474.543	438.955
33) L7 AR-1260-3	7.991	6.626	33941713	41227869	488.357	435.962
34) L7 AR-1260-4	8.225	7.101	42289420	31132160	495.731	434.367
35) L7 AR-1260-5	8.563	7.344	87496704	72044737	542.301	429.554


(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP101823\
Data File : PP061036.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 18 Oct 2023 16:52
Operator : YP\AJ
Sample : AR1660CCC500
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 18 17:05:11 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP101323.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat Oct 14 13:32:01 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.552 min
Delta R.T.: 0.001 min
Response: 110987265
Conc: 50.49 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.688 min
Delta R.T.: 0.001 min
Response: 65173937
Conc: 48.21 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.508 min
Delta R.T.: 0.000 min
Response: 84023170
Conc: 53.65 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.753 min
Delta R.T.: 0.000 min
Response: 64586398
Conc: 42.96 ng/ml



#3 AR-1016-1


R.T.: 5.743 min
Delta R.T.: 0.001 min
Response: 38178659
Conc: 531.81 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#3 AR-1016-1

R.T.: 4.786 min
Delta R.T.: 0.000 min
Response: 24194472
Conc: 464.88 ng/ml




#4 AR-1016-2

R.T.: 5.765 min
Delta R.T.: 0.000 min
Response: 54059148
Conc: 527.05 ng/ml



#4 AR-1016-2


R.T.: 4.805 min
Delta R.T.: 0.000 min
Response: 33584707
Conc: 468.94 ng/ml



#5 AR-1016-3


R.T.: 5.829 min
Delta R.T.: 0.001 min
Response: 33330927
Conc: 533.38 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#5 AR-1016-3

R.T.: 4.982 min
Delta R.T.: 0.000 min
Response: 18659139
Conc: 461.78 ng/ml




#6 AR-1016-4

R.T.: 5.929 min
Delta R.T.: 0.001 min
Response: 27353843
Conc: 535.47 ng/ml



#6 AR-1016-4


R.T.: 5.025 min
Delta R.T.: 0.000 min
Response: 14820873
Conc: 458.20 ng/ml



#7 AR-1016-5


R.T.: 6.228 min
Delta R.T.: 0.000 min
Response: 26211884
Conc: 518.44 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#7 AR-1016-5

R.T.: 5.240 min
Delta R.T.: 0.000 min
Response: 19645555
Conc: 470.51 ng/ml




#31 AR-1260-1

R.T.: 7.369 min
Delta R.T.: 0.001 min
Response: 44342487
Conc: 473.72 ng/ml



#31 AR-1260-1


R.T.: 6.282 min
Delta R.T.: 0.000 min
Response: 36517008
Conc: 441.29 ng/ml



#32 AR-1260-2


R.T.: 7.628 min
Delta R.T.: 0.000 min
Response: 51519975
Conc: 474.54 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#32 AR-1260-2

R.T.: 6.471 min
Delta R.T.: 0.000 min
Response: 43225026
Conc: 438.95 ng/ml




#33 AR-1260-3

R.T.: 7.991 min
Delta R.T.: 0.000 min
Response: 33941713
Conc: 488.36 ng/ml



#33 AR-1260-3


R.T.: 6.626 min
Delta R.T.: 0.000 min
Response: 41227869
Conc: 435.96 ng/ml



#34 AR-1260-4


R.T.: 8.225 min
Delta R.T.: 0.000 min
Response: 42289420
Conc: 495.73 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#34 AR-1260-4

R.T.: 7.101 min
Delta R.T.: -0.001 min
Response: 31132160
Conc: 434.37 ng/ml



#35 AR-1260-5

R.T.: 8.563 min
Delta R.T.: 0.003 min
Response: 87496704
Conc: 542.30 ng/ml



#35 AR-1260-5

R.T.: 7.344 min
Delta R.T.: 0.000 min
Response: 72044737
Conc: 429.55 ng/ml