

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP101724\
 Data File : PP068014.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 18 Oct 2024 04:02
 Operator : YP\AJ
 Sample : P4426-29
 Misc :
 ALS Vial : 30 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 PAD-15

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 18 05:21:19 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP100824.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Oct 09 05:48:32 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.753	4.049	17256599	17182876	18.643	17.006
2)	SA Decachlor...	10.662	9.209	25161222	22769572	21.855	20.304
Target Compounds							
3)	L1 AR-1016-1	5.914	5.156	374528	483908	11.326	13.831
4)	L1 AR-1016-2	5.944	5.177	229534	483318	4.706	10.034 #
5)	L1 AR-1016-3	6.003	5.357	366005	452191	11.615	16.787 #
6)	L1 AR-1016-4	6.099	5.396	280475	893807	10.895	37.110 #
7)	L1 AR-1016-5	6.397	5.614	446138	455000	16.483	14.962
8)	L2 AR-1221-1	4.953	4.255	357412	105920	27.901	8.042 #
9)	L2 AR-1221-2	5.057	4.357	656655	345167	69.410	34.651 #
10)	L2 AR-1221-3	5.110	4.428	327477	175650	11.328	5.642 #
11)	L3 AR-1232-1	5.110	4.428	327477	175650	14.325	7.219 #
12)	L3 AR-1232-2	5.648	5.177	194841	483318	16.536	21.610 #
13)	L3 AR-1232-3	5.937	5.357	262111	452191	11.767	37.860 #
14)	L3 AR-1232-4	6.099	5.440	280475	422704	24.848	36.657 #
15)	L3 AR-1232-5	6.190	5.614	772987	455000	84.756	33.828 #
16)	L4 AR-1242-1	5.914	5.156	374528	483908	13.406	16.643
17)	L4 AR-1242-2	5.937	5.177	262111	483318	6.450	12.038 #
18)	L4 AR-1242-3	6.003	5.357	366005	452191	13.573	20.102 #
19)	L4 AR-1242-4	6.099	5.440	280475	422704	13.075	17.807 #
20)	L4 AR-1242-5	6.837	5.972	563143	7807370	23.342	272.513 #
21)	L5 AR-1248-1	5.914	5.156	374528	483908	17.483	21.352
22)	L5 AR-1248-2	6.190	5.396	772987	893807	22.994	27.007
23)	L5 AR-1248-3	6.397	5.440	446138	422704	12.181	12.261
24)	L5 AR-1248-4	6.772	5.614	7887267	455000	194.985	11.109 #
25)	L5 AR-1248-5	6.837	6.014	563143	406755	14.199	10.863
26)	L6 AR-1254-1	6.772	5.972	7887267	7807370	176.124	132.140
27)	L6 AR-1254-2	6.987	6.118	11002956	10531525	166.465	200.737
28)	L6 AR-1254-3	7.353	6.543	5635352	25394236	81.139	305.190 #
29)	L6 AR-1254-4	7.636	6.756	5250979	2103468	105.220	43.573 #
30)	L6 AR-1254-5	8.052	7.177	44475541	42378289	750.101	576.174
31)	L7 AR-1260-1	7.518	6.659	30822017	30904527	570.528	542.547
32)	L7 AR-1260-2	7.771	6.845	38458934	38821575	607.847	581.994
33)	L7 AR-1260-3	8.132	7.002	30291755	35737834	584.232	561.024
34)	L7 AR-1260-4	8.370	7.477	35119909	31318833	583.441	566.419
35)	L7 AR-1260-5	8.707	7.715	67679714	73317157	628.339	599.493
36)	L8 AR-1262-1	8.370	7.214	35119909	32788509	489.331	416.767
37)	L8 AR-1262-2	8.707	7.477	67679714	31318833	556.667	441.346
38)	L8 AR-1262-3	9.054	8.001	45272641	16558805	511.473	290.000 #
39)	L8 AR-1262-4	9.136	8.065	27584358	50899689	394.373	506.785 #
40)	L8 AR-1262-5	9.839	8.593	19900633	19464846	425.307	389.159
41)	L9 AR-1268-1	9.054	8.001	45272641	16558805	296.581	104.051 #

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP101724\
Data File : PP068014.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 18 Oct 2024 04:02
Operator : YP\AJ
Sample : P4426-29
Misc :
ALS Vial : 30 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
PAD-15

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 18 05:21:19 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP100824.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Oct 09 05:48:32 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
42)	L9 AR-1268-2	9.136	8.065	27584358	50899689	201.942	355.479 #
43)	L9 AR-1268-3	9.392	8.285	728933	696425	6.028	5.462
44)	L9 AR-1268-4	9.839	8.593	19900633	19464846	391.591	344.657
45)	L9 AR-1268-5	10.290	8.918	4232190	4167708	11.332	11.025



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP101724\
Data File : PP068014.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 18 Oct 2024 04:02
Operator : YP\AJ
Sample : P4426-29
Misc :
ALS Vial : 30 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
PAD-15

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 18 05:21:19 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP100824.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Oct 09 05:48:32 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.753 min
Delta R.T.: -0.002 min
Response: 17256599
Conc: 18.64 ng/ml

Instrument :
ECD_P
ClientSampleId :
PAD-15




#1 Tetrachloro-m-xylene

R.T.: 4.049 min
Delta R.T.: -0.003 min
Response: 17182876
Conc: 17.01 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.662 min
Delta R.T.: -0.007 min
Response: 25161222
Conc: 21.85 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.209 min
Delta R.T.: -0.010 min
Response: 22769572
Conc: 20.30 ng/ml



#3 AR-1016-1


R.T.: 5.914 min
Delta R.T.: -0.005 min
Response: 374528
Conc: 11.33 ng/ml

Instrument :
ECD_P
ClientSampleId :
PAD-15




#3 AR-1016-1

R.T.: 5.156 min
Delta R.T.: -0.004 min
Response: 483908
Conc: 13.83 ng/ml




#4 AR-1016-2

R.T.: 5.944 min
Delta R.T.: 0.002 min
Response: 229534
Conc: 4.71 ng/ml



#4 AR-1016-2


R.T.: 5.177 min
Delta R.T.: -0.004 min
Response: 483318
Conc: 10.03 ng/ml



#5 AR-1016-3


R.T.: 6.003 min
Delta R.T.: -0.002 min
Response: 366005
Conc: 11.61 ng/ml

Instrument :
ECD_P
ClientSampleId :
PAD-15




#5 AR-1016-3

R.T.: 5.357 min
Delta R.T.: -0.004 min
Response: 452191
Conc: 16.79 ng/ml




#6 AR-1016-4

R.T.: 6.099 min
Delta R.T.: -0.004 min
Response: 280475
Conc: 10.89 ng/ml



#6 AR-1016-4


R.T.: 5.396 min
Delta R.T.: -0.005 min
Response: 893807
Conc: 37.11 ng/ml



#7 AR-1016-5


R.T.: 6.397 min
Delta R.T.: -0.001 min
Response: 446138
Conc: 16.48 ng/ml

Instrument :
ECD_P
ClientSampleId :
PAD-15




#7 AR-1016-5

R.T.: 5.614 min
Delta R.T.: -0.006 min
Response: 455000
Conc: 14.96 ng/ml




#8 AR-1221-1

R.T.: 4.953 min
Delta R.T.: -0.002 min
Response: 357412
Conc: 27.90 ng/ml



#8 AR-1221-1


R.T.: 4.255 min
Delta R.T.: -0.012 min
Response: 105920
Conc: 8.04 ng/ml



#9 AR-1221-2


R.T.: 5.057 min
Delta R.T.: 0.016 min
Response: 656655
Conc: 69.41 ng/ml

Instrument :
ECD_P
ClientSampleId :
PAD-15




#9 AR-1221-2

R.T.: 4.357 min
Delta R.T.: 0.001 min
Response: 345167
Conc: 34.65 ng/ml




#10 AR-1221-3

R.T.: 5.110 min
Delta R.T.: -0.009 min
Response: 327477
Conc: 11.33 ng/ml



#10 AR-1221-3


R.T.: 4.428 min
Delta R.T.: -0.006 min
Response: 175650
Conc: 5.64 ng/ml



#11 AR-1232-1


R.T.: 5.110 min
Delta R.T.: -0.008 min
Response: 327477
Conc: 14.33 ng/ml

Instrument :
ECD_P
ClientSampleId :
PAD-15




#11 AR-1232-1

R.T.: 4.428 min
Delta R.T.: -0.006 min
Response: 175650
Conc: 7.22 ng/ml




#12 AR-1232-2

R.T.: 5.648 min
Delta R.T.: -0.003 min
Response: 194841
Conc: 16.54 ng/ml



#12 AR-1232-2


R.T.: 5.177 min
Delta R.T.: -0.004 min
Response: 483318
Conc: 21.61 ng/ml



#13 AR-1232-3


R.T.: 5.937 min
Delta R.T.: -0.003 min
Response: 262111
Conc: 11.77 ng/ml

Instrument :
ECD_P
ClientSampleId :
PAD-15




#13 AR-1232-3

R.T.: 5.357 min
Delta R.T.: -0.004 min
Response: 452191
Conc: 37.86 ng/ml




#14 AR-1232-4

R.T.: 6.099 min
Delta R.T.: -0.002 min
Response: 280475
Conc: 24.85 ng/ml



#14 AR-1232-4


R.T.: 5.440 min
Delta R.T.: -0.005 min
Response: 422704
Conc: 36.66 ng/ml



#15 AR-1232-5


R.T.: 6.190 min
Delta R.T.: -0.001 min
Response: 772987
Conc: 84.76 ng/ml

Instrument :
ECD_P
ClientSampleId :
PAD-15




#15 AR-1232-5

R.T.: 5.614 min
Delta R.T.: -0.005 min
Response: 455000
Conc: 33.83 ng/ml




#16 AR-1242-1

R.T.: 5.914 min
Delta R.T.: -0.002 min
Response: 374528
Conc: 13.41 ng/ml



#16 AR-1242-1


R.T.: 5.156 min
Delta R.T.: -0.004 min
Response: 483908
Conc: 16.64 ng/ml



#17 AR-1242-2


R.T.: 5.937 min
Delta R.T.: -0.003 min
Response: 262111
Conc: 6.45 ng/ml

Instrument :
ECD_P
ClientSampleId :
PAD-15




#17 AR-1242-2

R.T.: 5.177 min
Delta R.T.: -0.004 min
Response: 483318
Conc: 12.04 ng/ml




#18 AR-1242-3

R.T.: 6.003 min
Delta R.T.: 0.001 min
Response: 366005
Conc: 13.57 ng/ml



#18 AR-1242-3


R.T.: 5.357 min
Delta R.T.: -0.004 min
Response: 452191
Conc: 20.10 ng/ml



#19 AR-1242-4


R.T.: 6.099 min
Delta R.T.: -0.002 min
Response: 280475
Conc: 13.08 ng/ml

Instrument :
ECD_P
ClientSampleId :
PAD-15




#19 AR-1242-4

R.T.: 5.440 min
Delta R.T.: -0.005 min
Response: 422704
Conc: 17.81 ng/ml




#20 AR-1242-5

R.T.: 6.837 min
Delta R.T.: 0.002 min
Response: 563143
Conc: 23.34 ng/ml



#20 AR-1242-5


R.T.: 5.972 min
Delta R.T.: -0.003 min
Response: 7807370
Conc: 272.51 ng/ml



#21 AR-1248-1


R.T.: 5.914 min
Delta R.T.: -0.003 min
Response: 374528
Conc: 17.48 ng/ml

Instrument :
ECD_P
ClientSampleId :
PAD-15




#21 AR-1248-1

R.T.: 5.156 min
Delta R.T.: -0.003 min
Response: 483908
Conc: 21.35 ng/ml




#22 AR-1248-2

R.T.: 6.190 min
Delta R.T.: 0.000 min
Response: 772987
Conc: 22.99 ng/ml



#22 AR-1248-2


R.T.: 5.396 min
Delta R.T.: -0.004 min
Response: 893807
Conc: 27.01 ng/ml



#23 AR-1248-3


R.T.: 6.397 min
Delta R.T.: 0.000 min
Response: 446138
Conc: 12.18 ng/ml

Instrument :
ECD_P
ClientSampleId :
PAD-15




#23 AR-1248-3

R.T.: 5.440 min
Delta R.T.: -0.004 min
Response: 422704
Conc: 12.26 ng/ml




#24 AR-1248-4

R.T.: 6.772 min
Delta R.T.: -0.024 min
Response: 7887267
Conc: 194.98 ng/ml



#24 AR-1248-4


R.T.: 5.614 min
Delta R.T.: -0.004 min
Response: 455000
Conc: 11.11 ng/ml



#25 AR-1248-5


R.T.: 6.837 min
Delta R.T.: 0.001 min
Response: 563143
Conc: 14.20 ng/ml

Instrument :
ECD_P
ClientSampleId :
PAD-15




#25 AR-1248-5

R.T.: 6.014 min
Delta R.T.: -0.003 min
Response: 406755
Conc: 10.86 ng/ml




#26 AR-1254-1

R.T.: 6.772 min
Delta R.T.: -0.001 min
Response: 7887267
Conc: 176.12 ng/ml



#26 AR-1254-1


R.T.: 5.972 min
Delta R.T.: -0.003 min
Response: 7807370
Conc: 132.14 ng/ml



#27 AR-1254-2


R.T.: 6.987 min
Delta R.T.: -0.003 min
Response: 11002956
Conc: 166.46 ng/ml

Instrument :
ECD_P
ClientSampleId :
PAD-15




#27 AR-1254-2

R.T.: 6.118 min
Delta R.T.: -0.004 min
Response: 10531525
Conc: 200.74 ng/ml




#28 AR-1254-3

R.T.: 7.353 min
Delta R.T.: -0.001 min
Response: 5635352
Conc: 81.14 ng/ml



#28 AR-1254-3


R.T.: 6.543 min
Delta R.T.: 0.010 min
Response: 25394236
Conc: 305.19 ng/ml



#29 AR-1254-4


R.T.: 7.636 min
Delta R.T.: -0.002 min
Response: 5250979
Conc: 105.22 ng/ml

Instrument :
ECD_P
ClientSampleId :
PAD-15




#29 AR-1254-4

R.T.: 6.756 min
Delta R.T.: -0.004 min
Response: 2103468
Conc: 43.57 ng/ml




#30 AR-1254-5

R.T.: 8.052 min
Delta R.T.: -0.002 min
Response: 44475541
Conc: 750.10 ng/ml



#30 AR-1254-5


R.T.: 7.177 min
Delta R.T.: -0.005 min
Response: 42378289
Conc: 576.17 ng/ml



#31 AR-1260-1


R.T.: 7.518 min
Delta R.T.: -0.004 min
Response: 30822017
Conc: 570.53 ng/ml

Instrument :
ECD_P
ClientSampleId :
PAD-15




#31 AR-1260-1

R.T.: 6.659 min
Delta R.T.: -0.006 min
Response: 30904527
Conc: 542.55 ng/ml




#32 AR-1260-2

R.T.: 7.771 min
Delta R.T.: -0.004 min
Response: 38458934
Conc: 607.85 ng/ml



#32 AR-1260-2


R.T.: 6.845 min
Delta R.T.: -0.005 min
Response: 38821575
Conc: 581.99 ng/ml



#33 AR-1260-3


R.T.: 8.132 min
Delta R.T.: -0.005 min
Response: 30291755
Conc: 584.23 ng/ml

Instrument :
ECD_P
ClientSampleId :
PAD-15




#33 AR-1260-3

R.T.: 7.002 min
Delta R.T.: -0.006 min
Response: 35737834
Conc: 561.02 ng/ml




#34 AR-1260-4

R.T.: 8.370 min
Delta R.T.: -0.005 min
Response: 35119909
Conc: 583.44 ng/ml



#34 AR-1260-4


R.T.: 7.477 min
Delta R.T.: -0.006 min
Response: 31318833
Conc: 566.42 ng/ml



#35 AR-1260-5


R.T.: 8.707 min
Delta R.T.: -0.005 min
Response: 67679714
Conc: 628.34 ng/ml

Instrument :
ECD_P
ClientSampleId :
PAD-15




#35 AR-1260-5

R.T.: 7.715 min
Delta R.T.: -0.006 min
Response: 73317157
Conc: 599.49 ng/ml




#36 AR-1262-1

R.T.: 8.370 min
Delta R.T.: -0.003 min
Response: 35119909
Conc: 489.33 ng/ml



#36 AR-1262-1


R.T.: 7.214 min
Delta R.T.: -0.005 min
Response: 32788509
Conc: 416.77 ng/ml



#37 AR-1262-2


R.T.: 8.707 min
Delta R.T.: -0.002 min
Response: 67679714
Conc: 556.67 ng/ml

Instrument :
ECD_P
ClientSampleId :
PAD-15




#37 AR-1262-2

R.T.: 7.477 min
Delta R.T.: -0.005 min
Response: 31318833
Conc: 441.35 ng/ml




#38 AR-1262-3

R.T.: 9.054 min
Delta R.T.: 0.006 min
Response: 45272641
Conc: 511.47 ng/ml



#38 AR-1262-3


R.T.: 8.001 min
Delta R.T.: -0.006 min
Response: 16558805
Conc: 290.00 ng/ml



#39 AR-1262-4


R.T.: 9.136 min
Delta R.T.: -0.004 min
Response: 27584358
Conc: 394.37 ng/ml

Instrument :
ECD_P
ClientSampleId :
PAD-15




#39 AR-1262-4

R.T.: 8.065 min
Delta R.T.: -0.005 min
Response: 50899689
Conc: 506.79 ng/ml




#40 AR-1262-5

R.T.: 9.839 min
Delta R.T.: -0.003 min
Response: 19900633
Conc: 425.31 ng/ml



#40 AR-1262-5


R.T.: 8.593 min
Delta R.T.: -0.006 min
Response: 19464846
Conc: 389.16 ng/ml



#41 AR-1268-1


R.T.: 9.054 min
Delta R.T.: 0.011 min
Response: 45272641
Conc: 296.58 ng/ml

Instrument :
ECD_P
ClientSampleId :
PAD-15




#41 AR-1268-1

R.T.: 8.001 min
Delta R.T.: -0.006 min
Response: 16558805
Conc: 104.05 ng/ml




#42 AR-1268-2

R.T.: 9.136 min
Delta R.T.: -0.009 min
Response: 27584358
Conc: 201.94 ng/ml



#42 AR-1268-2


R.T.: 8.065 min
Delta R.T.: -0.008 min
Response: 50899689
Conc: 355.48 ng/ml



#43 AR-1268-3


R.T.: 9.392 min
Delta R.T.: -0.003 min
Response: 728933
Conc: 6.03 ng/ml

Instrument :
ECD_P
ClientSampleId :
PAD-15




#43 AR-1268-3

R.T.: 8.285 min
Delta R.T.: -0.006 min
Response: 696425
Conc: 5.46 ng/ml




#44 AR-1268-4

R.T.: 9.839 min
Delta R.T.: -0.004 min
Response: 19900633
Conc: 391.59 ng/ml



#44 AR-1268-4


R.T.: 8.593 min
Delta R.T.: -0.007 min
Response: 19464846
Conc: 344.66 ng/ml



#45 AR-1268-5

R.T.: 10.290 min
Delta R.T.: -0.004 min
Response: 4232190
Conc: 11.33 ng/ml

Instrument :
ECD_P
ClientSampleId :
PAD-15



#45 AR-1268-5

R.T.: 8.918 min
Delta R.T.: -0.007 min
Response: 4167708
Conc: 11.02 ng/ml