

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP101923\
 Data File : PP061108.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 19 Oct 2023 16:01
 Operator : YP\AJ
 Sample : AR1254CCC500
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1254CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 19 18:02:07 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP101323.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sat Oct 14 13:32:01 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.554	3.690	122.9E6	72925455	55.922	53.939
2) SA Decachlor...	10.520	8.759	89331047	73097334	57.044	48.617
Target Compounds						
26) L6 AR-1254-1	6.615	5.599	40908156	38194763	517.242	494.359
27) L6 AR-1254-2	6.836	5.748	61780387	34441898	518.605	497.095
28) L6 AR-1254-3	7.208	6.154	63524349	56035097	519.244	495.222
29) L6 AR-1254-4	7.496	6.385	45603869	34491582	518.111	501.994
30) L6 AR-1254-5	7.920	6.806	51269150	54230189	516.305	506.970



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP101923\
Data File : PP061108.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 19 Oct 2023 16:01
Operator : YP\AJ
Sample : AR1254CCC500
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 19 18:02:07 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP101323.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat Oct 14 13:32:01 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.554 min
Delta R.T.: 0.003 min
Response: 122934873
Conc: 55.92 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.690 min
Delta R.T.: 0.003 min
Response: 72925455
Conc: 53.94 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.520 min
Delta R.T.: 0.013 min
Response: 89331047
Conc: 57.04 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.759 min
Delta R.T.: 0.005 min
Response: 73097334
Conc: 48.62 ng/ml



#26 AR-1254-1


R.T.: 6.615 min
Delta R.T.: 0.004 min
Response: 40908156
Conc: 517.24 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500




#26 AR-1254-1

R.T.: 5.599 min
Delta R.T.: 0.002 min
Response: 38194763
Conc: 494.36 ng/ml




#27 AR-1254-2

R.T.: 6.836 min
Delta R.T.: 0.004 min
Response: 61780387
Conc: 518.61 ng/ml



#27 AR-1254-2


R.T.: 5.748 min
Delta R.T.: 0.002 min
Response: 34441898
Conc: 497.09 ng/ml



#28 AR-1254-3


R.T.: 7.208 min
Delta R.T.: 0.005 min
Response: 63524349
Conc: 519.24 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500




#28 AR-1254-3

R.T.: 6.154 min
Delta R.T.: 0.001 min
Response: 56035097
Conc: 495.22 ng/ml




#29 AR-1254-4

R.T.: 7.496 min
Delta R.T.: 0.005 min
Response: 45603869
Conc: 518.11 ng/ml



#29 AR-1254-4


R.T.: 6.385 min
Delta R.T.: 0.002 min
Response: 34491582
Conc: 501.99 ng/ml



#30 AR-1254-5

R.T.: 7.920 min
Delta R.T.: 0.005 min
Response: 51269150
Conc: 516.30 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500



#30 AR-1254-5

R.T.: 6.806 min
Delta R.T.: 0.003 min
Response: 54230189
Conc: 506.97 ng/ml