

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102021\
 Data File : PP040294.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 21 Oct 2021 2:13
 Operator : AJ\MA
 Sample : AR1660CCC500
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 21 06:00:12 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP101221.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Oct 13 08:03:59 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.889	3.977	1605151	840580	52.856	49.721
2)	SA Decachlor...	10.904	9.302	1274389	1054189	51.560	49.672
Target Compounds							
3)	L1 AR-1016-1	6.212	5.240	524436	322374	531.122	493.984
4)	L1 AR-1016-2	6.236	5.260	787317	452514	550.959	494.965
5)	L1 AR-1016-3	6.302	5.453	489699	247121	561.924	488.459
6)	L1 AR-1016-4	6.410	5.505	401765	202267	564.641	493.993
7)	L1 AR-1016-5	6.725	5.736	373308	226098	583.661	451.598
8)	L2 AR-1221-1	5.134	4.236	53994	29438	155.040	143.846
9)	L2 AR-1221-2	5.239	4.336	83910	49057	376.626	291.646
10)	L2 AR-1221-3	5.326	4.425	300962	188797	373.317	348.630
11)	L3 AR-1232-1	5.326	4.425	300962	188797	483.670	464.835
12)	L3 AR-1232-2	5.916	5.260	423702	452514	1234.495	1157.516
13)	L3 AR-1232-3	6.236	5.453	787317	247121	1245.578	1193.038
14)	L3 AR-1232-4	6.410	5.550	401765	248995	1281.007	1315.969
15)	L3 AR-1232-5	6.507	5.736	338967	226098	1445.545	1095.114
16)	L4 AR-1242-1	6.212	5.240	524436	322374	667.915	647.884
17)	L4 AR-1242-2	6.236	5.260	787317	452514	694.104	662.540
18)	L4 AR-1242-3	6.302	5.453	489699	247121	704.593	654.982
19)	L4 AR-1242-4	6.410	5.550	401765	248995	697.276	672.185
20)	L4 AR-1242-5	7.194	6.116	65198	207840	124.662	435.756 #
21)	L5 AR-1248-1	6.212	5.240	524436	322374	850.116	826.861
22)	L5 AR-1248-2	6.507	5.505	338967	202267	423.436	376.311
23)	L5 AR-1248-3	6.725	5.550	373308	248995	416.336	440.474
24)	L5 AR-1248-4	7.154	5.736	76480	226098	83.109	353.797 #
25)	L5 AR-1248-5	7.194	6.158	65198	40427	73.704	64.804
26)	L6 AR-1254-1	7.123	6.116	290262	207840	312.851	216.946 #
27)	L6 AR-1254-2	7.352	6.276	247282	192597	178.624	218.215
28)	L6 AR-1254-3	7.736	6.716	161630	354919	113.452	260.325 #
29)	L6 AR-1254-4	8.031	6.938	143112	45505	135.243	53.054 #
30)	L6 AR-1254-5	8.461	7.369	834908	694648	714.305	538.149
31)	L7 AR-1260-1	7.902	6.836	614490	500536	562.519	499.717
32)	L7 AR-1260-2	8.169	7.033	697743	607325	510.506	512.897
33)	L7 AR-1260-3	8.535	7.189	481465	564053	528.357	477.402
34)	L7 AR-1260-4	8.766	7.674	578357	456873	547.943	519.812
35)	L7 AR-1260-5	9.090	7.922	1080715	1036474	521.868	507.155
36)	L8 AR-1262-1	8.535	7.369	481465	694648	353.068	988.861 #
37)	L8 AR-1262-2	9.090	7.922	1080715	1036474	435.794	460.569
38)	L8 AR-1262-3	9.421f	8.210	744679	228449	647.493	234.575 #
39)	L8 AR-1262-4	9.473f	8.273	456871	794170	559.147	447.972
40)	L8 AR-1262-5	10.157	8.773	279395	262480	294.066	289.851
41)	L9 AR-1268-1	9.421f	8.210	744679	228449	252.172	84.664 #

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102021\
 Data File : PP040294.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 21 Oct 2021 2:13
 Operator : AJ\MA
 Sample : AR1660CCC500
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 21 06:00:12 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP101221.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Oct 13 08:03:59 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
42) L9	AR-1268-2	9.473f	8.273	456871	794170	161.819	317.196 #
43) L9	AR-1268-3	0.000	8.483	0	14909	N.D.	6.809 #
44) L9	AR-1268-4	10.157	8.773	279395	262480	271.094	264.695
45) L9	AR-1268-5	10.569	9.056	114067	88547	13.262	12.420



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102021\
Data File : PP040294.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 21 Oct 2021 2:13
Operator : AJ\MA
Sample : AR1660CCC500
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 21 06:00:12 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP101221.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Oct 13 08:03:59 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.889 min
Delta R.T.: -0.004 min
Response: 1605151
Conc: 52.86 ng/ml

Instrument :
ECD_P
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 3.977 min
Delta R.T.: -0.005 min
Response: 840580
Conc: 49.72 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.904 min
Delta R.T.: -0.007 min
Response: 1274389
Conc: 51.56 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.302 min
Delta R.T.: -0.011 min
Response: 1054189
Conc: 49.67 ng/ml



#3 AR-1016-1


R.T.: 6.212 min
Delta R.T.: -0.005 min
Response: 524436
Conc: 531.12 ng/ml

Instrument :
ECD_P
ClientSampleId :




#3 AR-1016-1

R.T.: 5.240 min
Delta R.T.: -0.007 min
Response: 322374
Conc: 493.98 ng/ml




#4 AR-1016-2

R.T.: 6.236 min
Delta R.T.: -0.004 min
Response: 787317
Conc: 550.96 ng/ml



#4 AR-1016-2


R.T.: 5.260 min
Delta R.T.: -0.007 min
Response: 452514
Conc: 494.97 ng/ml



#5 AR-1016-3


R.T.: 6.302 min
Delta R.T.: -0.005 min
Response: 489699
Conc: 561.92 ng/ml

Instrument :
ECD_P
ClientSampleId :




#5 AR-1016-3

R.T.: 5.453 min
Delta R.T.: -0.007 min
Response: 247121
Conc: 488.46 ng/ml




#6 AR-1016-4

R.T.: 6.410 min
Delta R.T.: -0.005 min
Response: 401765
Conc: 564.64 ng/ml



#6 AR-1016-4


R.T.: 5.505 min
Delta R.T.: -0.007 min
Response: 202267
Conc: 493.99 ng/ml



#7 AR-1016-5


R.T.: 6.725 min
Delta R.T.: -0.005 min
Response: 373308
Conc: 583.66 ng/ml

Instrument :
ECD_P
ClientSampleId :




#7 AR-1016-5

R.T.: 5.736 min
Delta R.T.: -0.007 min
Response: 226098
Conc: 451.60 ng/ml




#8 AR-1221-1

R.T.: 5.134 min
Delta R.T.: -0.011 min
Response: 53994
Conc: 155.04 ng/ml



#8 AR-1221-1


R.T.: 4.236 min
Delta R.T.: -0.007 min
Response: 29438
Conc: 143.85 ng/ml



#9 AR-1221-2


R.T.: 5.239 min
Delta R.T.: -0.004 min
Response: 83910
Conc: 376.63 ng/ml

Instrument :
ECD_P
ClientSampleId :




#9 AR-1221-2

R.T.: 4.336 min
Delta R.T.: -0.007 min
Response: 49057
Conc: 291.65 ng/ml




#10 AR-1221-3

R.T.: 5.326 min
Delta R.T.: -0.005 min
Response: 300962
Conc: 373.32 ng/ml



#10 AR-1221-3


R.T.: 4.425 min
Delta R.T.: -0.007 min
Response: 188797
Conc: 348.63 ng/ml



#11 AR-1232-1


R.T.: 5.326 min
Delta R.T.: -0.004 min
Response: 300962
Conc: 483.67 ng/ml

Instrument :
ECD_P
ClientSampleId :




#11 AR-1232-1

R.T.: 4.425 min
Delta R.T.: -0.006 min
Response: 188797
Conc: 464.84 ng/ml




#12 AR-1232-2

R.T.: 5.916 min
Delta R.T.: -0.004 min
Response: 423702
Conc: 1234.49 ng/ml



#12 AR-1232-2


R.T.: 5.260 min
Delta R.T.: -0.007 min
Response: 452514
Conc: 1157.52 ng/ml



#13 AR-1232-3


R.T.: 6.236 min
Delta R.T.: -0.004 min
Response: 787317
Conc: 1245.58 ng/ml

Instrument :
ECD_P
ClientSampleId :




#13 AR-1232-3

R.T.: 5.453 min
Delta R.T.: -0.007 min
Response: 247121
Conc: 1193.04 ng/ml




#14 AR-1232-4

R.T.: 6.410 min
Delta R.T.: -0.005 min
Response: 401765
Conc: 1281.01 ng/ml



#14 AR-1232-4


R.T.: 5.550 min
Delta R.T.: -0.007 min
Response: 248995
Conc: 1315.97 ng/ml



#15 AR-1232-5


R.T.: 6.507 min
Delta R.T.: -0.004 min
Response: 338967
Conc: 1445.54 ng/ml

Instrument :
ECD_P
ClientSampleId :




#15 AR-1232-5

R.T.: 5.736 min
Delta R.T.: -0.007 min
Response: 226098
Conc: 1095.11 ng/ml




#16 AR-1242-1

R.T.: 6.212 min
Delta R.T.: -0.004 min
Response: 524436
Conc: 667.92 ng/ml



#16 AR-1242-1


R.T.: 5.240 min
Delta R.T.: -0.007 min
Response: 322374
Conc: 647.88 ng/ml



#17 AR-1242-2


R.T.: 6.236 min
Delta R.T.: -0.005 min
Response: 787317
Conc: 694.10 ng/ml

Instrument :
ECD_P
ClientSampleId :




#17 AR-1242-2

R.T.: 5.260 min
Delta R.T.: -0.007 min
Response: 452514
Conc: 662.54 ng/ml




#18 AR-1242-3

R.T.: 6.302 min
Delta R.T.: -0.004 min
Response: 489699
Conc: 704.59 ng/ml



#18 AR-1242-3


R.T.: 5.453 min
Delta R.T.: -0.007 min
Response: 247121
Conc: 654.98 ng/ml



#19 AR-1242-4


R.T.: 6.410 min
Delta R.T.: -0.004 min
Response: 401765
Conc: 697.28 ng/ml

Instrument :
ECD_P
ClientSampleId :




#19 AR-1242-4

R.T.: 5.550 min
Delta R.T.: -0.007 min
Response: 248995
Conc: 672.18 ng/ml




#20 AR-1242-5

R.T.: 7.194 min
Delta R.T.: -0.004 min
Response: 65198
Conc: 124.66 ng/ml



#20 AR-1242-5


R.T.: 6.116 min
Delta R.T.: -0.006 min
Response: 207840
Conc: 435.76 ng/ml



#21 AR-1248-1


R.T.: 6.212 min
Delta R.T.: -0.005 min
Response: 524436
Conc: 850.12 ng/ml

Instrument :
ECD_P
ClientSampleId :




#21 AR-1248-1

R.T.: 5.240 min
Delta R.T.: -0.006 min
Response: 322374
Conc: 826.86 ng/ml




#22 AR-1248-2

R.T.: 6.507 min
Delta R.T.: -0.004 min
Response: 338967
Conc: 423.44 ng/ml



#22 AR-1248-2


R.T.: 5.505 min
Delta R.T.: -0.007 min
Response: 202267
Conc: 376.31 ng/ml



#23 AR-1248-3


R.T.: 6.725 min
Delta R.T.: -0.005 min
Response: 373308
Conc: 416.34 ng/ml

Instrument :
ECD_P
ClientSampleId :




#23 AR-1248-3

R.T.: 5.550 min
Delta R.T.: -0.007 min
Response: 248995
Conc: 440.47 ng/ml




#24 AR-1248-4

R.T.: 7.154 min
Delta R.T.: -0.005 min
Response: 76480
Conc: 83.11 ng/ml



#24 AR-1248-4


R.T.: 5.736 min
Delta R.T.: -0.007 min
Response: 226098
Conc: 353.80 ng/ml



#25 AR-1248-5


R.T.: 7.194 min
Delta R.T.: -0.004 min
Response: 65198
Conc: 73.70 ng/ml

Instrument :
ECD_P
ClientSampleId :




#25 AR-1248-5

R.T.: 6.158 min
Delta R.T.: -0.008 min
Response: 40427
Conc: 64.80 ng/ml




#26 AR-1254-1

R.T.: 7.123 min
Delta R.T.: -0.005 min
Response: 290262
Conc: 312.85 ng/ml



#26 AR-1254-1


R.T.: 6.116 min
Delta R.T.: -0.007 min
Response: 207840
Conc: 216.95 ng/ml



#27 AR-1254-2


R.T.: 7.352 min
Delta R.T.: -0.006 min
Response: 247282
Conc: 178.62 ng/ml

Instrument :
ECD_P
ClientSampleId :




#27 AR-1254-2

R.T.: 6.276 min
Delta R.T.: -0.007 min
Response: 192597
Conc: 218.22 ng/ml




#28 AR-1254-3

R.T.: 7.736 min
Delta R.T.: -0.005 min
Response: 161630
Conc: 113.45 ng/ml



#28 AR-1254-3


R.T.: 6.716 min
Delta R.T.: 0.010 min
Response: 354919
Conc: 260.33 ng/ml



#29 AR-1254-4


R.T.: 8.031 min
Delta R.T.: -0.007 min
Response: 143112
Conc: 135.24 ng/ml

Instrument :
ECD_P
ClientSampleId :




#29 AR-1254-4

R.T.: 6.938 min
Delta R.T.: -0.007 min
Response: 45505
Conc: 53.05 ng/ml




#30 AR-1254-5

R.T.: 8.461 min
Delta R.T.: -0.006 min
Response: 834908
Conc: 714.30 ng/ml



#30 AR-1254-5


R.T.: 7.369 min
Delta R.T.: -0.007 min
Response: 694648
Conc: 538.15 ng/ml



#31 AR-1260-1


R.T.: 7.902 min
Delta R.T.: -0.007 min
Response: 614490
Conc: 562.52 ng/ml

Instrument :
ECD_P
ClientSampleId :




#31 AR-1260-1

R.T.: 6.836 min
Delta R.T.: -0.009 min
Response: 500536
Conc: 499.72 ng/ml




#32 AR-1260-2

R.T.: 8.169 min
Delta R.T.: -0.006 min
Response: 697743
Conc: 510.51 ng/ml



#32 AR-1260-2


R.T.: 7.033 min
Delta R.T.: -0.008 min
Response: 607325
Conc: 512.90 ng/ml



#33 AR-1260-3


R.T.: 8.535 min
Delta R.T.: -0.005 min
Response: 481465
Conc: 528.36 ng/ml

Instrument :
ECD_P
ClientSampleId :




#33 AR-1260-3

R.T.: 7.189 min
Delta R.T.: -0.009 min
Response: 564053
Conc: 477.40 ng/ml




#34 AR-1260-4

R.T.: 8.766 min
Delta R.T.: -0.006 min
Response: 578357
Conc: 547.94 ng/ml



#34 AR-1260-4


R.T.: 7.674 min
Delta R.T.: -0.009 min
Response: 456873
Conc: 519.81 ng/ml



#35 AR-1260-5


R.T.: 9.090 min
Delta R.T.: -0.007 min
Response: 1080715
Conc: 521.87 ng/ml

Instrument :
ECD_P
ClientSampleId :




#35 AR-1260-5

R.T.: 7.922 min
Delta R.T.: -0.009 min
Response: 1036474
Conc: 507.15 ng/ml




#36 AR-1262-1

R.T.: 8.535 min
Delta R.T.: -0.005 min
Response: 481465
Conc: 353.07 ng/ml



#36 AR-1262-1


R.T.: 7.369 min
Delta R.T.: -0.007 min
Response: 694648
Conc: 988.86 ng/ml



#37 AR-1262-2


R.T.: 9.090 min
Delta R.T.: -0.006 min
Response: 1080715
Conc: 435.79 ng/ml

Instrument :
ECD_P
ClientSampleId :




#37 AR-1262-2

R.T.: 7.922 min
Delta R.T.: -0.007 min
Response: 1036474
Conc: 460.57 ng/ml




#38 AR-1262-3

R.T.: 9.421 min
Delta R.T.: 0.016 min
Response: 744679
Conc: 647.49 ng/ml



#38 AR-1262-3


R.T.: 8.210 min
Delta R.T.: -0.007 min
Response: 228449
Conc: 234.57 ng/ml



#39 AR-1262-4


R.T.: 9.473 min
Delta R.T.: -0.025 min
Response: 456871
Conc: 559.15 ng/ml

Instrument :
ECD_P
ClientSampleId :




#39 AR-1262-4

R.T.: 8.273 min
Delta R.T.: -0.008 min
Response: 794170
Conc: 447.97 ng/ml




#40 AR-1262-5

R.T.: 10.157 min
Delta R.T.: -0.007 min
Response: 279395
Conc: 294.07 ng/ml



#40 AR-1262-5


R.T.: 8.773 min
Delta R.T.: -0.007 min
Response: 262480
Conc: 289.85 ng/ml



#41 AR-1268-1


R.T.: 9.421 min
Delta R.T.: 0.018 min
Response: 744679
Conc: 252.17 ng/ml

Instrument :
ECD_P
ClientSampleId :




#41 AR-1268-1

R.T.: 8.210 min
Delta R.T.: -0.008 min
Response: 228449
Conc: 84.66 ng/ml




#42 AR-1268-2

R.T.: 9.473 min
Delta R.T.: -0.027 min
Response: 456871
Conc: 161.82 ng/ml



#42 AR-1268-2


R.T.: 8.273 min
Delta R.T.: -0.010 min
Response: 794170
Conc: 317.20 ng/ml



#43 AR-1268-3


R.T.: 0.000 min
Exp R.T.: 9.725 min
Response: 0
Conc: N.D.

Instrument :
ECD_P
ClientSampleId :




#43 AR-1268-3

R.T.: 8.483 min
Delta R.T.: -0.008 min
Response: 14909
Conc: 6.81 ng/ml




#44 AR-1268-4

R.T.: 10.157 min
Delta R.T.: -0.006 min
Response: 279395
Conc: 271.09 ng/ml



#44 AR-1268-4


R.T.: 8.773 min
Delta R.T.: -0.007 min
Response: 262480
Conc: 264.69 ng/ml



#45 AR-1268-5

R.T.: 10.569 min
Delta R.T.: -0.008 min
Response: 114067
Conc: 13.26 ng/ml

Instrument :
ECD_P
ClientSampleId :



#45 AR-1268-5

R.T.: 9.056 min
Delta R.T.: -0.008 min
Response: 88547
Conc: 12.42 ng/ml