

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102023\
 Data File : PP061159.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 20 Oct 2023 23:47
 Operator : YP\AJ
 Sample : 04981-01
 Misc :
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 TR-04-10192023

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 21 02:23:29 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP101323.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sat Oct 14 13:32:01 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.559	3.693	38252416	22616053	17.401	16.728
2) SA Decachlor...	10.532	8.766	30890823	20627848	19.726	13.720 #

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102023\
Data File : PP061159.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 20 Oct 2023 23:47
Operator : YP\AJ
Sample : 04981-01
Misc :
ALS Vial : 28 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
TR-04-10192023

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 21 02:23:29 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP101323.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat Oct 14 13:32:01 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.559 min
Delta R.T.: 0.008 min
Response: 38252416
Conc: 17.40 ng/ml

Instrument :
ECD_P
ClientSampleId :
TR-04-10192023




#1 Tetrachloro-m-xylene

R.T.: 3.693 min
Delta R.T.: 0.006 min
Response: 22616053
Conc: 16.73 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.532 min
Delta R.T.: 0.014 min
Response: 30890823
Conc: 19.73 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.766 min
Delta R.T.: 0.013 min
Response: 20627848
Conc: 13.72 ng/ml