

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102122\
 Data File : PP052593.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 21 Oct 2022 18:06
 Operator : YP\AJ
 Sample : AR1254ICC500
 Misc :
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1254ICC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 23 14:46:16 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102122.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sun Oct 23 14:43:54 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.422	3.672	91341116	77255737	50.000	50.000
2) SA Decachlor...	10.237	8.763	85180290	68863979	50.000	50.000
Target Compounds						
26) L6 AR-1254-1	6.455	5.589	30175333	37059864	500.000	500.000
27) L6 AR-1254-2	6.676	5.738	46655303	32756049	500.000	500.000
28) L6 AR-1254-3	7.043	6.148	46510419	50892933	500.000	500.000
29) L6 AR-1254-4	7.330	6.379	29830792	26730600	500.000	500.000
30) L6 AR-1254-5	7.750	6.801	39109847	43215123	500.000	500.000



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102122\
Data File : PP052593.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Oct 2022 18:06
Operator : YP\AJ
Sample : AR1254ICC500
Misc :
ALS Vial : 22 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1254ICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 23 14:46:16 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102122.M
Quant Title : GC EXTRACTABLES
QLast Update : Sun Oct 23 14:43:54 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.422 min
Delta R.T.: 0.000 min
Response: 91341116
Conc: 50.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254ICC500




#1 Tetrachloro-m-xylene

R.T.: 3.672 min
Delta R.T.: 0.000 min
Response: 77255737
Conc: 50.00 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.237 min
Delta R.T.: 0.000 min
Response: 85180290
Conc: 50.00 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.763 min
Delta R.T.: 0.000 min
Response: 68863979
Conc: 50.00 ng/ml



#26 AR-1254-1


R.T.: 6.455 min
Delta R.T.: 0.000 min
Response: 30175333
Conc: 500.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254ICC500




#26 AR-1254-1

R.T.: 5.589 min
Delta R.T.: 0.000 min
Response: 37059864
Conc: 500.00 ng/ml




#27 AR-1254-2

R.T.: 6.676 min
Delta R.T.: 0.000 min
Response: 46655303
Conc: 500.00 ng/ml



#27 AR-1254-2


R.T.: 5.738 min
Delta R.T.: 0.000 min
Response: 32756049
Conc: 500.00 ng/ml



#28 AR-1254-3


R.T.: 7.043 min
Delta R.T.: 0.000 min
Response: 46510419
Conc: 500.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254ICC500




#28 AR-1254-3

R.T.: 6.148 min
Delta R.T.: 0.000 min
Response: 50892933
Conc: 500.00 ng/ml




#29 AR-1254-4

R.T.: 7.330 min
Delta R.T.: 0.000 min
Response: 29830792
Conc: 500.00 ng/ml



#29 AR-1254-4


R.T.: 6.379 min
Delta R.T.: 0.000 min
Response: 26730600
Conc: 500.00 ng/ml



#30 AR-1254-5

R.T.: 7.750 min
Delta R.T.: 0.000 min
Response: 39109847
Conc: 500.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254ICC500



#30 AR-1254-5

R.T.: 6.801 min
Delta R.T.: 0.000 min
Response: 43215123
Conc: 500.00 ng/ml