

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102220\
 Data File : PP030767.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 22 Oct 2020 13:20
 Operator : DD\AJ
 Sample : AR1221CCC500
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1221CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 23 08:08:29 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PP102120.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Oct 22 12:07:08 2020
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.814	4.090	3004784	2676655	45.280	43.132
2) SA Decachlor...	10.700	9.436	1689507	1748371	46.415	46.819
Target Compounds						
8) L2 AR-1221-1	5.060	4.351	337044	305747	523.722	539.194
9) L2 AR-1221-2	5.158	4.451	252592	233075	543.627	559.085
10) L2 AR-1221-3	5.243	4.540	775333	719867	521.626	524.050



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102220\
Data File : PP030767.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 22 Oct 2020 13:20
Operator : DD\AJ
Sample : AR1221CCC500
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1221CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 23 08:08:29 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PP102120.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Oct 22 12:07:08 2020
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.814 min
Delta R.T.: 0.010 min
Response: 3004784
Conc: 45.28 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1221CCC500




#1 Tetrachloro-m-xylene

R.T.: 4.090 min
Delta R.T.: 0.000 min
Response: 2676655
Conc: 43.13 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.700 min
Delta R.T.: 0.014 min
Response: 1689507
Conc: 46.41 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.436 min
Delta R.T.: -0.001 min
Response: 1748371
Conc: 46.82 ng/ml



#8 AR-1221-1


R.T.: 5.060 min
Delta R.T.: 0.009 min
Response: 337044
Conc: 523.72 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1221CCC500




#8 AR-1221-1

R.T.: 4.351 min
Delta R.T.: 0.000 min
Response: 305747
Conc: 539.19 ng/ml




#9 AR-1221-2

R.T.: 5.158 min
Delta R.T.: 0.011 min
Response: 252592
Conc: 543.63 ng/ml



#9 AR-1221-2


R.T.: 4.451 min
Delta R.T.: -0.001 min
Response: 233075
Conc: 559.08 ng/ml



#10 AR-1221-3

R.T.: 5.243 min
Delta R.T.: 0.010 min
Response: 775333
Conc: 521.63 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1221CCC500



#10 AR-1221-3

R.T.: 4.540 min
Delta R.T.: 0.000 min
Response: 719867
Conc: 524.05 ng/ml