

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102224\
 Data File : PP068141.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 22 Oct 2024 18:55
 Operator : YP\AJ
 Sample : P4472-05MS
 Misc :
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 BP-F-6MS

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 23 01:53:41 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP100824.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Oct 09 05:48:32 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.755	4.050	19474857	18462588	21.039	18.273
2) SA Decachlor...	10.666	9.209	23120073	20897680	20.082	18.634
Target Compounds						
3) L1 AR-1016-1	5.920	5.157	15304466	15548556	462.836	444.406
4) L1 AR-1016-2	5.942	5.177	23406774	21224356	479.934	440.627
5) L1 AR-1016-3	6.006	5.358	14535440	11958651	461.268	443.941
6) L1 AR-1016-4	6.104	5.397	11641708	10369169	452.217	430.522
7) L1 AR-1016-5	6.398	5.616	11649536	12823188	430.398	421.679
31) L7 AR-1260-1	7.521	6.659	24872592	24284973	460.402	426.337
32) L7 AR-1260-2	7.774	6.845	27811462	28495834	439.562	427.196
33) L7 AR-1260-3	8.135	7.003	19179490	27101124	369.912	425.442
34) L7 AR-1260-4	8.373	7.478	23126992	20146238	384.205	364.356
35) L7 AR-1260-5	8.710	7.716	41465889	45645131	384.970	373.227



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102224\
Data File : PP068141.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Oct 2024 18:55
Operator : YP\AJ
Sample : P4472-05MS
Misc :
ALS Vial : 18 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
BP-F-6MS

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 23 01:53:41 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP100824.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Oct 09 05:48:32 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.755 min
Delta R.T.: 0.000 min
Response: 19474857
Conc: 21.04 ng/ml

Instrument :
ECD_P
ClientSampleId :
BP-F-6MS




#1 Tetrachloro-m-xylene

R.T.: 4.050 min
Delta R.T.: -0.002 min
Response: 18462588
Conc: 18.27 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.666 min
Delta R.T.: -0.003 min
Response: 23120073
Conc: 20.08 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.209 min
Delta R.T.: -0.010 min
Response: 20897680
Conc: 18.63 ng/ml



#3 AR-1016-1


R.T.: 5.920 min
Delta R.T.: 0.000 min
Response: 15304466
Conc: 462.84 ng/ml

Instrument :
ECD_P
ClientSampleId :
BP-F-6MS




#3 AR-1016-1

R.T.: 5.157 min
Delta R.T.: -0.003 min
Response: 15548556
Conc: 444.41 ng/ml




#4 AR-1016-2

R.T.: 5.942 min
Delta R.T.: 0.000 min
Response: 23406774
Conc: 479.93 ng/ml



#4 AR-1016-2


R.T.: 5.177 min
Delta R.T.: -0.004 min
Response: 21224356
Conc: 440.63 ng/ml



#5 AR-1016-3


R.T.: 6.006 min
Delta R.T.: 0.001 min
Response: 14535440
Conc: 461.27 ng/ml

Instrument :
ECD_P
ClientSampleId :
BP-F-6MS




#5 AR-1016-3

R.T.: 5.358 min
Delta R.T.: -0.004 min
Response: 11958651
Conc: 443.94 ng/ml




#6 AR-1016-4

R.T.: 6.104 min
Delta R.T.: 0.000 min
Response: 11641708
Conc: 452.22 ng/ml



#6 AR-1016-4


R.T.: 5.397 min
Delta R.T.: -0.004 min
Response: 10369169
Conc: 430.52 ng/ml



#7 AR-1016-5


R.T.: 6.398 min
Delta R.T.: 0.000 min
Response: 11649536
Conc: 430.40 ng/ml

Instrument :
ECD_P
ClientSampleId :
BP-F-6MS




#7 AR-1016-5

R.T.: 5.616 min
Delta R.T.: -0.004 min
Response: 12823188
Conc: 421.68 ng/ml




#31 AR-1260-1

R.T.: 7.521 min
Delta R.T.: 0.000 min
Response: 24872592
Conc: 460.40 ng/ml



#31 AR-1260-1


R.T.: 6.659 min
Delta R.T.: -0.005 min
Response: 24284973
Conc: 426.34 ng/ml



#32 AR-1260-2


R.T.: 7.774 min
Delta R.T.: 0.000 min
Response: 27811462
Conc: 439.56 ng/ml

Instrument :
ECD_P
ClientSampleId :
BP-F-6MS




#32 AR-1260-2

R.T.: 6.845 min
Delta R.T.: -0.005 min
Response: 28495834
Conc: 427.20 ng/ml




#33 AR-1260-3

R.T.: 8.135 min
Delta R.T.: -0.002 min
Response: 19179490
Conc: 369.91 ng/ml



#33 AR-1260-3


R.T.: 7.003 min
Delta R.T.: -0.005 min
Response: 27101124
Conc: 425.44 ng/ml



#34 AR-1260-4


R.T.: 8.373 min
Delta R.T.: -0.002 min
Response: 23126992
Conc: 384.20 ng/ml

Instrument :
ECD_P
ClientSampleId :
BP-F-6MS




#34 AR-1260-4

R.T.: 7.478 min
Delta R.T.: -0.006 min
Response: 20146238
Conc: 364.36 ng/ml



#35 AR-1260-5

R.T.: 8.710 min
Delta R.T.: -0.002 min
Response: 41465889
Conc: 384.97 ng/ml



#35 AR-1260-5

R.T.: 7.716 min
Delta R.T.: -0.006 min
Response: 45645131
Conc: 373.23 ng/ml