

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102320\
 Data File : PP030810.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 23 Oct 2020 18:37
 Operator : DD\AJ
 Sample : AR1221CCC500
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1221CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 24 01:55:52 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PP102120.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Oct 22 12:07:08 2020
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.804	4.085	2985363	2797946	44.987	45.087
2)	SA Decachlor...	10.687	9.426	1652039	1757511	45.386	47.064
Target Compounds							
3)	L1 AR-1016-1	6.109	5.350	56497	50437	30.516	29.480
4)	L1 AR-1016-2	6.132	5.371	94871	85505	34.292	34.106
5)	L1 AR-1016-3	6.198	5.562	65918	37745	39.619	30.223
6)	L1 AR-1016-4	6.304	0.000	46998	0	32.397	N.D. #
8)	L2 AR-1221-1	5.049	4.346	329650	304975	512.232	537.834
9)	L2 AR-1221-2	5.147	4.445	245347	224048	528.033	537.431
10)	L2 AR-1221-3	5.232	4.534	767736	706334	516.514	514.197



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102320\
Data File : PP030810.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 23 Oct 2020 18:37
Operator : DD\AJ
Sample : AR1221CCC500
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1221CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 24 01:55:52 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PP102120.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Oct 22 12:07:08 2020
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.804 min
Delta R.T.: 0.000 min
Response: 2985363
Conc: 44.99 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1221CCC500




#1 Tetrachloro-m-xylene

R.T.: 4.085 min
Delta R.T.: -0.005 min
Response: 2797946
Conc: 45.09 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.687 min
Delta R.T.: 0.000 min
Response: 1652039
Conc: 45.39 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.426 min
Delta R.T.: -0.011 min
Response: 1757511
Conc: 47.06 ng/ml



#3 AR-1016-1


R.T.: 6.109 min
Delta R.T.: 0.000 min
Response: 56497
Conc: 30.52 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1221CCC500




#3 AR-1016-1

R.T.: 5.350 min
Delta R.T.: -0.007 min
Response: 50437
Conc: 29.48 ng/ml




#4 AR-1016-2

R.T.: 6.132 min
Delta R.T.: 0.000 min
Response: 94871
Conc: 34.29 ng/ml



#4 AR-1016-2


R.T.: 5.371 min
Delta R.T.: -0.006 min
Response: 85505
Conc: 34.11 ng/ml



#5 AR-1016-3


R.T.: 6.198 min
Delta R.T.: 0.000 min
Response: 65918
Conc: 39.62 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1221CCC500




#5 AR-1016-3

R.T.: 5.562 min
Delta R.T.: -0.007 min
Response: 37745
Conc: 30.22 ng/ml




#6 AR-1016-4

R.T.: 6.304 min
Delta R.T.: 0.001 min
Response: 46998
Conc: 32.40 ng/ml



#6 AR-1016-4


R.T.: 0.000 min
Exp R.T. : 5.622 min
Response: 0
Conc: N.D.



#8 AR-1221-1


R.T.: 5.049 min
Delta R.T.: -0.002 min
Response: 329650
Conc: 512.23 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1221CCC500




#8 AR-1221-1

R.T.: 4.346 min
Delta R.T.: -0.005 min
Response: 304975
Conc: 537.83 ng/ml




#9 AR-1221-2

R.T.: 5.147 min
Delta R.T.: 0.000 min
Response: 245347
Conc: 528.03 ng/ml



#9 AR-1221-2


R.T.: 4.445 min
Delta R.T.: -0.007 min
Response: 224048
Conc: 537.43 ng/ml



#10 AR-1221-3

R.T.: 5.232 min
Delta R.T.: 0.000 min
Response: 767736
Conc: 516.51 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1221CCC500



#10 AR-1221-3

R.T.: 4.534 min
Delta R.T.: -0.006 min
Response: 706334
Conc: 514.20 ng/ml