

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102323\
 Data File : PP061179.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 23 Oct 2023 13:26
 Operator : YP\AJ
 Sample : 04975-01
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 54

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 10/24/2023
 Supervised By :Ankita Jodhani 10/24/2023

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 23 21:31:30 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP101323.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sat Oct 14 13:32:01 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.546	3.694	57579926	22142668	26.193	16.378m#
2) SA Decachlor...	10.536	8.766	27011530	108.2E6	17.249	71.933 #

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102323\
Data File : PP061179.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 Oct 2023 13:26
Operator : YP\AJ
Sample : 04975-01
Misc :
ALS Vial : 15 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
54


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/24/2023
Supervised By :Ankita Jodhani 10/24/2023

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 23 21:31:30 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP101323.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat Oct 14 13:32:01 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.546 min
Delta R.T.: -0.005 min
Response: 57579926
Conc: 26.19 ng/ml

Instrument :
ECD_P
ClientSampleId :
54


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/24/2023
Supervised By :Ankita Jodhani 10/24/2023




#1 Tetrachloro-m-xylene

R.T.: 3.694 min
Delta R.T.: 0.007 min
Response: 22142668
Conc: 16.38 ng/ml m



#2 Decachlorobiphenyl

R.T.: 10.536 min
Delta R.T.: 0.018 min
Response: 27011530
Conc: 17.25 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.766 min
Delta R.T.: 0.012 min
Response: 108153982
Conc: 71.93 ng/ml