

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102323\
 Data File : PP061195.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 23 Oct 2023 22:26
 Operator : YP\AJ
 Sample : AR1660CCC500
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1660CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 24 03:53:22 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP101323.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sat Oct 14 13:32:01 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.556	3.692	127.4E6	70696702	57.932	52.290
2) SA Decachlor...	10.522	8.759	81832597	63831682	52.256	42.455
Target Compounds						
3) L1 AR-1016-1	5.748	4.790	41037756	26371927	571.638	506.721
4) L1 AR-1016-2	5.771	4.810	58379812	35915039	569.171	501.473
5) L1 AR-1016-3	5.834	4.988	35413529	20035972	566.706	495.852
6) L1 AR-1016-4	5.935	5.030	28900754	15875819	565.755	490.810
7) L1 AR-1016-5	6.234	5.246	28553458	20990606	564.757	502.729
31) L7 AR-1260-1	7.375	6.288	47941700	39479574	512.167	477.090
32) L7 AR-1260-2	7.634	6.477	56361278	46722593	519.135	474.473
33) L7 AR-1260-3	7.997	6.632	36238168	44593113	521.399	471.547
34) L7 AR-1260-4	8.228	7.107	74253215	32257090	870.422	450.062 #
35) L7 AR-1260-5	8.569	7.350	92375173	73944609	572.537	440.882



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102323\
Data File : PP061195.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 Oct 2023 22:26
Operator : YP\AJ
Sample : AR1660CCC500
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 24 03:53:22 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP101323.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat Oct 14 13:32:01 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.556 min
Delta R.T.: 0.005 min
Response: 127354131
Conc: 57.93 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.692 min
Delta R.T.: 0.005 min
Response: 70696702
Conc: 52.29 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.522 min
Delta R.T.: 0.004 min
Response: 81832597
Conc: 52.26 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.759 min
Delta R.T.: 0.005 min
Response: 63831682
Conc: 42.45 ng/ml



#3 AR-1016-1


R.T.: 5.748 min
Delta R.T.: 0.007 min
Response: 41037756
Conc: 571.64 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#3 AR-1016-1

R.T.: 4.790 min
Delta R.T.: 0.005 min
Response: 26371927
Conc: 506.72 ng/ml




#4 AR-1016-2

R.T.: 5.771 min
Delta R.T.: 0.007 min
Response: 58379812
Conc: 569.17 ng/ml



#4 AR-1016-2


R.T.: 4.810 min
Delta R.T.: 0.005 min
Response: 35915039
Conc: 501.47 ng/ml



#5 AR-1016-3


R.T.: 5.834 min
Delta R.T.: 0.006 min
Response: 35413529
Conc: 566.71 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#5 AR-1016-3

R.T.: 4.988 min
Delta R.T.: 0.005 min
Response: 20035972
Conc: 495.85 ng/ml




#6 AR-1016-4

R.T.: 5.935 min
Delta R.T.: 0.007 min
Response: 28900754
Conc: 565.76 ng/ml



#6 AR-1016-4


R.T.: 5.030 min
Delta R.T.: 0.004 min
Response: 15875819
Conc: 490.81 ng/ml



#7 AR-1016-5


R.T.: 6.234 min
Delta R.T.: 0.007 min
Response: 28553458
Conc: 564.76 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#7 AR-1016-5

R.T.: 5.246 min
Delta R.T.: 0.005 min
Response: 20990606
Conc: 502.73 ng/ml




#31 AR-1260-1

R.T.: 7.375 min
Delta R.T.: 0.007 min
Response: 47941700
Conc: 512.17 ng/ml



#31 AR-1260-1


R.T.: 6.288 min
Delta R.T.: 0.005 min
Response: 39479574
Conc: 477.09 ng/ml



#32 AR-1260-2


R.T.: 7.634 min
Delta R.T.: 0.007 min
Response: 56361278
Conc: 519.14 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#32 AR-1260-2

R.T.: 6.477 min
Delta R.T.: 0.006 min
Response: 46722593
Conc: 474.47 ng/ml




#33 AR-1260-3

R.T.: 7.997 min
Delta R.T.: 0.007 min
Response: 36238168
Conc: 521.40 ng/ml



#33 AR-1260-3


R.T.: 6.632 min
Delta R.T.: 0.005 min
Response: 44593113
Conc: 471.55 ng/ml



#34 AR-1260-4


R.T.: 8.228 min
Delta R.T.: 0.004 min
Response: 74253215
Conc: 870.42 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#34 AR-1260-4

R.T.: 7.107 min
Delta R.T.: 0.005 min
Response: 32257090
Conc: 450.06 ng/ml



#35 AR-1260-5

R.T.: 8.569 min
Delta R.T.: 0.009 min
Response: 92375173
Conc: 572.54 ng/ml



#35 AR-1260-5

R.T.: 7.350 min
Delta R.T.: 0.006 min
Response: 73944609
Conc: 440.88 ng/ml