

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102422\
 Data File : PP052680.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 25 Oct 2022 05:36
 Operator : YP\AJ
 Sample : PB148511BS
 Misc :
 ALS Vial : 59 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 PB148511BS

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 25 06:24:48 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102122.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sun Oct 23 15:44:19 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.422	3.672	28726798	28140535	17.363	21.120
2) SA Decachlor...	10.234	8.760	31616593	29768780	19.730	23.951
Target Compounds						
3) L1 AR-1016-1	5.595	4.774	18638972	18758337	424.721	407.807
4) L1 AR-1016-2	5.618	4.793	27995991	27872641	443.411	445.192
5) L1 AR-1016-3	5.681	4.973	18054542	14451608	447.204	419.128
6) L1 AR-1016-4	5.779	5.013	13514364	13158578	392.007	453.890
7) L1 AR-1016-5	6.076	5.231	14580608	16256292	381.413	435.272
31) L7 AR-1260-1	7.206	6.278	27822881	30722636	365.302	437.206
32) L7 AR-1260-2	7.464	6.468	29009857	33856436	346.555	422.529
33) L7 AR-1260-3	7.825	6.623	22157606	33332307	343.806	437.909 #
34) L7 AR-1260-4	8.051	7.100	26590988	24500968	352.029	419.205
35) L7 AR-1260-5	8.375	7.343	42951056	50984335	322.558	407.140 #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102422\
Data File : PP052680.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 25 Oct 2022 05:36
Operator : YP\AJ
Sample : PB148511BS
Misc :
ALS Vial : 59 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
PB148511BS

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 25 06:24:48 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102122.M
Quant Title : GC EXTRACTABLES
QLast Update : Sun Oct 23 15:44:19 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.422 min
Delta R.T.: 0.000 min
Response: 28726798
Conc: 17.36 ng/ml

Instrument :
ECD_P
ClientSampleId :
PB148511BS




#1 Tetrachloro-m-xylene

R.T.: 3.672 min
Delta R.T.: 0.000 min
Response: 28140535
Conc: 21.12 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.234 min
Delta R.T.: -0.003 min
Response: 31616593
Conc: 19.73 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.760 min
Delta R.T.: -0.003 min
Response: 29768780
Conc: 23.95 ng/ml



#3 AR-1016-1


R.T.: 5.595 min
Delta R.T.: 0.001 min
Response: 18638972
Conc: 424.72 ng/ml

Instrument :
ECD_P
ClientSampleId :
PB148511BS




#3 AR-1016-1

R.T.: 4.774 min
Delta R.T.: 0.000 min
Response: 18758337
Conc: 407.81 ng/ml




#4 AR-1016-2

R.T.: 5.618 min
Delta R.T.: 0.001 min
Response: 27995991
Conc: 443.41 ng/ml



#4 AR-1016-2


R.T.: 4.793 min
Delta R.T.: 0.000 min
Response: 27872641
Conc: 445.19 ng/ml



#5 AR-1016-3


R.T.: 5.681 min
Delta R.T.: 0.000 min
Response: 18054542
Conc: 447.20 ng/ml

Instrument :
ECD_P
ClientSampleId :
PB148511BS




#5 AR-1016-3

R.T.: 4.973 min
Delta R.T.: 0.000 min
Response: 14451608
Conc: 419.13 ng/ml




#6 AR-1016-4

R.T.: 5.779 min
Delta R.T.: 0.000 min
Response: 13514364
Conc: 392.01 ng/ml



#6 AR-1016-4


R.T.: 5.013 min
Delta R.T.: 0.000 min
Response: 13158578
Conc: 453.89 ng/ml



#7 AR-1016-5


R.T.: 6.076 min
Delta R.T.: 0.000 min
Response: 14580608
Conc: 381.41 ng/ml

Instrument :
ECD_P
ClientSampleId :
PB148511BS




#7 AR-1016-5

R.T.: 5.231 min
Delta R.T.: 0.000 min
Response: 16256292
Conc: 435.27 ng/ml




#31 AR-1260-1

R.T.: 7.206 min
Delta R.T.: 0.000 min
Response: 27822881
Conc: 365.30 ng/ml



#31 AR-1260-1


R.T.: 6.278 min
Delta R.T.: 0.000 min
Response: 30722636
Conc: 437.21 ng/ml



#32 AR-1260-2


R.T.: 7.464 min
Delta R.T.: 0.000 min
Response: 29009857
Conc: 346.55 ng/ml

Instrument :
ECD_P
ClientSampleId :
PB148511BS




#32 AR-1260-2

R.T.: 6.468 min
Delta R.T.: 0.000 min
Response: 33856436
Conc: 422.53 ng/ml




#33 AR-1260-3

R.T.: 7.825 min
Delta R.T.: 0.000 min
Response: 22157606
Conc: 343.81 ng/ml



#33 AR-1260-3


R.T.: 6.623 min
Delta R.T.: 0.000 min
Response: 33332307
Conc: 437.91 ng/ml



#34 AR-1260-4


R.T.: 8.051 min
Delta R.T.: 0.000 min
Response: 26590988
Conc: 352.03 ng/ml

Instrument :
ECD_P
ClientSampleId :
PB148511BS




#34 AR-1260-4

R.T.: 7.100 min
Delta R.T.: 0.000 min
Response: 24500968
Conc: 419.20 ng/ml



#35 AR-1260-5

R.T.: 8.375 min
Delta R.T.: -0.001 min
Response: 42951056
Conc: 322.56 ng/ml



#35 AR-1260-5

R.T.: 7.343 min
Delta R.T.: 0.000 min
Response: 50984335
Conc: 407.14 ng/ml