

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102422\
 Data File : PP052681.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 25 Oct 2022 05:53
 Operator : YP\AJ
 Sample : N5234-08
 Misc :
 ALS Vial : 60 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 LSP-SS-05-08-13

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 25 06:25:16 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102122.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sun Oct 23 15:44:19 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.422	3.672	22934532	23609260	13.862	17.720 #
2) SA Decachlor...	10.235	8.761	24653524	19453136	15.385	15.651

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102422\
Data File : PP052681.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 25 Oct 2022 05:53
Operator : YP\AJ
Sample : N5234-08
Misc :
ALS Vial : 60 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
LSP-SS-05-08-13

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 25 06:25:16 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102122.M
Quant Title : GC EXTRACTABLES
QLast Update : Sun Oct 23 15:44:19 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.422 min
Delta R.T.: 0.000 min
Response: 22934532
Conc: 13.86 ng/ml

Instrument :
ECD_P
ClientSampleId :
LSP-SS-05-08-13




#1 Tetrachloro-m-xylene

R.T.: 3.672 min
Delta R.T.: 0.000 min
Response: 23609260
Conc: 17.72 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.235 min
Delta R.T.: -0.003 min
Response: 24653524
Conc: 15.38 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.761 min
Delta R.T.: -0.001 min
Response: 19453136
Conc: 15.65 ng/ml