

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102422\
 Data File : PP052689.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 25 Oct 2022 08:38
 Operator : YP\AJ
 Sample : AR1248CCC500
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1248CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 25 11:06:48 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102122.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sun Oct 23 15:44:19 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.422	3.671	69879555	73864406	42.237	55.438 #
2) SA Decachlor...	10.239	8.760	69474284	67763819	43.354	54.521 #
Target Compounds						
21) L5 AR-1248-1	5.595	4.773	13429954	14072843	462.866	495.909
22) L5 AR-1248-2	5.870	5.013	20042408	20430895	472.898	498.453
23) L5 AR-1248-3	6.075	5.055	21588558	21198456	417.679	504.155
24) L5 AR-1248-4	6.481	5.230	21276915	24832096	386.097	506.336 #
25) L5 AR-1248-5	6.520	5.630	22346713	20059246	393.173	482.848



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102422\
Data File : PP052689.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 25 Oct 2022 08:38
Operator : YP\AJ
Sample : AR1248CCC500
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 25 11:06:48 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102122.M
Quant Title : GC EXTRACTABLES
QLast Update : Sun Oct 23 15:44:19 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.422 min
Delta R.T.: 0.000 min
Response: 69879555
Conc: 42.24 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.671 min
Delta R.T.: 0.000 min
Response: 73864406
Conc: 55.44 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.239 min
Delta R.T.: 0.001 min
Response: 69474284
Conc: 43.35 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.760 min
Delta R.T.: -0.002 min
Response: 67763819
Conc: 54.52 ng/ml



#21 AR-1248-1


R.T.: 5.595 min
Delta R.T.: 0.000 min
Response: 13429954
Conc: 462.87 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500




#21 AR-1248-1

R.T.: 4.773 min
Delta R.T.: 0.000 min
Response: 14072843
Conc: 495.91 ng/ml




#22 AR-1248-2

R.T.: 5.870 min
Delta R.T.: 0.000 min
Response: 20042408
Conc: 472.90 ng/ml



#22 AR-1248-2


R.T.: 5.013 min
Delta R.T.: -0.001 min
Response: 20430895
Conc: 498.45 ng/ml



#23 AR-1248-3


R.T.: 6.075 min
Delta R.T.: 0.000 min
Response: 21588558
Conc: 417.68 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500




#23 AR-1248-3

R.T.: 5.055 min
Delta R.T.: -0.002 min
Response: 21198456
Conc: 504.16 ng/ml




#24 AR-1248-4

R.T.: 6.481 min
Delta R.T.: 0.000 min
Response: 21276915
Conc: 386.10 ng/ml



#24 AR-1248-4


R.T.: 5.230 min
Delta R.T.: -0.001 min
Response: 24832096
Conc: 506.34 ng/ml



#25 AR-1248-5

R.T.: 6.520 min
Delta R.T.: 0.000 min
Response: 22346713
Conc: 393.17 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500



#25 AR-1248-5

R.T.: 5.630 min
Delta R.T.: -0.001 min
Response: 20059246
Conc: 482.85 ng/ml