

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102423\
 Data File : PP061222.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 24 Oct 2023 16:50
 Operator : YP\AJ
 Sample : 05010-01
 Misc :
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 343

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 10/25/2023
 Supervised By :Ankita Jodhani 10/25/2023

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 25 01:14:08 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP101323.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sat Oct 14 13:32:01 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.564	3.697	41763197	26660268	18.998	19.719m
2) SA Decachlor...	10.537	8.770	19400868	16400602	12.389	10.908

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102423\
Data File : PP061222.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 24 Oct 2023 16:50
Operator : YP\AJ
Sample : 05010-01
Misc :
ALS Vial : 14 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
343


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/25/2023
Supervised By :Ankita Jodhani 10/25/2023

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 25 01:14:08 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP101323.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat Oct 14 13:32:01 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.564 min
Delta R.T.: 0.013 min
Response: 41763197
Conc: 19.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
343


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/25/2023
Supervised By :Ankita Jodhani 10/25/2023




#1 Tetrachloro-m-xylene

R.T.: 3.697 min
Delta R.T.: 0.010 min
Response: 26660268
Conc: 19.72 ng/ml m



#2 Decachlorobiphenyl

R.T.: 10.537 min
Delta R.T.: 0.017 min
Response: 19400868
Conc: 12.39 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.770 min
Delta R.T.: 0.016 min
Response: 16400602
Conc: 10.91 ng/ml