

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102424\
 Data File : PP068192.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 24 Oct 2024 10:49
 Operator : YP\AJ
 Sample : AR1254CCC500
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1254CCC500

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 10/25/2024
 Supervised By :Ankita Jodhani 10/25/2024

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 25 02:35:21 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP100824.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Oct 09 05:48:32 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.756	4.050	49949159	50153959	53.962	49.638
2) SA Decachlor...	10.666	9.209	60513641	57881786	52.561	51.613
Target Compounds						
26) L6 AR-1254-1	6.773	5.971	21861267	29532791	488.166m	499.845m
27) L6 AR-1254-2	6.990	6.118	31772080	25485267	480.683m	485.765m
28) L6 AR-1254-3	7.355	6.527	34787443	38655273	500.880m	464.562m
29) L6 AR-1254-4	7.639	6.756	25685321	24690495	514.689	511.463
30) L6 AR-1254-5	8.057	7.178	28961158	36193848	488.444	492.090

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102424\
Data File : PP068192.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 24 Oct 2024 10:49
Operator : YP\AJ
Sample : AR1254CCC500
Misc :
ALS Vial : 6 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
AR1254CCC500


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/25/2024
Supervised By :Ankita Jodhani 10/25/2024

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 25 02:35:21 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP100824.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Oct 09 05:48:32 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.756 min
Delta R.T.: 0.000 min
Response: 49949159
Conc: 53.96 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/25/2024
Supervised By :Ankita Jodhani 10/25/2024




#1 Tetrachloro-m-xylene

R.T.: 4.050 min
Delta R.T.: -0.002 min
Response: 50153959
Conc: 49.64 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.666 min
Delta R.T.: -0.002 min
Response: 60513641
Conc: 52.56 ng/ml



#2 Decachlorobiphenyl

R.T.: 9.209 min
Delta R.T.: -0.010 min
Response: 57881786
Conc: 51.61 ng/ml




#26 AR-1254-1

R.T.: 6.773 min
Delta R.T.: 0.000 min
Response: 21861267
Conc: 488.17 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/25/2024
Supervised By :Ankita Jodhani 10/25/2024




#26 AR-1254-1

R.T.: 5.971 min
Delta R.T.: -0.004 min
Response: 29532791
Conc: 499.84 ng/ml




#27 AR-1254-2

R.T.: 6.990 min
Delta R.T.: 0.000 min
Response: 31772080
Conc: 480.68 ng/ml



#27 AR-1254-2

R.T.: 6.118 min
Delta R.T.: -0.004 min
Response: 25485267
Conc: 485.77 ng/ml




#28 AR-1254-3

R.T.: 7.355 min
Delta R.T.: 0.000 min
Response: 34787443
Conc: 500.88 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/25/2024
Supervised By :Ankita Jodhani 10/25/2024




#28 AR-1254-3

R.T.: 6.527 min
Delta R.T.: -0.005 min
Response: 38655273
Conc: 464.56 ng/ml




#29 AR-1254-4

R.T.: 7.639 min
Delta R.T.: 0.001 min
Response: 25685321
Conc: 514.69 ng/ml



#29 AR-1254-4

R.T.: 6.756 min
Delta R.T.: -0.004 min
Response: 24690495
Conc: 511.46 ng/ml




#30 AR-1254-5

R.T.: 8.057 min
Delta R.T.: 0.002 min
Response: 28961158
Conc: 488.44 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/25/2024
Supervised By :Ankita Jodhani 10/25/2024



#30 AR-1254-5

R.T.: 7.178 min
Delta R.T.: -0.004 min
Response: 36193848
Conc: 492.09 ng/ml