

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102425\
 Data File : PP076039.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 24 Oct 2025 17:21
 Operator : YP\AJ
 Sample : Q3446-02DL 100X
 Misc :
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 SB-15DL

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 10/27/2025
 Supervised By :mohammad ahmed 10/28/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 24 23:37:45 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Oct 24 15:40:32 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
----------	------	------	--------	--------	-------	-------

System Monitoring Compounds

Target Compounds								
16)	L4	AR-1242-1	5.791	4.875	12513316	149.7E6	261.202m	260.436
17)	L4	AR-1242-2	5.812	4.932	17108267	69281032	241.670m	249.651
18)	L4	AR-1242-3	5.875	5.053	11233319	39073565	246.119m	248.225
19)	L4	AR-1242-4	5.972	5.138	9311654	48383290	256.101m	261.373
20)	L4	AR-1242-5	6.702	5.655	3598527	30166650	87.889m	131.024 #

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102425\
Data File : PP076039.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 24 Oct 2025 17:21
Operator : YP\AJ
Sample : Q3446-02DL 100X
Misc :
ALS Vial : 24 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
SB-15DL


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/27/2025
Supervised By :mohammad ahmed 10/28/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 24 23:37:45 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 24 15:40:32 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 0.000 min
Exp R.T. : 4.642 min
Response: 0
Conc: N.D.

Instrument :
ECD_P
ClientSampleId :
SB-15DL


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/27/2025
Supervised By :mohammad ahmed 10/28/2025




#1 Tetrachloro-m-xylene

R.T.: 0.000 min
Exp R.T. : 3.781 min
Response: 0
Conc: N.D.




#2 Decachlorobiphenyl

R.T.: 0.000 min
Exp R.T. : 10.419 min
Response: 0
Conc: N.D.



#2 Decachlorobiphenyl

R.T.: 0.000 min
Exp R.T. : 8.782 min
Response: 0
Conc: N.D.




#16 AR-1242-1

R.T.: 5.791 min
Delta R.T.: -0.008 min
Response: 12513316
Conc: 261.20 ng/ml

Instrument :
ECD_P
ClientSampleId :
SB-15DL


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/27/2025
Supervised By :mohammad ahmed 10/28/2025




#16 AR-1242-1

R.T.: 4.875 min
Delta R.T.: -0.005 min
Response: 149686162
Conc: 260.44 ng/ml




#17 AR-1242-2

R.T.: 5.812 min
Delta R.T.: -0.007 min
Response: 17108267
Conc: 241.67 ng/ml m



#17 AR-1242-2

R.T.: 4.932 min
Delta R.T.: -0.005 min
Response: 69281032
Conc: 249.65 ng/ml




#18 AR-1242-3

R.T.: 5.875 min
Delta R.T.: -0.007 min
Response: 11233319
Conc: 246.12 ng/ml

Instrument :
ECD_P
ClientSampleId :
SB-15DL


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/27/2025
Supervised By :mohammad ahmed 10/28/2025




#18 AR-1242-3

R.T.: 5.053 min
Delta R.T.: -0.005 min
Response: 39073565
Conc: 248.23 ng/ml




#19 AR-1242-4

R.T.: 5.972 min
Delta R.T.: -0.007 min
Response: 9311654
Conc: 256.10 ng/ml m



#19 AR-1242-4

R.T.: 5.138 min
Delta R.T.: -0.002 min
Response: 48383290
Conc: 261.37 ng/ml




#20 AR-1242-5

R.T.: 6.702 min
Delta R.T.: -0.008 min
Response: 3598527
Conc: 87.89 ng/ml

Instrument :
ECD_P
ClientSampleId :
SB-15DL

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/27/2025
Supervised By :mohammad ahmed 10/28/2025



#20 AR-1242-5

R.T.: 5.655 min
Delta R.T.: -0.006 min
Response: 30166650
Conc: 131.02 ng/ml