

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_0\methods\
 Method File : AR1232PP102520.M
 Title : GC EXTRACTABLES
 Last Update : Mon Oct 26 04:46:19 2020
 Response Via : Initial Calibration

Calibration Files

500 =PP030834.D 1000 =PP030832.D 750 =PP030833.D
 250 =PP030835.D 50 =PP030836.D

	Compound	500	1000	750	250	50	Avg		%RSD
1)	SA Tetrachloro-m-xylene	6.651	6.313	6.432	6.666	7.455	6.703	E4	6.65
2)	SA Decachlorobiphenyl	3.659	3.515	3.653	3.763	3.775	3.673	E4	2.86
3)	L1 AR-1016-1						0.000		-1.00
4)	L1 AR-1016-2						0.000		-1.00
5)	L1 AR-1016-3						0.000		-1.00
6)	L1 AR-1016-4						0.000		-1.00
7)	L1 AR-1016-5						0.000		-1.00
8)	L2 AR-1221-1						0.000		-1.00
9)	L2 AR-1221-2						0.000		-1.00
10)	L2 AR-1221-3						0.000		-1.00
11)	L3 AR-1232-1	1.283	1.172	1.219	1.339	1.499	1.302	E3	9.74
12)	L3 AR-1232-2	6.275	5.724	5.950	6.475	6.620	6.209	E2	5.95
13)	L3 AR-1232-3	1.151	1.054	1.096	1.199	1.308	1.162	E3	8.48
14)	L3 AR-1232-4	5.868	5.372	5.630	6.101	6.015	5.797	E2	5.13
15)	L3 AR-1232-5	3.498	3.201	3.365	3.631	4.628	3.665	E2	15.33
16)	L4 AR-1242-1						0.000		-1.00
17)	L4 AR-1242-2						0.000		-1.00
18)	L4 AR-1242-3						0.000		-1.00
19)	L4 AR-1242-4						0.000		-1.00
20)	L4 AR-1242-5						0.000		-1.00
21)	L5 AR-1248-1						0.000		-1.00
22)	L5 AR-1248-2						0.000		-1.00
23)	L5 AR-1248-3						0.000		-1.00
24)	L5 AR-1248-4						0.000		-1.00
25)	L5 AR-1248-5						0.000		-1.00
26)	L6 AR-1254-1						0.000		-1.00
27)	L6 AR-1254-2						0.000		-1.00
28)	L6 AR-1254-3						0.000		-1.00
29)	L6 AR-1254-4						0.000		-1.00
30)	L6 AR-1254-5						0.000		-1.00
31)	L7 AR-1260-1						0.000		-1.00
32)	L7 AR-1260-2						0.000		-1.00
33)	L7 AR-1260-3						0.000		-1.00
34)	L7 AR-1260-4						0.000		-1.00
35)	L7 AR-1260-5						0.000		-1.00
36)	L8 AR-1262-1						0.000		-1.00
37)	L8 AR-1262-2						0.000		-1.00
38)	L8 AR-1262-3						0.000		-1.00
39)	L8 AR-1262-4						0.000		-1.00
40)	L8 AR-1262-5						0.000		-1.00
41)	L9 AR-1268-1						0.000		-1.00
42)	L9 AR-1268-2						0.000		-1.00
43)	L9 AR-1268-3						0.000		-1.00
44)	L9 AR-1268-4						0.000		-1.00
45)	L9 AR-1268-5						0.000		-1.00

Signal #2 Calibration Files

500 =PP030834.D 1000 =PP030832.D 750 =PP030833.D
 250 =PP030835.D 50 =PP030836.D

	Compound	500	1000	750	250	50	Avg		%RSD
1)	SA Tetrachloro-m-xylene	6.199	5.887	6.004	6.193	6.807	6.218	E4	5.71
2)	SA Decachlorobiphenyl	3.908	3.851	3.980	3.988	4.056	3.956	E4	2.00
3)	L1 AR-1016-1						0.000		-1.00
4)	L1 AR-1016-2						0.000		-1.00
5)	L1 AR-1016-3						0.000		-1.00

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\
 Method File : AR1232PP102520.M
 Title : GC EXTRACTABLES
 Last Update : Mon Oct 26 04:46:19 2020
 Response Via : Initial Calibration

Calibration Files

500 =PP030834.D 1000 =PP030832.D 750 =PP030833.D
 250 =PP030835.D 50 =PP030836.D

	Compound	500	1000	750	250	50	Avg	%RSD
6) L1	AR-1016-4						0.000	-1.00
7) L1	AR-1016-5						0.000	-1.00
8) L2	AR-1221-1						0.000	-1.00
9) L2	AR-1221-2						0.000	-1.00
10) L2	AR-1221-3						0.000	-1.00
11) L3	AR-1232-1	1.189	1.084	1.127	1.213	1.307	1.184 E3	7.22
12) L3	AR-1232-2	1.035	0.961	0.995	1.087	1.164	1.049 E3	7.63
13) L3	AR-1232-3	5.521	5.075	5.302	5.550	5.371	5.364 E2	3.57
14) L3	AR-1232-4	4.384	4.056	4.238	3.715	4.176	4.114 E2	6.13
15) L3	AR-1232-5	4.854	4.490	4.701	4.975	5.220	4.848 E2	5.69
16) L4	AR-1242-1						0.000	-1.00
17) L4	AR-1242-2						0.000	-1.00
18) L4	AR-1242-3						0.000	-1.00
19) L4	AR-1242-4						0.000	-1.00
20) L4	AR-1242-5						0.000	-1.00
21) L5	AR-1248-1						0.000	-1.00
22) L5	AR-1248-2						0.000	-1.00
23) L5	AR-1248-3						0.000	-1.00
24) L5	AR-1248-4						0.000	-1.00
25) L5	AR-1248-5						0.000	-1.00
26) L6	AR-1254-1						0.000	-1.00
27) L6	AR-1254-2						0.000	-1.00
28) L6	AR-1254-3						0.000	-1.00
29) L6	AR-1254-4						0.000	-1.00
30) L6	AR-1254-5						0.000	-1.00
31) L7	AR-1260-1						0.000	-1.00
32) L7	AR-1260-2						0.000	-1.00
33) L7	AR-1260-3						0.000	-1.00
34) L7	AR-1260-4						0.000	-1.00
35) L7	AR-1260-5						0.000	-1.00
36) L8	AR-1262-1						0.000	-1.00
37) L8	AR-1262-2						0.000	-1.00
38) L8	AR-1262-3						0.000	-1.00
39) L8	AR-1262-4						0.000	-1.00
40) L8	AR-1262-5						0.000	-1.00
41) L9	AR-1268-1						0.000	-1.00
42) L9	AR-1268-2						0.000	-1.00
43) L9	AR-1268-3						0.000	-1.00
44) L9	AR-1268-4						0.000	-1.00
45) L9	AR-1268-5						0.000	-1.00

(#) = Out of Range