

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS102618\
 Data File : PS002153.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 27 Oct 2018 04:47
 Operator : EI\SJ
 Sample : J5204-14
 Misc :
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 29 09:04:26 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS102618.M
 Quant Title : 8080.M
 QLast Update : Sat Oct 27 01:34:10 2018
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	11.435	10.912	110.2E6	99559924	350.120	277.250
Target Compounds							
1) T	Dalapon	2.895	2.338	37567608	62811372	81.906	139.115 #
2) T	3,5-DICHL...	0.000	9.075	0	3732282	N.D.	8.414 #
3) T	4-Nitroph...	11.086	10.069	4605338	7229561	27.605	40.545 #
5) T	DICAMBA	0.000	11.154	0	8432261	N.D.	6.747 #
6) T	MCP	12.117	11.417	13091941	632123	13.253	0.735 #
7) T	MCPA	0.000	11.814	0	4898803	N.D.	3.620 #
8) T	DICHLORPROP	0.000	12.267	0	1547458	N.D.	4.819 #
9) T	2,4-D	0.000	12.778	0	678784	N.D.	1.594 #
11) T	2,4,5-TP ...	0.000	13.978	0	844869	N.D.	0.462 #
13) T	2,4-DB	0.000	15.111	0	120322	N.D.	0.543 #
14) T	DINOSEB	0.000	15.646	0	1779127	N.D.	1.405 #
15) T	Picloram	16.934	16.771	5798427	1411796	2.431	0.624 #
16) T	DCPA	0.000	16.833	0	4591183	N.D.	1.939 #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS102618\
Data File : PS002153.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Oct 2018 04:47
Operator : EI\SJ
Sample : J5204-14
Misc :
ALS Vial : 17 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 29 09:04:26 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS102618.M
Quant Title : 8080.M
QLast Update : Sat Oct 27 01:34:10 2018
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Dalapon


R.T.: 2.895 min
Delta R.T.: -0.047 min
Response: 37567608
Conc: 81.91 ng/ml

Instrument :
ECD_P
ClientSampleId :




#1 Dalapon

R.T.: 2.338 min
Delta R.T.: -0.036 min
Response: 62811372
Conc: 139.11 ng/ml




#2 3,5-DICHLOROBENZOIC ACID

R.T.: 0.000 min
Exp R.T. : 9.870 min
Response: 0
Conc: N.D.



#2 3,5-DICHLOROBENZOIC ACID


R.T.: 9.075 min
Delta R.T.: 0.005 min
Response: 3732282
Conc: 8.41 ng/ml



#3 4-Nitrophenol


R.T.: 11.086 min
Delta R.T.: 0.046 min
Response: 4605338
Conc: 27.61 ng/ml

Instrument :
ECD_P
ClientSampleId :




#3 4-Nitrophenol

R.T.: 10.069 min
Delta R.T.: 0.005 min
Response: 7229561
Conc: 40.54 ng/ml




#4 2,4-DCAA

R.T.: 11.435 min
Delta R.T.: -0.006 min
Response: 110177866
Conc: 350.12 ng/ml



#4 2,4-DCAA


R.T.: 10.912 min
Delta R.T.: -0.002 min
Response: 99559924
Conc: 277.25 ng/ml



#5 DICAMBA


R.T.: 0.000 min
Exp R.T.: 11.750 min
Response: 0
Conc: N.D.

Instrument :
ECD_P
ClientSampleId :




#5 DICAMBA

R.T.: 11.154 min
Delta R.T.: -0.043 min
Response: 8432261
Conc: 6.75 ng/ml




#6 MCP

R.T.: 12.117 min
Delta R.T.: 0.034 min
Response: 13091941
Conc: 13.25 ug/ml



#6 MCP


R.T.: 11.417 min
Delta R.T.: -0.003 min
Response: 632123
Conc: 0.73 ug/ml



#7 MCPA


R.T.: 0.000 min
Exp R.T.: 12.333 min
Response: 0
Conc: N.D.

Instrument :
ECD_P
ClientSampleId :




#7 MCPA

R.T.: 11.814 min
Delta R.T.: 0.019 min
Response: 4898803
Conc: 3.62 ug/ml




#8 DICHLORPROP

R.T.: 0.000 min
Exp R.T.: 12.937 min
Response: 0
Conc: N.D.



#8 DICHLORPROP


R.T.: 12.267 min
Delta R.T.: -0.052 min
Response: 1547458
Conc: 4.82 ng/ml



#9 2,4-D


R.T.: 0.000 min
Exp R.T.: 13.298 min
Response: 0
Conc: N.D.

Instrument :
ECD_P
ClientSampleId :




#9 2,4-D

R.T.: 12.778 min
Delta R.T.: -0.013 min
Response: 678784
Conc: 1.59 ng/ml




#11 2,4,5-TP (SILVEX)

R.T.: 0.000 min
Exp R.T.: 14.542 min
Response: 0
Conc: N.D.



#11 2,4,5-TP (SILVEX)


R.T.: 13.978 min
Delta R.T.: 0.025 min
Response: 844869
Conc: 0.46 ng/ml



#13 2,4-DB


R.T.: 0.000 min
Exp R.T.: 15.688 min
Response: 0
Conc: N.D.

Instrument :
ECD_P
ClientSampleId :




#13 2,4-DB

R.T.: 15.111 min
Delta R.T.: -0.056 min
Response: 120322
Conc: 0.54 ng/ml




#14 DINOSEB

R.T.: 0.000 min
Exp R.T.: 17.168 min
Response: 0
Conc: N.D.



#14 DINOSEB


R.T.: 15.646 min
Delta R.T.: 0.051 min
Response: 1779127
Conc: 1.41 ng/ml



#15 Picloram


R.T.: 16.934 min
Delta R.T.: -0.004 min
Response: 5798427
Conc: 2.43 ng/ml

Instrument :
ECD_P
ClientSampleId :




#15 Picloram

R.T.: 16.771 min
Delta R.T.: 0.000 min
Response: 1411796
Conc: 0.62 ng/ml



#16 DCPA

R.T.: 0.000 min
Exp R.T. : 17.509 min
Response: 0
Conc: N.D.



#16 DCPA

R.T.: 16.833 min
Delta R.T.: 0.006 min
Response: 4591183
Conc: 1.94 ng/ml